

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

## LAB CHRONICLE

<b>OrderID:</b>	Q2320	<b>OrderDate:</b>	6/13/2025 12:11:26 PM
<b>Client:</b>	First Environment, Inc.	<b>Project:</b>	EDGEW001 – Veterans Field Edgewater, NJ
<b>Contact:</b>	Ken Cwieka	<b>Location:</b>	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2320-01	WC	SOIL			06/13/25 09:30			06/13/25
			Corrosivity	9045D			06/17/25 15:50	
			Ignitability	1030			06/16/25 12:22	
			Reactive Cyanide	9012B		06/16/25	06/16/25 12:22	
			Reactive Sulfide	9034		06/16/25	06/16/25 13:40	



# SAMPLE DATA

## Report of Analysis

Client:	First Environment, Inc.	Date Collected:	06/13/25 09:30
Project:	EDGEW001 – Veterans Field Edgewater, NJ	Date Received:	06/13/25
Client Sample ID:	WC	SDG No.:	Q2320
Lab Sample ID:	Q2320-01	Matrix:	SOIL
		% Solid:	88.7

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	7.80	H	1	0	0	pH		06/17/25 15:50	9045D
Ignitability	NO		1	0	0	oC		06/16/25 12:22	1030
Reactive Cyanide	0.014	J	1	0.0083	0.049	mg/Kg	06/16/25 08:30	06/16/25 12:22	9012B
Reactive Sulfide	3.16	J	1	0.20	10.0	mg/Kg	06/16/25 08:30	06/16/25 13:40	9034

Comments: pH result reported at temperature 20.9 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



# QC RESULT SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** First Environment, Inc.

**SDG No.:** Q2320

**Project:** EDGEW001 – Veterans Field Edgewater, NJ

**RunNo.:** LB136162

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV1</b> Reactive Cyanide	mg/L	0.094	0.099	95	85–115	06/16/2025
Sample ID: <b>CCV1</b> Reactive Cyanide	mg/L	0.23	0.25	92	90–110	06/16/2025
Sample ID: <b>CCV2</b> Reactive Cyanide	mg/L	0.23	0.25	92	90–110	06/16/2025

## Initial and Continuing Calibration Verification

**Client:** First Environment, Inc.

**SDG No.:** Q2320

**Project:** EDGEW001 – Veterans Field Edgewater, NJ

**RunNo.:** LB136178

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.00	7	100	90-110	06/17/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	06/17/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	06/17/2025
Sample ID: Corrosivity	CCV3	pH	2.02	2.00	101	90-110	06/17/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Blank Summary

**Client:** First Environment, Inc.

**SDG No.:** Q2320

**Project:** EDGEW001 – Veterans Field Edgewater, NJ

**RunNo.:** LB136162

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>ICB1</b> Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/16/2025
Sample ID: <b>CCB1</b> Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/16/2025
Sample ID: <b>CCB2</b> Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	06/16/2025



## Preparation Blank Summary

**Client:** First Environment, Inc.

**SDG No.:** Q2320

**Project:** EDGEW001 – Veterans Field Edgewater, NJ

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>PB168487BL</b>							
Reactive Cyanide	mg/Kg	0.013	0.0250	J	0.0083	0.05	06/16/2025
Sample ID: <b>PB168488BL</b>							
Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.201	10	06/16/2025

## Duplicate Sample Summary

<b>Client:</b> First Environment, Inc.	<b>SDG No.:</b> Q2320
<b>Project:</b> EDGEW001 – Veterans Field Edgewater, NJ	<b>Sample ID:</b> Q2301-03
<b>Client ID:</b> WC-URBAN-FILL-CDUP	<b>Percent Solids for Spike Sample:</b> 100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		06/16/2025
Reactive Cyanide	mg/Kg	+/-20	0.013	J	0.012	J	1	8		06/16/2025
Reactive Sulfide	mg/Kg	+/-20	4.77	J	4.79	J	1	0.42		06/16/2025

## Duplicate Sample Summary

<b>Client:</b> First Environment, Inc. <b>Project:</b> EDGEW001 – Veterans Field Edgewater, NJ <b>Client ID:</b> MH-9DUP	<b>SDG No.:</b> Q2320 <b>Sample ID:</b> Q2347-04 <b>Percent Solids for Spike Sample:</b> 100
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	7.40		7.41		1	0.14		06/17/2025



# RAW DATA

## Analytical Summary Report

Analysis Method: 1030  
Parameter: Ignitability  
Run Number: LB136161

Reviewed By: Eman

Supervisor Review By: Iwona

Seq	LabID	ClientID	DF	matrix	Result Status	Burning Rate	Anal Date	Anal Time
1	Q2301-03	WC-URBAN-FILL-C	1	Solid	NO	0.00	06/16/2025	10:45
2	Q2301-03DUP	WC-URBAN-FILL-CDUP	1	Solid	NO	0.00	06/16/2025	10:52
3	Q2307-01	LINDEN-SAA	1	Solid	NO	0.00	06/16/2025	11:00
4	Q2310-01	TP-7	1	Solid	NO	0.00	06/16/2025	11:07
5	Q2310-04	TP-7	1	Solid	NO	0.00	06/16/2025	11:14
6	Q2311-01	TP03-MH2MH3-WC	1	Solid	NO	0.00	06/16/2025	11:22
7	Q2311-04	TP03-MH2MH3-WC	1	Solid	NO	0.00	06/16/2025	11:30
8	Q2311-05	TP04-MH2MH3-WC	1	Solid	NO	0.00	06/16/2025	11:37
9	Q2311-08	TP04-MH2MH3-WC	1	Solid	NO	0.00	06/16/2025	11:45
10	Q2312-01	TP-1	1	Solid	NO	0.00	06/16/2025	11:52
11	Q2312-04	TP-1	1	Solid	NO	0.00	06/16/2025	12:00
12	Q2319-01	MH-B	1	Solid	NO	0.00	06/16/2025	12:07
13	Q2319-04	MH-B	1	Solid	NO	0.00	06/16/2025	12:15
14	Q2320-01	WC	1	Solid	NO	0.00	06/16/2025	12:22
15	Q2325-01	TP-8	1	Solid	NO	0.00	06/16/2025	12:30
16	Q2325-04	TP-8	1	Solid	NO	0.00	06/16/2025	12:37

$$\text{Burning Rate} = \frac{\text{Length (mm)}}{\text{Total Time (sec)}}$$

# WORKLIST(Hardcopy Internal Chain)

18136161

WorkList Name : IGN-061625

WorkList ID : 190200

Department : Wet-Chemistry

Date : 06-16-2025 08:28:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2301-03	WC-URBAN-FILL-C	Solid	Ignitability	Cool 4 deg C	ENTA05	D41	06/11/2025	1030
Q2307-01	LINDEN-SAA	Solid	Ignitability	Cool 4 deg C	PSEG03	D51	06/12/2025	1030
Q2310-01	TP-7	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/12/2025	1030
Q2310-04	TP-7	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/12/2025	1030
Q2311-01	TP03-MH2MH3-WC	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/11/2025	1030
Q2311-04	TP03-MH2MH3-WC	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/11/2025	1030
Q2311-05	TP04-MH2MH3-WC	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/11/2025	1030
Q2311-08	TP04-MH2MH3-WC	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/11/2025	1030
Q2312-01	TP-1	Solid	Ignitability	Cool 4 deg C	PSEG03	D51	06/13/2025	1030
Q2312-04	TP-1	Solid	Ignitability	Cool 4 deg C	PSEG03	D51	06/13/2025	1030
Q2319-01	MH-B	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/13/2025	1030
Q2319-04	MH-B	Solid	Ignitability	Cool 4 deg C	PSEG03	D41	06/13/2025	1030
Q2320-01	WC	Solid	Ignitability	Cool 4 deg C	FIRS02	D41	06/13/2025	1030
Q2325-01	TP-8	Solid	Ignitability	Cool 4 deg C	PSEG03	D51	06/13/2025	1030
Q2325-04	TP-8	Solid	Ignitability	Cool 4 deg C	PSEG03	D51	06/13/2025	1030

Date/Time 06/16/25 08:40  
Raw Sample Received by: EM (WSE)  
Raw Sample Relinquished by: CP (WSE)

Date/Time 06/16/25 12:50  
Raw Sample Received by: CP (WSE)  
Raw Sample Relinquished by: EM (WSE)

Test results

Aquakem 7.2AQ1

Page:

LB136162

CHEMTECH CONSULTING GROUP INC  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

6/16/2025 12:27

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.005	0.0	0.078	
ICB1	1.284	0.0	0.001	
CCV1	234.000	0.0	0.194	
CCB1	1.251	0.0	0.001	
PB168487BL	1.283	0.0	0.001	
Q2301-03	1.333	0.0	0.001	
Q2301-03DUP	1.234	0.0	0.001	
Q2310-04	1.246	0.0	0.001	
Q2311-04	1.284	0.0	0.001	
Q2311-08	1.161	0.0	0.001	
Q2312-04	1.438	0.0	0.001	
Q2319-04	1.209	0.0	0.001	
Q2320-01	1.388	0.0	0.001	
Q2325-04	1.186	0.0	0.001	
CCV2	230.887	0.0	0.192	
CCB2	1.336	0.0	0.001	
N	16			
Mean	35.970			
SD	80.0921			
CV%	222.66			

Aquakem v. 7.2AQ1

Results from time period:

Mon Jun 16 11:25:27 2025

Mon Jun 16 12:24:44 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Total CN	P	1.5054	µg/l	6/16/2025 11:42:50
5.OPPBCN	A	Total CN	P	5.9181	µg/l	6/16/2025 11:42:51
10PPBCN	A	Total CN	P	10.9305	µg/l	6/16/2025 11:42:52
50PPBCN	A	Total CN	P	48.7206	µg/l	6/16/2025 11:42:53
100PPBCN	A	Total CN	P	99.7388	µg/l	6/16/2025 11:42:54
250PPBCN	A	Total CN	P	246.0683	µg/l	6/16/2025 11:42:55
500PPBCN	A	Total CN	P	502.1182	µg/l	6/16/2025 11:42:56
ICV1	S	Total CN	P	94.0052	µg/l	6/16/2025 12:14:22
ICB1	S	Total CN	P	1.2836	µg/l	6/16/2025 12:14:23
CCV1	S	Total CN	P	234.0004	µg/l	6/16/2025 12:14:26
CCB1	S	Total CN	P	1.2511	µg/l	6/16/2025 12:14:27
PB168487BL	S	Total CN	P	1.2827	µg/l	6/16/2025 12:14:30
Q2301-03	S	Total CN	P	1.3335	µg/l	6/16/2025 12:14:31
Q2301-03DUP	S	Total CN	P	1.2344	µg/l	6/16/2025 12:21:56
Q2310-04	S	Total CN	P	1.2455	µg/l	6/16/2025 12:21:57
Q2311-04	S	Total CN	P	1.2835	µg/l	6/16/2025 12:21:58
Q2311-08	S	Total CN	P	1.1609	µg/l	6/16/2025 12:21:59
Q2312-04	S	Total CN	P	1.4376	µg/l	6/16/2025 12:22:00
Q2319-04	S	Total CN	P	1.2092	µg/l	6/16/2025 12:22:01
Q2320-01	S	Total CN	P	1.3884	µg/l	6/16/2025 12:22:02
Q2325-04	S	Total CN	P	1.1863	µg/l	6/16/2025 12:22:03
CCV2	S	Total CN	P	230.8865	µg/l	6/16/2025 12:22:06
CCB2	S	Total CN	P	1.3365	µg/l	6/16/2025 12:24:44



Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

6/16/2025 11:43


Test Total CN

Accepted 6/16/2025 11:43

Factor 1204  
Bias 0

Coeff. of det. 0.999875

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	1.5054	0.0000	-
2	5.0PPBCN	0.005	5.9181	5.0000	18.4
3	10PPBCN	0.009	10.9305	10.0000	9.3
4	50PPBCN	0.040	48.7206	50.0000	-2.6
5	100PPBCN	0.083	99.7388	100.0000	-0.3
6	250PPBCN	0.204	246.0683	250.0000	-1.6
7	500PPBCN	0.417	502.1182	500.0000	0.4

12  
6/16/25

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB136168

ANALYST: Iwona

SUPERVISOR REVIEW BY: jignesh

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3213
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB168488BL		1	5.05	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	06/16/2025	13:15
2	Q2301-03		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.06	4.77	06/16/2025	13:18
3	Q2301-03DUP		1	5.01	50	2.00	0.00	1.88	1.88	0.12	0.06	4.79	06/16/2025	13:21
4	Q2310-04		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	06/16/2025	13:25
5	Q2311-04		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.08	6.32	06/16/2025	13:28
6	Q2311-08		1	5.02	50	2.00	0.00	1.88	1.88	0.12	0.06	4.78	06/16/2025	13:31
7	Q2312-04		1	5.03	50	2.00	0.00	1.86	1.86	0.14	0.08	6.36	06/16/2025	13:34
8	Q2319-04		1	5.04	50	2.00	0.00	1.92	1.92	0.08	0.02	1.59	06/16/2025	13:37
9	Q2320-01		1	5.07	50	2.00	0.00	1.90	1.90	0.10	0.04	3.16	06/16/2025	13:40
10	Q2325-04		1	5.01	50	2.00	0.00	1.88	1.88	0.12	0.06	4.79	06/16/2025	13:44

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136178

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/17/2025	15:15
2	CAL2	1	Water	NA	NA	20.2	7.00	06/17/2025	15:16
3	CAL3	1	Water	NA	NA	20.2	10.02	06/17/2025	15:16
4	ICV	1	Water	NA	NA	20.3	7.00	06/17/2025	15:19
5	CCV1	1	Water	NA	NA	20.3	2.02	06/17/2025	15:20
6	Q2311-04	1	Solid	20.02	20	20.2	9.17	06/17/2025	15:30
7	Q2311-08	1	Solid	20.03	20	20.1	10.23	06/17/2025	15:35
8	Q2312-04	1	Solid	20.02	20	20.6	8.13	06/17/2025	15:40
9	Q2319-04	1	Solid	20.03	20	20.3	5.76	06/17/2025	15:44
10	Q2320-01	1	Solid	20.04	20	20.9	7.80	06/17/2025	15:50
11	Q2325-04	1	Solid	20.02	20	20.2	6.61	06/17/2025	16:00
12	Q2339-04	1	Solid	20.03	20	20.7	6.20	06/17/2025	16:10
13	Q2340-04	1	Solid	20.04	20	20.4	8.92	06/17/2025	16:11
14	Q2341-04	1	Solid	20.03	20	20.6	8.63	06/17/2025	16:20
15	Q2341-08	1	Solid	20.04	20	20.5	7.04	06/17/2025	16:30
16	CCV2	1	Water	NA	NA	20.2	12.02	06/17/2025	16:35
17	Q2347-04	1	Solid	20.02	20	20.2	7.40	06/17/2025	16:45
18	Q2347-04DUP	1	Solid	20.03	20	20.3	7.41	06/17/2025	16:47
19	CCV3	1	Water	NA	NA	20.3	2.02	06/17/2025	16:50

# WORKLIST(Hardcopy Internal Chain)

136178

WorkList Name : corrosivity Q2347      WorkList ID : 190244      Department : Wet-Chemistry      Date : 06-17-2025 14:40:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2311-04	TP03-MH2MH3-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/11/2025	9045D
Q2311-08	TP04-MH2MH3-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/11/2025	9045D
Q2312-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	D51	06/13/2025	9045D
Q2319-04	MH-B	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/13/2025	9045D
Q2320-01	WC	Solid	Corrosivity	Cool 4 deg C	FIRS02	D41	06/13/2025	9045D
Q2325-04	TP-8	Solid	Corrosivity	Cool 4 deg C	PSEG03	D51	06/13/2025	9045D
Q2339-04	MH-C	Solid	Corrosivity	Cool 4 deg C	PSEG03	D51	06/16/2025	9045D
Q2340-04	TP05-MH17A-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/13/2025	9045D
Q2341-04	TP-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	D52	06/16/2025	9045D
Q2341-08	EP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	D52	06/16/2025	9045D
Q2347-04	MH-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	D42	06/17/2025	9045D

Date/Time 06/17/25 15:00  
Raw Sample Received by: SA WNC  
Raw Sample Relinquished by: CP SM

Date/Time 06/17/25 18:00  
Raw Sample Received by: CP SM  
Raw Sample Relinquished by: SA WNC

SOP ID : M9012B-Total, Amenable and Reactive Cyanide-20

SDG No : N/A

Start Digest Date: 06/16/2025 Time : 08:30 Temp : N/A

Matrix : SOIL

End Digest Date: 06/16/2025 Time : 10:00 Temp : N/A

Pipette ID : N/A

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : N/A

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: EM

Weigh By : JP

pH Meter ID : N/A

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP111294
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Comment

## Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/16/25 10:15	EM (WC)	12 (JC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/Nitrite	Comment	Prep Pos
PB168487BL	PB168487BL	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2301-03DUP	WC-URBAN-FILL-CDUP	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2301-03	WC-URBAN-FILL-C	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2310-04	TP-7	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2311-04	TP03-MH2MH3-WC	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2311-08	TP04-MH2MH3-WC	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2312-04	TP-1	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2319-04	MH-B	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2320-01	WC	5.08	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2325-04	TP-8	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCN-061625

WorkList ID : 190217

Department : Distillation

Date : 06-16-2025 08:10:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2301-03	WC-URBAN-FILL-C	Solid	Reactive Cyanide	Cool 4 deg C	ENTA05	D41	06/11/2025	9012B
Q2310-04	TP-7	Solid	Reactive Cyanide	Cool 4 deg C	PSEG03	D41	06/12/2025	9012B
Q2311-04	TP03-MH2MH3-WC	Solid	Reactive Cyanide	Cool 4 deg C	PSEG03	D41	06/11/2025	9012B
Q2311-08	TP04-MH2MH3-WC	Solid	Reactive Cyanide	Cool 4 deg C	PSEG03	D41	06/11/2025	9012B
Q2312-04	TP-1	Solid	Reactive Cyanide	Cool 4 deg C	PSEG03	D51	06/13/2025	9012B
Q2319-04	MH-B	Solid	Reactive Cyanide	Cool 4 deg C	PSEG03	D41	06/13/2025	9012B
Q2320-01	WC	Solid	Reactive Cyanide	Cool 4 deg C	FIRS02	D41	06/13/2025	9012B
Q2325-04	TP-8	Solid	Reactive Cyanide	Cool 4 deg C	PSEG03	D51	06/13/2025	9012B

Date/Time 06/16/25 08:15  
 Raw Sample Received by: EM (VSC)  
 Raw Sample Relinquished by: CP PM

Date/Time 06/16/25 09:00  
 Raw Sample Received by: CP PM  
 Raw Sample Relinquished by: EM (VSC)

SOP ID : M9030B-Sulfide-12

SDG No : N/A

Start Digest Date: 06/16/2025 Time : 08:30 Temp : N/A

Matrix : SOIL

End Digest Date: 06/16/2025 Time : 10:00 Temp : N/A

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : N/A

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: EM

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.5M ZINC ACETATE	5.0ML	WP113086
FORMALDEHYDE	2.0ML	W2725
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
	Preparation Group	Analysis Group

6/16/25



Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/Nitrite	Comment	Prep Pos
PB168488BL	PB168488BL	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2301-03DUP	WC-URBAN-FILL-CDUP	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2301-03	WC-URBAN-FILL-C	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2310-04	TP-7	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2311-04	TP03-MH2MH3-WC	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2311-08	TP04-MH2MH3-WC	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2312-04	TP-1	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2319-04	MH-B	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2320-01	WC	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2325-04	TP-8	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A

**Instrument ID:** FLAME

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136161**

Review By	Eman	Review On	6/16/2025 1:49:13 PM
Supervise By	Iwona	Supervise On	6/16/2025 1:50:22 PM
SubDirectory	LB136161	Test	Ignitability
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	Q2301-03	WC-URBAN-FILL-C	SAM	06/16/25 10:45		Eman	OK
2	Q2301-03DUP	WC-URBAN-FILL-CD	DUP	06/16/25 10:52		Eman	OK
3	Q2307-01	LINDEN-SAA	SAM	06/16/25 11:00		Eman	OK
4	Q2310-01	TP-7	SAM	06/16/25 11:07		Eman	OK
5	Q2310-04	TP-7	SAM	06/16/25 11:14		Eman	OK
6	Q2311-01	TP03-MH2MH3-WC	SAM	06/16/25 11:22		Eman	OK
7	Q2311-04	TP03-MH2MH3-WC	SAM	06/16/25 11:30		Eman	OK
8	Q2311-05	TP04-MH2MH3-WC	SAM	06/16/25 11:37		Eman	OK
9	Q2311-08	TP04-MH2MH3-WC	SAM	06/16/25 11:45		Eman	OK
10	Q2312-01	TP-1	SAM	06/16/25 11:52		Eman	OK
11	Q2312-04	TP-1	SAM	06/16/25 12:00		Eman	OK
12	Q2319-01	MH-B	SAM	06/16/25 12:07		Eman	OK
13	Q2319-04	MH-B	SAM	06/16/25 12:15		Eman	OK
14	Q2320-01	WC	SAM	06/16/25 12:22		Eman	OK
15	Q2325-01	TP-8	SAM	06/16/25 12:30		Eman	OK
16	Q2325-04	TP-8	SAM	06/16/25 12:37		Eman	OK

**Instrument ID:** KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136162**

Review By	Iwona	Review On	6/17/2025 10:27:14 AM
Supervise By	Sohil	Supervise On	6/17/2025 10:35:41 AM
SubDirectory	LB136162	Test	Reactive Cyanide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	WP113536,WP113537,WP113538,WP113540,WP113541,WP113542,WP113543		
ICV Standard	WP113544		
CCV Standard	WP113538		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP112643,WP112900,WP113548		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	06/16/25 11:42		Iwona	OK
2	5.0PPBCN	5.0PPBCN	CAL2	06/16/25 11:42		Iwona	OK
3	10PPBCN	10PPBCN	CAL3	06/16/25 11:42		Iwona	OK
4	50PPBCN	50PPBCN	CAL4	06/16/25 11:42		Iwona	OK
5	100PPBCN	100PPBCN	CAL5	06/16/25 11:42		Iwona	OK
6	250PPBCN	250PPBCN	CAL6	06/16/25 11:42		Iwona	OK
7	500PPBCN	500PPBCN	CAL7	06/16/25 11:42		Iwona	OK
8	ICV1	ICV1	ICV	06/16/25 12:14		Iwona	OK
9	ICB1	ICB1	ICB	06/16/25 12:14		Iwona	OK
10	CCV1	CCV1	CCV	06/16/25 12:14		Iwona	OK
11	CCB1	CCB1	CCB	06/16/25 12:14		Iwona	OK
12	PB168487BL	PB168487BL	MB	06/16/25 12:14		Iwona	OK
13	Q2301-03	WC-URBAN-FILL-C	SAM	06/16/25 12:14		Iwona	OK
14	Q2301-03DUP	WC-URBAN-FILL-CD	DUP	06/16/25 12:21		Iwona	OK
15	Q2310-04	TP-7	SAM	06/16/25 12:21		Iwona	OK
16	Q2311-04	TP03-MH2MH3-WC	SAM	06/16/25 12:21		Iwona	OK
17	Q2311-08	TP04-MH2MH3-WC	SAM	06/16/25 12:21		Iwona	OK
18	Q2312-04	TP-1	SAM	06/16/25 12:22		Iwona	OK

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136162**

Review By	Iwona	Review On	6/17/2025 10:27:14 AM
Supervise By	Sohil	Supervise On	6/17/2025 10:35:41 AM
SubDirectory	LB136162	Test	Reactive Cyanide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	WP113536,WP113537,WP113538,WP113540,WP113541,WP113542,WP113543		
ICV Standard	WP113544		
CCV Standard	WP113538		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP112643,WP112900,WP113548		

19	Q2319-04	MH-B	SAM	06/16/25 12:22		Iwona	OK
20	Q2320-01	WC	SAM	06/16/25 12:22		Iwona	OK
21	Q2325-04	TP-8	SAM	06/16/25 12:22		Iwona	OK
22	CCV2	CCV2	CCV	06/16/25 12:22		Iwona	OK
23	CCB2	CCB2	CCB	06/16/25 12:24		Iwona	OK

**Instrument ID:** TITRAMETRIC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136168**

Review By	Iwona	Review On	6/17/2025 10:02:45 AM
Supervise By	jignesh	Supervise On	6/17/2025 10:07:25 AM
SubDirectory	LB136168	Test	Reactive Sulfide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3105,W3213,W3149		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	PB168488BL	PB168488BL	MB	06/16/25 13:15		Iwona	OK
2	Q2301-03	WC-URBAN-FILL-C	SAM	06/16/25 13:18		Iwona	OK
3	Q2301-03DUP	WC-URBAN-FILL-CD	DUP	06/16/25 13:21		Iwona	OK
4	Q2310-04	TP-7	SAM	06/16/25 13:25		Iwona	OK
5	Q2311-04	TP03-MH2MH3-WC	SAM	06/16/25 13:28		Iwona	OK
6	Q2311-08	TP04-MH2MH3-WC	SAM	06/16/25 13:31		Iwona	OK
7	Q2312-04	TP-1	SAM	06/16/25 13:34		Iwona	OK
8	Q2319-04	MH-B	SAM	06/16/25 13:37		Iwona	OK
9	Q2320-01	WC	SAM	06/16/25 13:40		Iwona	OK
10	Q2325-04	TP-8	SAM	06/16/25 13:44		Iwona	OK

**Instrument ID:** WC PH METER-1

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136178**

Review By	jignesh	Review On	6/18/2025 8:57:10 AM
Supervise By	Iwona	Supervise On	6/18/2025 9:44:38 AM
SubDirectory	LB136178	Test	Corrosivity
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3071,W3161,W3200		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	06/17/25 15:15		Jignesh	OK
2	CAL2	CAL2	CAL	06/17/25 15:16		Jignesh	OK
3	CAL3	CAL3	CAL	06/17/25 15:16		Jignesh	OK
4	ICV	ICV	ICV	06/17/25 15:19		Jignesh	OK
5	CCV1	CCV1	CCV	06/17/25 15:20		Jignesh	OK
6	Q2311-04	TP03-MH2MH3-WC	SAM	06/17/25 15:30		Jignesh	OK
7	Q2311-08	TP04-MH2MH3-WC	SAM	06/17/25 15:35		Jignesh	OK
8	Q2312-04	TP-1	SAM	06/17/25 15:40		Jignesh	OK
9	Q2319-04	MH-B	SAM	06/17/25 15:44		Jignesh	OK
10	Q2320-01	WC	SAM	06/17/25 15:50		Jignesh	OK
11	Q2325-04	TP-8	SAM	06/17/25 16:00		Jignesh	OK
12	Q2339-04	MH-C	SAM	06/17/25 16:10		Jignesh	OK
13	Q2340-04	TP05-MH17A-WC	SAM	06/17/25 16:11		Jignesh	OK
14	Q2341-04	TP-9	SAM	06/17/25 16:20		Jignesh	OK
15	Q2341-08	EP-3	SAM	06/17/25 16:30		Jignesh	OK
16	CCV2	CCV2	CCV	06/17/25 16:35		Jignesh	OK
17	Q2347-04	TP-10-9	SAM	06/17/25 16:45		Jignesh	OK
18	Q2347-04DUP	TP-10-9	DUP	06/17/25 16:47		Jignesh	OK

**Instrument ID:** WC PH METER-1

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136178**

Review By	jignesh	Review On	6/18/2025 8:57:10 AM
Supervise By	Iwona	Supervise On	6/18/2025 9:44:38 AM
SubDirectory	LB136178	Test	Corrosivity
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3071,W3161,W3200		

19	CCV3	CCV3	CCV	06/17/25 16:50		Jignesh	OK
----	------	------	-----	----------------	--	---------	----

## Prep Standard - Chemical Standard Summary

**Order ID :** Q2320

**Test :** Corrosivity,Ignitability,Percent Solids,Reactive Cyanide,Reactive Sulfide

**Prepbatch ID :** PB168487,PB168488,

**Sequence ID/Qc Batch ID:** LB136161,LB136162,LB136168,LB136178,

**Standard ID :**

WP111294,WP112643,WP112900,WP112995,WP113086,WP113535,WP113536,WP113537,WP113538,WP113540,WP113541,WP113542,WP113543,WP113544,WP113548,

**Chemical ID :**

M6151,W2668,W2725,W2926,W3019,W3071,W3093,W3105,W3112,W3113,W3139,W3149,W3161,W3173,W3178,W3191,W3200,W3203,W3213,W3214,





<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
11	Sodium hydroxide absorbing solution 0.25 N	<a href="#">WP111294</a>	01/07/2025	07/07/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WC-SC-5)	None	Iwona Zarych 01/07/2025
<b><u>FROM</u></b>	21.00000L of W3112 + 210.00000gram of W3113 = Final Quantity: 21.000 L							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
539	CN BUFFER	<a href="#">WP112643</a>	04/09/2025	10/09/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WCS-5)	None	Iwona Zarych 04/09/2025
<b><u>FROM</u></b>	138.00000gram of W2668 + 862.00000ml of W3112 = Final Quantity: 1000.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
607	PYRIDINE-BARBITURIC ACID	<a href="#">WP112900</a>	05/01/2025	08/18/2025	Rubina Mughal	WETCHEM_SCALE_8 (WCS-7)	Glass Pipette-A	Iwona Zarych 05/01/2025
<u>FROM</u>	145.00000ml of W3112 + 15.00000gram of W3203 + 15.00000ml of M6151 + 75.00000ml of W3019 = Final Quantity: 250.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3371	Cyanide LCS Spike Solution, 5PPM	<a href="#">WP112995</a>	05/07/2025	07/07/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 05/07/2025
<u>FROM</u>	1.00000ml of W3173 + 199.00000ml of WP111294 = Final Quantity: 200.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
160	0.5M ZINC ACETATE	<a href="#">WP113086</a>	05/15/2025	08/18/2025	Rubina Mughal	WETCHEM_SCALE_8 (WCS-7)	None	Iwona Zarych 05/15/2025
<b><u>FROM</u></b> 0.88900L of W3112 + 1.00000ml of M6151 + 110.00000gram of W2926 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3456	Cyanide Intermediate Working Std, 5PPM	<a href="#">WP113535</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 06/17/2025
<b><u>FROM</u></b> 0.25000ml of W3214 + 49.75000ml of WP111294 = Final Quantity: 50.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
167	0 ppb CN calibration std	<a href="#">WP113536</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 06/17/2025
<b><u>FROM</u></b> 50.00000ml of WP111294 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4	Calibration standard 500 ppb	<a href="#">WP113537</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh
<p>(WC)</p> <p><b>FROM</b> 45.00000ml of WP111294 + 5.00000ml of WP113535 = Final Quantity: 50.000 ml</p>								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3761	Calibration-CCV CN Standard 250 ppb	<a href="#">WP113538</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh
(WC)								
<u>FROM</u>	2.50000ml of WP113535 + 47.50000ml of WP111294 = Final Quantity: 50.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
6	Calibration Standard 100 ppb	<a href="#">WP113540</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 06/17/2025
<u>FROM</u>	1.00000ml of WP113535 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
7	Calibration Standard 50 ppb	<a href="#">WP113541</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 06/17/2025
<b><u>FROM</u></b>	0.50000ml of WP113535 + 49.50000ml of WP111294 = Final Quantity: 50.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
8	Calibration Standard 10 ppb	<a href="#">WP113542</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 06/17/2025
<u>FROM</u>	1.00000ml of WP113537 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
9	Calibration Standard 5 ppb	<a href="#">WP113543</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh
<b>FROM</b> 0.50000ml of WP113537 + 49.50000ml of WP111294 = Final Quantity: 50.000 ml <div></div>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2168	RCN ICV STD, 100 PPB	<a href="#">WP113544</a>	06/16/2025	06/17/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh
<b>FROM</b> 1.00000ml of WP112995 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml <div></div>								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1582	Chloramine T solution, 0.014M	<a href="#">WP113548</a>	06/16/2025	06/17/2025	Iwona Zarych	WETCHEM_SCALE_5 (WCS-5)	None	Jignesh Parikh 06/17/2025
<b><u>FROM</u></b> 0.08000gram of W3139 + 20.00000ml of W3112 = Final Quantity: 20.000 ml								



## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	22G2862015	08/18/2025	02/18/2025 / Sagar	01/15/2025 / Sagar	M6151

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3818-5 / SODIUM PHOSPHATE, MONOBAS/HYD, CRYST, ACS, 2.5 KG	0000225799	12/03/2025	04/05/2021 / Alexander	02/10/2020 / apatel	W2668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EMD-FX0410-5 / FORMALDEHYDE SOLUTION 450ML	60045	06/22/2025	08/19/2024 / lwona	06/22/2020 / apatel	W2725

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J4296-1 / ZINC ACETATE, DIHYD, CRYST, ACS, 500G	383058	07/05/2027	07/05/2022 / ketankumar	07/05/2022 / ketankumar	W2926

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
SIGMA ALDRICH	270970-1L / Pyridine 1L	SHBQ2113	04/03/2028	04/03/2023 / lwona	04/03/2023 / lwona	W3019

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14455-3 / buffer solution pH 7 yellow	4308H30	07/31/2025	01/02/2024 / JIGNESH	12/06/2023 / lwona	W3071

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	566002 / BUFFER PH 7.00 GREEN 1PINT PK6	44001f99	12/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3093

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE	4403S13	09/30/2025	04/22/2024 / lwona	04/22/2024 / lwona	W3105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-7 / Sodium Hydroxide Pellets 12 Kg	23B1556310	12/31/2025	07/08/2024 / lwona	07/08/2024 / lwona	W3113

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	JTE494-6 / CHLORAMINE-T BAKER 250GM	10239484	09/09/2029	09/09/2024 / lwona	09/09/2024 / lwona	W3139

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL70850-8 / Starch Solution, 4L	4408P62	08/31/2026	10/16/2024 / lwona	10/16/2024 / lwona	W3149

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL13850-1 / Buffer Solution, PH2 (500ml)	2411E26	10/31/2026	12/09/2024 / lwona	12/09/2024 / lwona	W3161

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	LC135457 / Cyanide Standard, 1000 PPM, Second Source	45010168	07/17/2025	01/24/2025 / lwona	01/24/2025 / lwona	W3173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14055-3 / PH 4 BUFFER SOLUTION	2411A93	10/30/2026	04/01/2025 / JIGNESH	01/27/2025 / jignesh	W3178

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	1601-1 / PH 10.01 BUFFER,COLOR CD 475ML	2410F80	03/31/2026	04/01/2025 / JIGNESH	03/13/2025 / jignesh	W3191

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
RICCA CHEMICAL COMPANY	1615-16 / pH 12.00 Buffer	2504F20	09/30/2026	04/11/2025 / lwona	04/11/2025 / lwona	W3200

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EM-BX0035-3 / Barbituric Acid, 100 gms	WXBF3271V	05/16/2029	04/21/2025 / lwona	04/21/2025 / lwona	W3203

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL35830-4 / IODINE SOLUTION .025N 1L	MK25A21527	01/20/2029	05/21/2025 / Iwona	05/21/2025 / Iwona	W3213

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	RC2543-4 / CYANIDE STD 1000PPM 4OZ	1505H73	11/30/2025	05/21/2025 / Iwona	05/21/2025 / Iwona	W3214


W3071  
Rec 12/6/23

## Certificate of Analysis 12

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4308H30

Product Number: 1551

Manufacture Date: AUG 09, 2023

Expiration Date: JUL 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Yellow liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.002	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (08/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W3019  
rec 4/3/23

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: [www.sigmaaldrich.com](http://www.sigmaaldrich.com)Email USA: [techserv@sial.com](mailto:techserv@sial.com)Outside USA: [eurtechserv@sial.com](mailto:eurtechserv@sial.com)

## Certificate of Analysis

Product Name:

Pyridine - anhydrous, 99.8%

Product Number:

270970

Batch Number:

SHBQ2113

Brand:

SIAL

CAS Number:

110-86-1

MDL Number:

MFCD00011732

Formula:


C<sub>5</sub>H<sub>5</sub>N

Formula Weight:

79.10 g/mol

Quality Release Date:

15 DEC 2022



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Purity (GC)	≥ 99.75 %	99.99 %
Water (by Karl Fischer)	≤ 0.003 %	0.002 %
Residue on Evaporation	≤ 0.0005 %	< 0.0001 %

  
Larry Coers, Director  
Quality Control  
Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





## Certificate of Analysis

Date of Release: 2/26/2020

Name: Formaldehyde Solution  
GR ACS  
Meets ACS Specifications

Item No: FX0410 all size codes

Lot / Batch No: 60045

Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Assay	36.5	38.0	36.71	%
Chloride (Cl)		5	<5	ppm
Color (APHA)		10	<10	
Form			Passes test	
Heavy metals (as Pb)		5	<5	ppm
Iron (Fe)		5	0.6	ppm
Residue after ignition		0.005	<0.0050	%
Sulfate (SO <sub>4</sub> )		0.002	<0.0020	%
Titration acid		0.006	<0.0060	meq/g

Heather Sinn,

-----  
Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Millipore Corporation, an affiliate of Merck KGaA, Darmstadt, Germany  
290 Concord Road  
Billerica, MA 01821  
U.S.A

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.



Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantor™**



M6151

R → 11/15/25

Material No.: 9530-33  
Batch No.: 22G2862015  
Manufactured Date: 2022-06-15  
Retest Date: 2027-06-14  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.9 %
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.191
ACS – Bromide (Br)	≤ 0.005 %	< 0.005 %
ACS – Extractable Organic Substances	≤ 5 ppm	< 1 ppm
ACS – Free Chlorine (as Cl <sub>2</sub> )	≤ 0.5 ppm	< 0.5 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.05 ppm	< 0.03 ppm
Sulfate (SO <sub>4</sub> )	≤ 0.5 ppm	< 0.3 ppm
Sulfite (SO <sub>3</sub> )	≤ 0.8 ppm	0.3 ppm
Ammonium (NH <sub>4</sub> )	≤ 3 ppm	< 1 ppm
Trace Impurities – Arsenic (As)	≤ 0.010 ppm	< 0.003 ppm
Trace Impurities – Aluminum (Al)	≤ 10.0 ppb	1.3 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 3.0 ppb
Trace Impurities – Barium (Ba)	≤ 1.0 ppb	0.2 ppb
Trace Impurities – Beryllium (Be)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Bismuth (Bi)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 20.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	163.0 ppb
Trace Impurities – Chromium (Cr)	≤ 1.0 ppb	0.7 ppb
Trace Impurities – Cobalt (Co)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gallium (Ga)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0 ppb
Trace Impurities – Gold (Au)	≤ 4.0 ppb	0.6 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 15 ppb	6 ppb

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantorsm**



Material No.: 9530-33  
Batch No.: 22G2862015

Test	Specification	Result
Trace Impurities – Lead (Pb)	≤ 1.0 ppb	< 0.5 ppb
Trace Impurities – Lithium (Li)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Magnesium (Mg)	≤ 10.0 ppb	2.9 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	0.1 ppb
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 3.0 ppb
Trace Impurities – Nickel (Ni)	≤ 4.0 ppb	< 0.3 ppb
Trace Impurities – Niobium (Nb)	≤ 1.0 ppb	0.8 ppb
Trace Impurities – Potassium (K)	≤ 9.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se), For Information Only		< 1.0 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	0.5 ppb
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	2.3 ppb
Trace Impurities – Strontium (Sr)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Tantalum (Ta)	≤ 1.0 ppb	1.6 ppb
Trace Impurities – Thallium (Tl)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	4.0 ppb
Trace Impurities – Titanium (Ti)	≤ 1.0 ppb	1.5 ppb
Trace Impurities – Vanadium (V)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.8 ppb
Trace Impurities – Zirconium (Zr)	≤ 1.0 ppb	0.3 ppb

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA–ANALYZED® Reagent  
For Trace Metal Analysis



Material No.: 9530-33  
Batch No.: 22G2862015

Test	Specification	Result
------	---------------	--------

For Laboratory, Research, or Manufacturing Use  
Product Information (not specifications):  
Appearance (clear, fuming liquid)  
Meets ACS Specifications  
Storage Condition: Store below 25 °C.

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.  
Jamie Ethier  
Vice President Global Quality

Sodium Phosphate, Monobasic, Monohydrate,  
Crystal  
BAKER ANALYZED® A.C.S. Reagent

(sodium dihydrogen phosphate, monohydrate)



Material No.: 3818-05  
Batch No.: 0000225799  
Manufactured Date: 2018/12/05  
Retest Date: 2025/12/03  
Revision No: 1

## Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay ( $\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$ )	98.0 – 102.0 %	99.5
pH of 5% Solution at 25°C	4.1 – 4.5	4.3
Insoluble Matter	$\leq 0.01$ %	$< 0.01$
Chloride (Cl)	$\leq 5$ ppm	$< 5$
ACS – Sulfate ( $\text{SO}_4$ )	$\leq 0.003$ %	$< 0.003$
Calcium (Ca)	$\leq 0.005$ %	$< 0.005$
Potassium (K)	$\leq 0.01$ %	$< 0.01$
Heavy Metals (as Pb)	$\leq 0.001$ %	$< 0.001$
Trace Impurities – Iron (Fe)	$\leq 0.001$ %	$< 0.001$

For Laboratory, Research or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: IN  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality


For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

## Certificate of Analysis

Product Name:


Zinc acetate dihydrate - ACS reagent,  $\geq 98\%$ 

Product Number: 383058  
Batch Number: MKCQ9159  
Brand: SIGALD  
CAS Number: 5970-45-6  
MDL Number: MFCD00066961  
Formula:  $C_4H_6O_4Zn \cdot 2H_2O$   
Formula Weight: 219.51 g/mol  
Quality Release Date: 06 JAN 2022



W2926  
open 7/5/22  
received  
on 7/5/22

Test	Specification	Result
Appearance (Color)	White	White
Appearance (Form)	Powder or Crystal or Chunk(s)	Powder
Infrared Spectrum	Conforms to Structure	Conforms
Insoluble Matter	$\leq 0.005\%$	0.003 %
Calcium (Ca)	$\leq 0.005\%$	0.003 %
Chloride (Cl)	$\leq 5$ ppm	$< 5$ ppm
Iron (Fe)	$\leq 5$ ppm	$< 5$ ppm
Potassium (K)	$\leq 0.01\%$	0.00 %
Magnesium (Mg)	$\leq 0.005\%$	0.003 %
Sodium (Na)	$\leq 0.05\%$	0.03 %
Lead (Pb)	$\leq 0.002\%$	$< 0.001\%$
pH	6.0 - 7.0	6.1
Sulfate (SO <sub>4</sub> )	$\leq 0.005\%$	$< 0.005\%$
Complexometric EDTA	98.0 - 101.0 %	100.3 %
Meets ACS Requirements	Meets Requirements	Meets Requirements

  
Larry Coers, Director  
Quality Control  
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





## Certificate of Analysis

W3093  
094121  
04/03/2024  
16

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4401F99

Product Number: 1551

Manufacture Date: JAN 08, 2024

Expiration Date: DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	

Test	Specification	Result
Appearance	Yellow liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.004	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-1	4 L natural poly	24 months
1551-1CT	4 L Cubitainer®	24 months
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

# Certificate of Analysis

## Sodium Thiosulfate, 0.0250 Normal (N/40)

**Lot Number:** 4403S13

**Product Number:** 7900

**Manufacture Date:** MAR 29, 2024

**Expiration Date:** SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Thiosulfate Pentahydrate	10102-17-7	ACS
Organic Preservative	Proprietary	
Sodium Carbonate	497-19-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.02499-0.02501 N at 20°C	0.02501 N at 20°C	136

Specification	Reference
Standard Sodium Thiosulfate Solution, 0.0250 N	APHA (4500-S2- F)
Standard Sodium Thiosulfate Titrant	APHA (4500-O D)
Standard Sodium Thiosulfate Titrant	APHA (4500-O E)
Standard Sodium Thiosulfate Titrant	APHA (4500-O F)
Standard Sodium Thiosulfate Titrant, 0.025 N	APHA (4500-CI B)
Standard Sodium Thiosulfate Titrant	APHA (4500-O C)
Standard Sodium Thiosulfate Titrant, 0.025 M	APHA (5530 C)
Standard Sodium Thiosulfate Solution (0.025 N)	EPA (SW-846) (9031)
Standard Sodium Thiosulfate solution (0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7900-1	4 L natural poly	18 months
7900-16	500 mL natural poly	18 months
7900-1CT	4 L Cubitainer®	18 months
7900-32	1 L natural poly	18 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)





Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# Certificate of Analysis



## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

Chemical Formula: NaOH  
Molecular Weight: 40  
CAS #: 1310-73-2  
Appearance:

Manufacture Date: 12/14/2022  
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

Chemical Formula: NaOH  
Molecular Weight: 40  
CAS #: 1310-73-2  
Appearance:

Manufacture Date: 12/14/2022  
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

W3139 Received on 9/9/24 by IZ

Product No.: A12044  
Product: Chloramine-T trihydrate, 98%  
Lot No.: 10239484

Appearance:	White powder
Melting Point:	166°C(dec)
Assay (Iodometric titration):	100.5%
Identification (FTIR):	Conforms

Order our products online [thermofisher.com/chemicals](https://thermofisher.com/chemicals)

**This document has been electronically generated and does not require a signature.**

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.



# Certificate of Analysis

## Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Starch, soluble	9005-84-9	ACS
Salicylic Acid	69-72-7	ACS

Test	Specification	Result
Appearance	White translucent liquid	Passed
Suitability for Use	Colorless (Iodine absent) - Blue (Iodine present)	Passed

Specification	Reference
Starch Solution	APHA (4500-S2- F)
Starch Indicator Solution	APHA (4500-CI B)
Starch Indicator	APHA (4500-SO32- B)
Starch indicator solution	APHA (2350 B)
Starch indicator solution	APHA (2350 E)
Starch Solution	APHA (510 B)
Starch Solution	APHA (5530 C)
Starch Indicator	APHA (4500-CI C)
Starch Indicator	EPA (345.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8000-1	4 L natural poly	24 months
8000-16	500 mL natural poly	24 months
8000-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)  
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# Certificate of Analysis

**Buffer, Reference Standard, pH 2.00 ± 0.01 at 25°C****Lot Number:** 2411E26**Product Number:** 1493**Manufacture Date:** NOV 11, 2024**Expiration Date:** OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	10	15	20	25	30	35	40	45	50
pH	1.93	1.98	1.98	2.00	2.01	2.03	2.03	2.04	2.04

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	1.994	0.02	185i, 186-I-g, 186-II-g

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1493-1	4 L natural poly	24 months
1493-16	500 mL natural poly	24 months
1493-1CT	4 L Cubitainer®	24 months
1493-2.5	10 L Cubitainer®	24 months
1493-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (11/11/2024)  
Operations Manager

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.





Part of TCP Analytical Group

Jackson's Pointe Commerce Park- Building 1000  
1010 Jackson's Pointe Court, Zelienople, PA 16063

## Certificate of Analysis

### Cyanide Standard 1000 ppm (1ml = 1mg CN)

Product Code: **LC13545**

Manufacture Date: January 16, 2025

Lot Number: **45010168**

Expiration Date: July 17, 2025

Test	Specification	Result
Appearance (clarity)	clear solution	clear solution
Appearance (color)	colorless	colorless
Concentration (CN)	0.990 - 1.010mg/mL	1.000mg/mL
Concentration (CN)	990 - 1,010ppm	1,000ppm
Traceable to NIST SRM	Report	999b

**Intended Use** - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

**Storage Information** - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

**Instructions for Handling and Use** - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

**Preparation** - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST

\*The suffix of the product code may differ from what is on your product label. The suffix will designate the size and be associated with a numeric digit(s). Visit [LabChem.com](http://LabChem.com) for more information\*

Suffix	1	2	3/3S/36/36S	4/4C	5	6	7	8	9	20	44	200	246	486
Size	500mL org	1L or 1kg	2.5L/2.5L Coated/6x2.5L/6x2.5L Coated	4L	20L	10L	125mL	25g	100g	20x20mL	4x4L	200L	24x6mL	48x6mL

*Michael Monteleone*

Michael Monteleone  
Chemistry Supervisor - Quality Control  
2025011610:36:11bsturges-0-0

ISO9001:2015 Registration #0306-01

**RICCA CHEMICAL COMPANY®**

1841 Broad Street  
Pocomoke City, MD 21851  
<http://www.riccachemical.com>  
1-888-GO-RICCA  
[customerservice@riccachemical.com](mailto:customerservice@riccachemical.com)

# Certificate of Analysis

W21758 58

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 2411A93

Product Number: 1501

Manufacture Date: NOV 04, 2024

Expiration Date: OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	4.00	4.00	4.00	4.00	4.00	4.00	4.01	4.02	4.03	4.04	4.06

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Preservative	Proprietary	Commercial
Red Dye	Proprietary	Purified

Test	Specification	Result
Appearance	Red liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	4.008	0.02	185i, 186-I-g, 186-II-g

Specification	Reference
Commercial Buffer Solutions	
Buffer B	ASTM (D 1293 B)
Buffer B	ASTM (D 5464)
Buffer B	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1501-16	500 mL natural poly	24 months
1501-2.5	10 L Cubitainer®	24 months
1501-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



RICCA CHEMICAL COMPANY®

1841 Broad Street  
Pocomoke City, MD 21851  
<http://www.riccachemical.com>  
1-888-GO-RICCA  
[customerservice@riccachemical.com](mailto:customerservice@riccachemical.com)

Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 2410F80

Product Number: 1601

Manufacture Date: OCT 09, 2024

Expiration Date: MAR 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.  
The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	50
pH	10.31	10.23	10.17	10.11	10.05	10.00	9.95	9.91	9.87	9.81

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	Reagent
Preservative	Proprietary	
Blue Dye	Proprietary	

Test	Specification	Result
Appearance	Blue liquid	Passed

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	10.009	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	
Buffer C	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-1	4 L natural poly	18 months
1601-16	500 mL natural poly	18 months
1601-1CT	4 L Cubitainer®	18 months
1601-2.5	10 L Cubitainer®	18 months
1601-32	1 L natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Version: 1.3

Lot Number: 2410F80

Product Number: 1601

Page 1 of 2



# Certificate of Analysis

**Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C****Lot Number:** 2504F20**Product Number:** 1615**Manufacture Date:** APR 08, 2025**Expiration Date:** SEP 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	15	20	25	30	35	40
pH	12.35	12.17	11.99	11.78	11.62	11.46

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Sodium Hydroxide	1310-73-2	Reagent (from ACS)

Test	Specification	Result
Appearance	Colorless liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	12.009	0.02	186-I-g, 186-II-g, 191d

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1615-1	4 L natural poly	18 months
1615-16	500 mL clear PET-G	18 months
1615-5	20 L Cubitainer®	18 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (04/08/2025)  
Operations Manager

**This product was tested in an ISO 17025 Accredited Laboratory**


This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

## Certificate of Analysis

Product Name:

Barbituric acid - ReagentPlus®, 99%

Product Number: 185698  
Batch Number: WXBFB3271V  
Brand: SIAL  
CAS Number: 67-52-7  
Formula: C<sub>4</sub>H<sub>4</sub>N<sub>2</sub>O<sub>3</sub>  
Formula Weight: 128.09 g/mol  
Quality Release Date: 16 MAY 2024



Test	Specification	Result
Appearance (Colour)	White to Off-White	White
Appearance (Form)	Powder	Powder
Infrared spectrum	Conforms to Structure	Conforms
Purity (Titration by NaOH)	98.5 - 101.5 %	100.4 %
GC (area %)	≥ 98 %	100 %
VPCT		



Kang Chen  
Quality Manager  
Wuxi, China CN

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





Material	BDHVBDH7206-1
Material Description	IODINE SOLUTION 0.025N
Lot	25A2461008
Expires end of	2029-Jan-20
Molecular mass	0
Last Quality Control	2025-Jan-24
Date of manufacture	2025-Jan-21
Made in	United States
Manufacturer Source Batch	MK25A21527

Additional information

Characteristics	Specifications	Measured values
Prepared to formulation on file	Confirmed	Confirmed
Appearance	Passes Test	Passes Test
Normality, N	0.0200 - 0.0300	0.0268

Signature

We certify that this batch conforms to the specifications listed above.

This document has been produced electronically and is valid without a signature.

Michelle Bales - Sr. Manager Quality Assurance  
 Avantor Performance Materials, LLC

For Professional use in Laboratory or Manufacturing. Not for use as an Active Pharmaceutical Ingredient or Food or Animal Feed. Suitability and intended use of the product remains the responsibility of the user.

VWR International LLC, Radnor Corporate Center, Building One, Suite 200, 100 Matsonford Road, Radnor, PA 19087, USA.

VWR International bv, Haasrode Research Park Zone 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium

BDHVBDH72 25A2461008 Page 1 / 1

# Certificate of Analysis

## Cyanide Standard, 1000 ppm CN<sup>-</sup>

**Lot Number:** 1505H73

**Product Number:** 2543

**Manufacture Date:** MAY 08, 2025

**Expiration Date:** NOV 2025

This standard is prepared using accurate volumetric techniques from material that has been assayed against Silver Nitrate solution certified traceable to NIST Standard Reference Material 999. The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is the combined uncertainty based on the stability of the assayed Potassium Cyanide, and the uncertainty in the mass and volume measurements.

Use 0.16% (w/v) (0.04 N) Sodium Hydroxide or 0.225 % (w/v) (0.04 N) Potassium Hydroxide to make dilutions of this standard. Restandardize weekly if extreme accuracy is required.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Cyanide	151-50-8	ACS
Sodium Hydroxide	1310-73-2	Reagent (from ACS)

Test	Specification	Result
Appearance	Colorless liquid	Passed
Cyanide (CN <sup>-</sup> )	995-1005 ppm	1000 ppm

Specification	Reference
Stock Standard Cyanide Solution	APHA (4500-CN- F)
Stock Cyanide Solution	APHA (4500-CN- E)
Stock Cyanide Solution	APHA (4500-CN- K)
Stock Cyanide Solution	APHA (4500-CN- H)
Cyanide Reference Solution (1000 mg/L)	EPA (SW-846) (7.3.3.2)
Cyanide Calibration Stock Solution (1,000 mg/L CN <sup>-</sup> )	EPA (SW-846) (9213)
Stock Cyanide Solution	EPA (335.3)
Stock Cyanide Solution	EPA (335.2)
Cyanide Solution Stock	ASTM (D 4282)
Simple Cyanide Solution, Stock (1.0 g/L CN <sup>-</sup> )	ASTM (D 4374)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2543-16	500 mL amber poly	6 months
2543-32	1 L amber poly	6 months
2543-4	120 mL amber poly	6 months

**Recommended Storage:** 2°C - 8°C (36°F - 46°F)





Ernest Mahan (05/08/2025)  
Plant Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



PERCENT SOLID

Supervisor: Iwona  
Analyst: Jignesh  
Date: 6/16/2025

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 06/13/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 105  
Time OUT: 09:30  
Out Date: 06/14/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB136154

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2306-01	KC-NSP-SG-001	1	1.00	1.00	2.00	2.00	100.0	sub out
Q2306-02	KC-NSP-SG-002	2	1.00	1.00	2.00	2.00	100.0	sub out
Q2306-03	KC-NSP-SG-003	3	1.00	1.00	2.00	2.00	100.0	sub out
Q2306-04	KC-NSP-SG-004	4	1.00	1.00	2.00	2.00	100.0	sub out
Q2311-01	TP03-MH2MH3-WC	5	1.14	10.18	11.32	10.15	88.5	
Q2311-02	TP03-MH2MH3-VOC	6	1.13	10.38	11.51	10.27	88.1	
Q2311-03	TP03-MH2MH3-EPH	7	1.13	10.13	11.26	10.2	89.5	
Q2311-05	TP04-MH2MH3-WC	8	1.18	10.49	11.67	10.81	91.8	
Q2311-06	TP04-MH2MH3-VOC	9	1.15	9.83	10.98	9.98	89.8	
Q2311-07	TP04-MH2MH3-EPH	10	1.13	10.31	11.44	10.55	91.4	
Q2312-01	TP-1	11	1.13	10.86	11.99	10.48	86.1	
Q2312-02	TP-1-EPH	12	1.13	10.85	11.98	10.58	87.1	
Q2312-03	TP-1-VOC	13	1.14	10.65	11.79	10.25	85.5	
Q2318-05	SVOC-GPC-BLANK	14	1.00	1.00	2.00	2.00	100.0	
Q2318-06	PEST-GPC-BLANK	15	1.00	1.00	2.00	2.00	100.0	
Q2318-07	PEST-GPC-BLANK-SPIKE	16	1.00	1.00	2.00	2.00	100.0	
Q2318-08	PCB-GPC-BLANK	17	1.00	1.00	2.00	2.00	100.0	
Q2318-09	PCB-GPC-BLANK-SPIKE	18	1.00	1.00	2.00	2.00	100.0	
Q2318-10	SVOC-GPC2-BLANK	19	1.00	1.00	2.00	2.00	100.0	
Q2318-11	PEST-GPC2-BLANK	20	1.00	1.00	2.00	2.00	100.0	
Q2318-12	PEST-GPC2-BLANK-SPIKE	21	1.00	1.00	2.00	2.00	100.0	
Q2318-13	PCB-GPC2-BLANK	22	1.00	1.00	2.00	2.00	100.0	
Q2318-14	PCB-GCP2-BLANK-SPIKE	23	1.00	1.00	2.00	2.00	100.0	
Q2319-01	MH-B	24	1.18	10.42	11.6	10.24	86.9	
Q2319-02	MH-B-EPH	25	1.12	10.41	11.53	10.17	86.9	
Q2319-03	MH-B-VOC	26	1.18	10.51	11.69	10.12	85.1	
Q2320-01	WC	27	1.17	10.47	11.64	10.46	88.7	
Q2321-01	61125	28	1.00	1.00	2.00	2.00	100.0	wipe sample



# PERCENT SOLID

Supervisor: Iwona  
Analyst: Jignesh  
Date: 6/16/2025

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 06/13/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 105  
Time OUT: 09:30  
Out Date: 06/14/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB136154

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2322-01	CL-01-061325	29	1.18	10.18	11.36	10.08	87.4	
Q2322-02	CL-01-061325-E2	30	1.19	10.37	11.56	10.16	86.5	
Q2323-01	PL-01-06132025	31	1.19	10.28	11.47	10.46	90.2	
Q2323-02	PL-01-06132025-E2	32	1.19	10.00	11.19	10.2	90.1	
Q2324-01	HD-01-6132025	33	1.15	10.30	11.45	10.83	94.0	
Q2324-02	HD-01-6132025-E2	34	1.13	10.17	11.3	10.77	94.8	
Q2325-01	TP-8	35	1.13	10.35	11.48	9.71	82.9	
Q2325-02	TP-8-EPH	36	1.15	10.83	11.98	9.9	80.8	
Q2325-03	TP-8-VOC	37	1.18	10.24	11.42	10.47	90.7	
Q2326-01	BC274770-1-1	38	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-02	BC274770-1-2	39	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-03	BC274770-2-1	40	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-04	BC274770-2-2	41	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-05	BC274770-3-1	42	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-06	BC274770-3-2	43	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-07	ECA45N-1-1	44	1.00	1.00	2.00	2.00	100.0	pilc
Q2326-08	ECA45N-12	45	1.00	1.00	2.00	2.00	100.0	pilc
Q2327-01	HEH700H-1-1	46	1.00	1.00	2.00	2.00	100.0	pilc
Q2327-02	HEH700H-1-2	47	1.00	1.00	2.00	2.00	100.0	pilc
Q2327-03	HEH700H-2-1	48	1.00	1.00	2.00	2.00	100.0	pilc
Q2327-04	HEH700H-2-2	49	1.00	1.00	2.00	2.00	100.0	pilc
Q2328-01	CHRT25653	50	1.14	10.52	11.66	9.84	82.7	
Q2328-02	CHRT25653-E2	51	1.11	10.29	11.4	9.17	78.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

136154

WorkList Name : %1-061325

WorkList ID : 190180

Department : Wet-Chemistry

Date : 06-13-2025 08:14:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2306-01	KC-NSP-SG-001	Solid	Percent Solids	Cool 4 deg C	ENTA08	D41	06/11/2025	Chemtech -SO
Q2306-02	KC-NSP-SG-002	Solid	Percent Solids	Cool 4 deg C	ENTA08	D41	06/11/2025	Chemtech -SO
Q2306-03	KC-NSP-SG-003	Solid	Percent Solids	Cool 4 deg C	ENTA08	D41	06/11/2025	Chemtech -SO
Q2306-04	KC-NSP-SG-004	Solid	Percent Solids	Cool 4 deg C	ENTA08	D41	06/11/2025	Chemtech -SO
Q2311-01	TP03-MH2MH3-WC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2311-02	TP03-MH2MH3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2311-03	TP03-MH2MH3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2311-05	TP04-MH2MH3-WC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2311-06	TP04-MH2MH3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2311-07	TP04-MH2MH3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2312-01	TP-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2312-02	TP-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2312-03	TP-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2318-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-10	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-11	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-12	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO

Date/Time 06-13-25 15:00

Raw Sample Received by: SA WCI

Raw Sample Relinquished by: ASM

Date/Time 06-13-25

Raw Sample Received by: CP M

Raw Sample Relinquished by: SA WCI

17:10

# WORKLIST(Hardcopy Internal Chain)

136154

WorkList Name : %1-061325      WorkList ID : 190180      Department : Wet-Chemistry      Date : 06-13-2025 08:14:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2318-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2318-14	PCB-GCP2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	06/06/2025	Chemtech -SO
Q2319-01	MH-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/13/2025	Chemtech -SO
Q2319-02	MH-B-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/13/2025	Chemtech -SO
Q2319-03	MH-B-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/13/2025	Chemtech -SO
Q2320-01	WC	Solid	Percent Solids	Cool 4 deg C	FIRS02	D41	06/13/2025	Chemtech -SO
Q2321-01	61125	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2322-01	CL-01-061325	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/13/2025	Chemtech -SO
Q2322-02	CL-01-061325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/13/2025	Chemtech -SO
Q2323-01	PL-01-06132025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/13/2025	Chemtech -SO
Q2323-02	PL-01-06132025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/13/2025	Chemtech -SO
Q2324-01	HD-01-6132025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/13/2025	Chemtech -SO
Q2324-02	HD-01-6132025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/13/2025	Chemtech -SO
Q2325-01	TP-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2325-02	TP-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2325-03	TP-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2326-01	BC274770-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2326-02	BC274770-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2326-03	BC274770-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2326-04	BC274770-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2326-05	BC274770-3-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO

Date/Time 06-13-25 15:00  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

# WORKLIST(Hardcopy Internal Chain)

236154

WorkList Name : %1-061325

WorkList ID : 190180

Department : Wet-Chemistry

Date : 06-13-2025 08:14:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2326-06	BC274770-3-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2326-07	ECA45N-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2326-08	ECA45N-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2327-01	HEH700H-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2327-02	HEH700H-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2327-03	HEH700H-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2327-04	HEH700H-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D52	06/13/2025	Chemtech -SO
Q2328-01	CHRT25653	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO
Q2328-02	CHRT25653-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/13/2025	Chemtech -SO

Date/Time 06-13-25 15:00  
Raw Sample Received by: SS w/c  
Raw Sample Relinquished by: cpm

Date/Time 06-13-25 17:10  
Raw Sample Received by: cpm  
Raw Sample Relinquished by: SS w/c



# SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:  
COMPANY: FIRST ENVIRONMENT INC  
ADDRESS: 10 PARK PL  
CITY: BUTLER STATE: N.J. ZIP: 07405  
ATTENTION:  
PHONE: 973-334-0003 FAX: 973-334-0928

CLIENT PROJECT INFORMATION

PROJECT NAME: EDGEWOOD  
PROJECT NO.: LOCATION: EDGEWOOD N.J.  
PROJECT MANAGER: KEN Cwirka  
e-mail: KMCC@FIRSTENVIRONMENT.COM  
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: FIRST ENVIRONMENT PO#:  
ADDRESS: 10 PARK PL Suite 504  
CITY: BUTLER STATE: N.J. ZIP: 07405  
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) \_\_\_\_\_ DAYS\*  
HARDCOPY (DATA PACKAGE): \_\_\_\_\_ DAYS\*  
EDD: STD DAYS\*  
\*TO BE APPROVED BY CHEMTECH  
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP  
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other \_\_\_\_\_  
☒ EDD FORMAT

1. CORROSION  
2. TENSILE  
3. REACTIVE  
4. REACTIVE  
5. REACTIVE  
6. REACTIVE  
7. REACTIVE  
8. REACTIVE  
9. REACTIVE

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	WC	S	X		6-13	0930	3	X	X	X	X	X	X	X	X	X	
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>GARY McLEOD</u>	DATE/TIME: <u>6/13/25/11:35</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP <u>2.8</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: <u>IR-Gun #1</u>
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page ____ of CLIENT: <input type="checkbox"/> Hand Delivered <input type="checkbox"/> Other Shipment Complete <input type="checkbox"/> YES <input type="checkbox"/> NO



### Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488