



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: First Environment, Inc.

SDG No.: Q2320

Project: EDGEW001 – Veterans Field Edgewater, NJ

RunNo.: LB136162

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.094	0.099	95	85–115	06/16/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.23	0.25	92	90–110	06/16/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90–110	06/16/2025

Initial and Continuing Calibration Verification

Client: First Environment, Inc.

SDG No.: Q2320

Project: EDGEW001 – Veterans Field Edgewater, NJ

RunNo.: LB136178

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.00	7	100	90-110	06/17/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	06/17/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	06/17/2025
Sample ID: Corrosivity	CCV3	pH	2.02	2.00	101	90-110	06/17/2025