


LB136181

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
B= 2.2702e-004
C= 6.8206e-003
Rho= 0.9998853
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	167	0.000	167	0			
0.2	0.200	0.218	0.018	931	0.0 %	931				
2.5	2.500	2.497	-0.003	10969	0.0 %	10969				
5.0	5.000	4.889	-0.111	21506	0.0 %	21506				
7.5	7.500	7.488	-0.012	32954	0.0 %	32954				
10.0	10.000	10.063	0.063	44296	0.0 %	44296				

0.000
10.000
-0.000
0.000

LB136181 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jun 2025 09:57:26

Sample ID	Extended ID	µ Abs	Conc	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	167	-	0.000	7470A	PPB	18 Jun 2025 10:20:25	S	Std
0.2 - 1	50-2	931	-	0.200	7470A	PPB	18 Jun 2025 10:25:36	S	Std
2.5 - 1	52-5	10968	-	2.500	7470A	PPB	18 Jun 2025 10:27:53	S	Std
5.0 - 1	55	21506	-	5.000	7470A	PPB	18 Jun 2025 10:30:10	S	Std
7.5 - 1	57-5	32954	-	7.500	7470A	PPB	18 Jun 2025 10:32:28	S	Std
10.0 - 1	51-5	44296	-	10.000	7470A	PPB	18 Jun 2025 10:34:46	S	Std
ICV43 - 1	ICV43	16489	3.7502	-	7470A	PPB	18 Jun 2025 10:37:36	U	SMPL
ICB43 - 1	ICB43	-117	-0.0197	-	7470A	PPB	18 Jun 2025 10:39:52	U	SMPL
CCV57 - 1	CCV57	20009	4.5493	-	7470A	PPB	18 Jun 2025 10:42:09	U	SMPL
CCB57 - 1	CCB57	-151	-0.0275	-	7470A	PPB	18 Jun 2025 10:44:24	U	SMPL
CRA - 1	CRA	879	0.2064	-	7470A	PPB	18 Jun 2025 10:46:41	U	SMPL
HighStd - 1	HighStd	43812	9.9530	-	7470A	PPB	18 Jun 2025 10:48:56	U	SMPL
ChkStd - 1	ChkStd	28107	6.3877	-	7470A	PPB	18 Jun 2025 10:51:12	U	SMPL
PB168519BL - 1	PBW	-384	-0.0804	-	7470A	PPB	18 Jun 2025 10:53:31	U	SMPL
PB168519BS - 1	LCSW	16101	3.6621	-	7470A	PPB	18 Jun 2025 10:55:50	U	SMPL
Q2311-04 - 1	TP03-MH2MH3-WC	1462	0.3387	-	7470A	PPB	18 Jun 2025 11:04:08	U	SMPL
Q2311-08 - 1	TP04-MH2MH3-WC	1426	0.3306	-	7470A	PPB	18 Jun 2025 11:06:24	U	SMPL
Q2312-04 - 1	TP-1	879	0.2064	-	7470A	PPB	18 Jun 2025 11:14:28	U	SMPL
Q2319-04 - 1	MH-B	475	0.1147	-	7470A	PPB	18 Jun 2025 11:16:44	U	SMPL
Q2320-01 - 1	WC	2627	0.6032	-	7470A	PPB	18 Jun 2025 11:18:59	U	SMPL
CCV58 - 1	CCV58	23705	5.3883	-	7470A	PPB	18 Jun 2025 11:21:15	U	SMPL
CCB58 - 1	CCB58	-116	-0.0195	-	7470A	PPB	18 Jun 2025 11:23:30	U	SMPL
Q2325-04 - 1	TP-8	717	0.1696	-	7470A	PPB	18 Jun 2025 11:28:38	U	SMPL
Q2325-04DUP - 1	TP-8DUP	600	0.1430	-	7470A	PPB	18 Jun 2025 11:30:52	U	SMPL
Q2325-04MS - 1	TP-8MS	17948	4.0814	-	7470A	PPB	18 Jun 2025 11:33:08	U	SMPL
Q2325-04MSD - 1	TP-8MSD	17486	3.9765	-	7470A	PPB	18 Jun 2025 11:35:24	U	SMPL
PB168530BL - 1	PBW	52	0.0186	-	7470A	PPB	18 Jun 2025 11:37:42	U	SMPL
PB168530BS - 1	LCSW	17053	3.8782	-	7470A	PPB	18 Jun 2025 11:40:01	U	SMPL
Q2126-07 - 1	LOD-MDL-WATER-01-QT2-2025	149	0.0406	-	7470A	PPB	18 Jun 2025 11:47:32	U	SMPL
Q2126-08 - 1	LOQ-WATER-02-QT2-2025	760	0.1794	-	7470A	PPB	18 Jun 2025 11:49:49	U	SMPL
Q2307-03 - 1	LINDEN-SAA-WATER	240276	54.5544	-	7470A	PPB	18 Jun 2025 11:52:06	U	SMPL
Q2307-03DUP - 1	LINDEN-SAA-WATERDUP	227381	51.6269	-	7470A	PPB	18 Jun 2025 11:54:22	U	SMPL
CCV59 - 1	CCV59	20695	4.7050	-	7470A	PPB	18 Jun 2025 11:57:21	U	SMPL
CCB59 - 1	CCB59	-304	-0.0622	-	7470A	PPB	18 Jun 2025 12:00:07	U	SMPL
Q2307-03MS - 1	LINDEN-SAA-WATERMS	247821	56.2672	-	7470A	PPB	18 Jun 2025 12:02:25	U	SMPL
Q2307-03MSD - 1	LINDEN-SAA-WATERMSD	241248	54.7750	-	7470A	PPB	18 Jun 2025 12:04:41	U	SMPL
PB168497TB - 1	PB168497TB	121	0.0343	-	7470A	PPB	18 Jun 2025 12:10:44	U	SMPL
Q2325-04LX5 - 1		112	0.0322	-	7470A	PPB	18 Jun 2025 12:13:00	U	SMPL
Q2325-04A - 1		18011	4.0957	-	7470A	PPB	18 Jun 2025 12:15:16	U	SMPL
Q2307-03LX5 - 1		46738	10.6173	-	7470A	PPB	18 Jun 2025 12:17:35	U	SMPL
Q2307-03A - 1		241309	54.7889	-	7470A	PPB	18 Jun 2025 12:19:54	U	SMPL
CCV60 - 1	CCV60	22857	5.1958	-	7470A	PPB	18 Jun 2025 12:26:12	U	SMPL
CCB60 - 1	CCB60	-149	-0.0270	-	7470A	PPB	18 Jun 2025 12:28:28	U	SMPL
Q2307-03DLX10 - 1		25846	5.8744	-	7470A	PPB	18 Jun 2025 12:30:47	U	SMPL
Q2307-03DUPDLX10 - 1		27408	6.2290	-	7470A	PPB	18 Jun 2025 12:33:02	U	SMPL
Q2307-03MSDLX10 - 1		30499	6.9307	-	7470A	PPB	18 Jun 2025 12:35:20	U	SMPL
Q2307-03MSDDLX10 - 1		30245	6.8731	-	7470A	PPB	18 Jun 2025 12:37:38	U	SMPL
Q2307-03LDLX50 - 1		5663	1.2924	-	7470A	PPB	18 Jun 2025 12:44:43	U	SMPL
Q2307-03ADLX10 - 1		27392	6.2254	-	7470A	PPB	18 Jun 2025 12:46:58	U	SMPL
CCV61 - 1	CCV61	22035	5.0092	-	7470A	PPB	18 Jun 2025 12:49:14	U	SMPL
CCB61 - 1	CCB61	-60	-0.0068	-	7470A	PPB	18 Jun 2025 12:51:32	U	SMPL