

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136181

Review By	MOHAN	Review On	6/19/2025 10:26:06 AM
Supervise By	jaswal	Supervise On	6/19/2025 3:46:53 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85991,MP85992,MP85993,MP85994,MP85995,MP85996		
ICV Standard	MP85997		
CCV Standard	MP85999		
ICSA Standard			
CRI Standard	MP86001		
LCS Standard			
Chk Standard	MP85998,MP86000,MP86002,MP86006		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/18/25 10:20		MOHAN	OK
2	S0.2	S0.2	CAL2	06/18/25 10:25		MOHAN	OK
3	S2.5	S2.5	CAL3	06/18/25 10:27		MOHAN	OK
4	S5	S5	CAL4	06/18/25 10:30		MOHAN	OK
5	S7.5	S7.5	CAL5	06/18/25 10:32		MOHAN	OK
6	S10	S10	CAL6	06/18/25 10:34		MOHAN	OK
7	ICV43	ICV43	ICV	06/18/25 10:37		MOHAN	OK
8	ICB43	ICB43	ICB	06/18/25 10:39		MOHAN	OK
9	CCV57	CCV57	CCV	06/18/25 10:42		MOHAN	OK
10	CCB57	CCB57	CCB	06/18/25 10:44		MOHAN	OK
11	CRA	CRA	CRDL	06/18/25 10:46		MOHAN	OK
12	HighStd	HighStd	HIGH STD	06/18/25 10:48		MOHAN	OK
13	ChkStd	ChkStd	SAM	06/18/25 10:51		MOHAN	OK
14	PB168519BL	PB168519BL	MB	06/18/25 10:53		MOHAN	OK
15	PB168519BS	PB168519BS	LCS	06/18/25 10:55		MOHAN	OK
16	Q2311-04	TP03-MH2MH3-WC	SAM	06/18/25 11:04		MOHAN	OK
17	Q2311-08	TP04-MH4-WC	SAM	06/18/25 11:06		MOHAN	OK
18	Q2312-04	TP-1	SAM	06/18/25 11:14		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136181

Review By	MOHAN	Review On	6/19/2025 10:26:06 AM
Supervise By	jaswal	Supervise On	6/19/2025 3:46:53 PM

STD. NAME	STD REF.#
ICAL Standard	MP85991,MP85992,MP85993,MP85994,MP85995,MP85996
ICV Standard	MP85997
CCV Standard	MP85999
ICSA Standard	
CRI Standard	MP86001
LCS Standard	
Chk Standard	MP85998,MP86000,MP86002,MP86006

19	Q2319-04	MH-B	SAM	06/18/25 11:16		MOHAN	OK
20	Q2320-01	WC	SAM	06/18/25 11:18		MOHAN	OK
21	CCV58	CCV58	CCV	06/18/25 11:21		MOHAN	OK
22	CCB58	CCB58	CCB	06/18/25 11:23		MOHAN	OK
23	Q2325-04	TP-8	SAM	06/18/25 11:28		MOHAN	OK
24	Q2325-04DUP	TP-8DUP	DUP	06/18/25 11:30		MOHAN	OK
25	Q2325-04MS	TP-8MS	MS	06/18/25 11:33		MOHAN	OK
26	Q2325-04MSD	TP-8MSD	MSD	06/18/25 11:35		MOHAN	OK
27	PB168530BL	PB168530BL	MB	06/18/25 11:37		MOHAN	OK
28	PB168530BS	PB168530BS	LCS	06/18/25 11:40		MOHAN	OK
29	Q2126-07	LOD-MDL-WATER-01	SAM	06/18/25 11:47		MOHAN	OK
30	Q2126-08	LOQ-WATER-02-QT2	LOQ	06/18/25 11:49		MOHAN	OK
31	Q2307-03	LINDEN-SAA-WATER	SAM	06/18/25 11:52	Hg high	MOHAN	Dilution
32	Q2307-03DUP	LINDEN-SAA-WATER	DUP	06/18/25 11:54	Hg high	MOHAN	Dilution
33	CCV59	CCV59	CCV	06/18/25 11:57		MOHAN	OK
34	CCB59	CCB59	CCB	06/18/25 12:00		MOHAN	OK
35	Q2307-03MS	LINDEN-SAA-WATER	MS	06/18/25 12:02	Hg high	MOHAN	Dilution
36	Q2307-03MSD	LINDEN-SAA-WATER	MSD	06/18/25 12:04	Hg high	MOHAN	Dilution
37	PB168497TB	PB168497TB	MB	06/18/25 12:10		MOHAN	OK
38	Q2325-04L	TP-8L	SD	06/18/25 12:13		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136181

Review By	MOHAN	Review On	6/19/2025 10:26:06 AM
Supervise By	jaswal	Supervise On	6/19/2025 3:46:53 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85991,MP85992,MP85993,MP85994,MP85995,MP85996		
ICV Standard	MP85997		
CCV Standard	MP85999		
ICSA Standard			
CRI Standard	MP86001		
LCS Standard			
Chk Standard	MP85998,MP86000,MP86002,MP86006		

39	Q2325-04A	TP-8A	PS	06/18/25 12:15		MOHAN	OK
40	Q2307-03L	LINDEN-SAA-WATER	SD	06/18/25 12:17	Hg high	MOHAN	Dilution
41	Q2307-03A	LINDEN-SAA-WATER	PS	06/18/25 12:19	Hg high	MOHAN	Dilution
42	CCV60	CCV60	CCV	06/18/25 12:26		MOHAN	OK
43	CCB60	CCB60	CCB	06/18/25 12:28		MOHAN	OK
44	Q2307-03DL	LINDEN-SAA-WATER	SAM	06/18/25 12:30	10X For Hg high	MOHAN	Confirms
45	Q2307-03DUPDL	LINDEN-SAA-WATER	DUP	06/18/25 12:33	10X For Hg high	MOHAN	Confirms
46	Q2307-03MSDL	LINDEN-SAA-WATER	MS	06/18/25 12:35	10X For Hg high	MOHAN	Confirms
47	Q2307-03MSDDL	LINDEN-SAA-WATER	MSD	06/18/25 12:37	10X For Hg high	MOHAN	Confirms
48	Q2307-03LDL	LINDEN-SAA-WATER	SD	06/18/25 12:44	50X For Hg high	MOHAN	Confirms
49	Q2307-03ADL	LINDEN-SAA-WATER	PS	06/18/25 12:46	10X For Hg high	MOHAN	Confirms
50	CCV61	CCV61	CCV	06/18/25 12:49		MOHAN	OK
51	CCB61	CCB61	CCB	06/18/25 12:51		MOHAN	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136210

Review By	Janvi	Review On	6/20/2025 12:25:55 PM
Supervise By	MOHAN	Supervise On	6/23/2025 5:20:13 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/19/25 14:10		Jaswal	OK
2	S1	S1	CAL2	06/19/25 14:14		Jaswal	OK
3	S2	S2	CAL3	06/19/25 14:19		Jaswal	OK
4	S3	S3	CAL4	06/19/25 14:23		Jaswal	OK
5	S4	S4	CAL5	06/19/25 14:27		Jaswal	OK
6	S5	S5	CAL6	06/19/25 14:31		Jaswal	OK
7	ICV01	ICV01	ICV	06/19/25 14:45	ICV fail for Sb,As,Ag (200.7) (95-105)	Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/19/25 14:57		Jaswal	OK
9	ICB01	ICB01	ICB	06/19/25 15:01		Jaswal	OK
10	CRI01	CRI01	CRDL	06/19/25 15:06		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/19/25 15:10		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/19/25 15:16		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/19/25 15:23		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/19/25 15:28		Jaswal	OK
15	CCV01	CCV01	CCV	06/19/25 15:33		Jaswal	OK
16	CCB01	CCB01	CCB	06/19/25 15:43		Jaswal	OK
17	Q2336-01	WBR-1	SAM	06/19/25 15:47		Jaswal	OK
18	Q2336-03	WBR-2	SAM	06/19/25 15:51		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136210

Review By	Janvi	Review On	6/20/2025 12:25:55 PM
Supervise By	MOHAN	Supervise On	6/23/2025 5:20:13 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	Q2337-01	OR-02-061625	SAM	06/19/25 15:56		Jaswal	OK
20	Q2338-01	OK-03-061625	SAM	06/19/25 16:00		Jaswal	OK
21	Q2340-01	TP05-MH17A-WC	SAM	06/19/25 16:04		Jaswal	OK
22	Q2339-01	MH-C	SAM	06/19/25 16:09		Jaswal	OK
23	Q2341-05	EP-3	SAM	06/19/25 16:13		Jaswal	OK
24	Q2341-01	TP-9	SAM	06/19/25 16:18	RPD fail for more than 50% elements	Jaswal	Not Ok
25	Q2340-04	TP05-MH17A-WC	SAM	06/19/25 16:22		Jaswal	OK
26	Q2348-01	001-WILLETS-PT-BLV	SAM	06/19/25 16:26		Jaswal	OK
27	CCV02	CCV02	CCV	06/19/25 16:31		Jaswal	OK
28	CCB02	CCB02	CCB	06/19/25 16:35		Jaswal	OK
29	Q2348-02	002-35TH-AVE(MAY)	SAM	06/19/25 16:40		Jaswal	OK
30	CCV03	CCV03	CCV	06/19/25 16:44		Jaswal	OK
31	CCB03	CCB03	CCB	06/19/25 16:48		Jaswal	OK
32	Q2348-02DUP	002-35TH-AVE(MAY)	DUP	06/19/25 16:53		Jaswal	OK
33	Q2348-02L	002-35TH-AVE(MAY)	SD	06/19/25 16:57		Jaswal	OK
34	Q2348-02MS	002-35TH-AVE(MAY)	MS	06/19/25 17:02		Jaswal	OK
35	Q2348-02MSD	002-35TH-AVE(MAY)	MSD	06/19/25 17:06		Jaswal	OK
36	Q2348-02A	002-35TH-AVE(MAY)	PS	06/19/25 17:10	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
37	Q2349-01	001-WILLETS-PT-BLV	SAM	06/19/25 17:15		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136210

Review By	Janvi	Review On	6/20/2025 12:25:55 PM
Supervise By	MOHAN	Supervise On	6/23/2025 5:20:13 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

38	Q2349-04	002-35TH-AVE(JUNE	SAM	06/19/25 17:19		Jaswal	OK
39	PB168539BS	PB168539BS	LCS	06/19/25 17:28		Jaswal	OK
40	CCV04	CCV04	CCV	06/19/25 17:32		Jaswal	OK
41	CCB04	CCB04	CCB	06/19/25 17:36		Jaswal	OK
42	PB168539BL	PB168539BL	MB	06/19/25 17:41		Jaswal	OK
43	Q2341-01DUP	TP-9DUP	DUP	06/19/25 17:45	RPD fail for more than 50% elements	Jaswal	Not Ok
44	Q2341-01L	TP-9L	SD	06/19/25 17:50	RPD fail for more than 50% elements	Jaswal	Not Ok
45	Q2341-01MS	TP-9MS	MS	06/19/25 17:54	RPD fail for more than 50% elements	Jaswal	Not Ok
46	Q2341-01MSD	TP-9MSD	MSD	06/19/25 17:58	RPD fail for more than 50% elements	Jaswal	Not Ok
47	Q2341-01A	TP-9A	PS	06/19/25 18:03	RPD fail for more than 50% elements	Jaswal	Not Ok
48	PB168518BL	PB168518BL	MB	06/19/25 18:07		Jaswal	OK
49	PB168518BS	PB168518BS	LCS	06/19/25 18:11		Jaswal	OK
50	Q2312-04	TP-1	SAM	06/19/25 18:16	Bad injection	Jaswal	Not Ok
51	Q2319-04	MH-B	SAM	06/19/25 18:20	Bad injection	Jaswal	Not Ok
52	CCV05	CCV05	CCV	06/19/25 18:25		Jaswal	OK
53	CCB05	CCB05	CCB	06/19/25 18:29		Jaswal	OK
54	Q2320-01	WC	SAM	06/19/25 18:34		Jaswal	OK
55	Q2325-04	TP-8	SAM	06/19/25 18:38		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136210

Review By	Janvi	Review On	6/20/2025 12:25:55 PM
Supervise By	MOHAN	Supervise On	6/23/2025 5:20:13 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

56	Q2325-04DUP	TP-8DUP	DUP	06/19/25 18:43		Jaswal	OK
57	Q2325-04L	TP-8L	SD	06/19/25 18:47		Jaswal	OK
58	Q2325-04MS	TP-8MS	MS	06/19/25 18:52		Jaswal	OK
59	Q2325-04MSD	TP-8MSD	MSD	06/19/25 18:56		Jaswal	OK
60	Q2325-04A	TP-8A	PS	06/19/25 19:01	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
61	PB168497TB	PB168497TB	MB	06/19/25 19:05		Jaswal	OK
62	Q2328-05	CHRT25653	SAM	06/19/25 19:14		Jaswal	OK
63	CCV06	CCV06	CCV	06/19/25 19:19		Jaswal	OK
64	CCB06	CCB06	CCB	06/19/25 19:23		Jaswal	OK
65	Q2358-01	NB-07-06182025	SAM	06/19/25 19:27		Jaswal	OK
66	Q2358-01DUP	NB-07-06182025DUP	DUP	06/19/25 19:32		Jaswal	OK
67	Q2358-01L	NB-07-06182025L	SD	06/19/25 19:36		Jaswal	OK
68	Q2358-01MS	NB-07-06182025MS	MS	06/19/25 19:40		Jaswal	OK
69	Q2358-01MSD	NB-07-06182025MSD	MSD	06/19/25 19:45		Jaswal	OK
70	Q2358-01A	NB-07-06182025A	PS	06/19/25 19:49	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
71	CCV07	CCV07	CCV	06/19/25 19:53		Jaswal	OK
72	CCB07	CCB07	CCB	06/19/25 19:58		Jaswal	OK