



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: First Environment, Inc. **SDG No.:** Q2320
Contract: FIRS02 **Lab Code:** CHEM **Case No.:** Q2320 **SAS No.:** Q2320
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	103	70 - 130	CV	06/18/2025	10:46	LB136181
CRI01	Arsenic	19.0	20.0	95	65 - 135	P	06/19/2025	15:06	LB136210
	Barium	97.8	100	98	65 - 135	P	06/19/2025	15:06	LB136210
	Cadmium	5.58	6.0	93	65 - 135	P	06/19/2025	15:06	LB136210
	Chromium	9.17	10.0	92	65 - 135	P	06/19/2025	15:06	LB136210
	Lead	10.2	12.0	85	65 - 135	P	06/19/2025	15:06	LB136210
	Selenium	22.0	20.0	110	65 - 135	P	06/19/2025	15:06	LB136210
	Silver	8.80	10.0	88	65 - 135	P	06/19/2025	15:06	LB136210