



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	First Environment, Inc.				SDG No.:	Q2320				
Contract:	FIRS02	Lab Code:	CHEM	Case No.:	Q2320	SAS No.:	Q2320			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB43	Mercury	0.076	+/-0.2	U	0.20	CV	06/18/2025	10:39	LB136181	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: First Environment, Inc.

SDG No.: Q2320

Contract: FIRS02

Lab Code: CHEM

Case No.: Q2320

SAS No.: Q2320

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB57	Mercury	0.076	+/-0.2	U	0.20	CV	06/18/2025	10:44	LB136181
CCB58	Mercury	0.076	+/-0.2	U	0.20	CV	06/18/2025	11:23	LB136181
CCB59	Mercury	0.076	+/-0.2	U	0.20	CV	06/18/2025	12:00	LB136181
CCB60	Mercury	0.076	+/-0.2	U	0.20	CV	06/18/2025	12:28	LB136181
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	06/18/2025	12:51	LB136181

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>First Environment, Inc.</u>		SDG No.: <u>Q2320</u>							
Contract: <u>FIRS02</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q2320</u>		SAS No.: <u>Q2320</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	15:01	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	15:01	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	15:01	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	15:01	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	15:01	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	15:01	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	15:01	LB136210

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: First Environment, Inc.

SDG No.: Q2320

Contract: FIRS02

Lab Code: CHEM

Case No.: Q2320

SAS No.: Q2320

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	15:43	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	15:43	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	15:43	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	15:43	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	15:43	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	15:43	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	15:43	LB136210
CCB02	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	16:35	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	16:35	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	16:35	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	16:35	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	16:35	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	16:35	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	16:35	LB136210
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	16:48	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	16:48	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	16:48	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	16:48	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	16:48	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	16:48	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	16:48	LB136210
CCB04	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	17:36	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	17:36	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	17:36	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	17:36	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	17:36	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	17:36	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	17:36	LB136210
CCB05	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	18:29	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	18:29	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	18:29	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	18:29	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	18:29	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	18:29	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	18:29	LB136210
CCB06	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	19:23	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	19:23	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	19:23	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	19:23	LB136210

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: First Environment, Inc.

SDG No.: Q2320

Contract: FIRS02

Lab Code: CHEM

Case No.: Q2320

SAS No.: Q2320

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	06/19/2025	19:23	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	19:23	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	19:23	LB136210
CCB07	Arsenic	5.12	+/-10	U	20.0	P	06/19/2025	19:58	LB136210
	Barium	14.6	+/-50	U	100	P	06/19/2025	19:58	LB136210
	Cadmium	0.50	+/-3	U	6.00	P	06/19/2025	19:58	LB136210
	Chromium	2.12	+/-5	U	10.0	P	06/19/2025	19:58	LB136210
	Lead	2.30	+/-6	U	12.0	P	06/19/2025	19:58	LB136210
	Selenium	9.64	+/-10	U	20.0	P	06/19/2025	19:58	LB136210
	Silver	1.62	+/-5	U	10.0	P	06/19/2025	19:58	LB136210