

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: First Environment, Inc. **Level:** LOW **SDG No.:** Q2320
Contract: FIRS02 **Lab Code:** CHEM **Case No.:** Q2320 **SAS No.:** Q2320
Matrix: Water **Sample ID:** Q2325-04 **Client ID:** TP-8DUP
Percent Solids for Sample: NA **Duplicate ID** Q2325-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	2720		2320		16		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	295		251		16		P
Mercury	ug/L	20	1.70	J	1.43	J	17		CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: First Environment, Inc. **Level:** LOW **SDG No.:** Q2320
Contract: FIRS02 **Lab Code:** CHEM **Case No.:** Q2320 **SAS No.:** Q2320
Matrix: Water **Sample ID:** Q2325-04MS **Client ID:** TP-8MSD
Percent Solids for Sample: NA **Duplicate ID** Q2325-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4320		3990		8		P
Barium	ug/L	20	3990		3260		20		P
Cadmium	ug/L	20	1080		1010		7		P
Chromium	ug/L	20	2170		2030		7		P
Lead	ug/L	20	5410		5010		8		P
Mercury	ug/L	20	40.8		39.8		2		CV
Selenium	ug/L	20	10500		9800		7		P
Silver	ug/L	20	385		359		7		P