

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** First Environment, Inc. **SDG No.:** Q2320  
**Contract:** FIRS02 **Lab Code:** CHEM **Case No.:** Q2320 **SAS No.:** Q2320

| Analyte    | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|------------|-------|------------|--------|---|---------------|----------------------|---|
| PB168518BS |       |            |        |   |               |                      |   |
| Arsenic    | ug/L  | 4000       | 3560   |   | 89            | 80 - 120             | P |
| Barium     | ug/L  | 1000       | 1020   |   | 102           | 80 - 120             | P |
| Cadmium    | ug/L  | 1000       | 939    |   | 94            | 80 - 120             | P |
| Chromium   | ug/L  | 2000       | 2010   |   | 100           | 80 - 120             | P |
| Lead       | ug/L  | 5000       | 4700   |   | 94            | 80 - 120             | P |
| Selenium   | ug/L  | 10000      | 9960   |   | 100           | 80 - 120             | P |
| Silver     | ug/L  | 380        | 368    |   | 97            | 80 - 120             | P |

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** First Environment, Inc.

**SDG No.:** Q2320

**Contract:** FIRS02

**Lab Code:** CHEM

**Case No.:** Q2320

**SAS No.:** Q2320

| Analyte               | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M  |
|-----------------------|-------|------------|--------|---|---------------|----------------------|----|
| PB168519BS<br>Mercury | ug/L  | 4.0        | 3.66   |   | 92            | 80 - 120             | CV |