

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

TP-8L

Lab Name: Chemtech Consulting Group

Contract: FIRS02

Lab Code: CHEM **Lb No.:** lb136210

Lab Sample ID : Q2325-04L

SDG No.: Q2320

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	2720		2570		6		P
Cadmium	30.0	U	150	U			P
Chromium	50.0	U	250	U			P
Lead	295		278	J	6		P
Mercury	1.70	J	10.0	U	100.0		CV
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P