

CALIBRATION VERIFICATION SUMMARY

Contract: FIRS02

Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/16/2025

Lab Sample No.: AR1660CCC500 Data File : PP072958.D Time Analyzed: 10:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.656	5.552	5.752	413.160	500.000	-17.4
Aroclor-1016-2	5.677	5.574	5.774	460.960	500.000	-7.8
Aroclor-1016-3	5.739	5.636	5.836	448.800	500.000	-10.2
Aroclor-1016-4	5.837	5.734	5.934	467.540	500.000	-6.5
Aroclor-1016-5	6.129	6.027	6.227	461.030	500.000	-7.8
Aroclor-1260-1	7.245	7.147	7.347	444.870	500.000	-11.0
Aroclor-1260-2	7.498	7.401	7.601	445.180	500.000	-11.0
Aroclor-1260-3	7.856	7.760	7.960	469.640	500.000	-6.1
Aroclor-1260-4	8.080	7.985	8.185	457.440	500.000	-8.5
Aroclor-1260-5	8.399	8.303	8.503	479.310	500.000	-4.1
Decachlorobiphenyl	10.199	10.113	10.313	59.090	50.000	18.2
Tetrachloro-m-xylene	4.504	4.396	4.596	48.180	50.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Contract: FIRS02

Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/16/2025

Lab Sample No.: AR1660CCC500 Data File : PP072958.D Time Analyzed: 10:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.873	4.777	4.977	479.300	500.000	-4.1
Aroclor-1016-2	4.891	4.795	4.995	496.470	500.000	-0.7
Aroclor-1016-3	5.068	4.973	5.173	498.160	500.000	-0.4
Aroclor-1016-4	5.109	5.015	5.215	508.080	500.000	1.6
Aroclor-1016-5	5.324	5.229	5.429	499.200	500.000	-0.2
Aroclor-1260-1	6.354	6.264	6.464	522.780	500.000	4.6
Aroclor-1260-2	6.543	6.453	6.653	505.780	500.000	1.2
Aroclor-1260-3	6.695	6.606	6.806	524.460	500.000	4.9
Aroclor-1260-4	7.164	7.077	7.277	518.500	500.000	3.7
Aroclor-1260-5	7.407	7.319	7.519	530.380	500.000	6.1
Decachlorobiphenyl	8.804	8.723	8.923	67.090	50.000	34.2
Tetrachloro-m-xylene	3.793	3.691	3.891	50.830	50.000	1.7