

CALIBRATION VERIFICATION SUMMARY

Contract: FIRS02

Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/16/2025

Lab Sample No.: AR1660CCC500 Data File : PP072973.D Time Analyzed: 15:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.646	5.552	5.752	427.360	500.000	-14.5
Aroclor-1016-2	5.669	5.574	5.774	427.440	500.000	-14.5
Aroclor-1016-3	5.731	5.636	5.836	425.490	500.000	-14.9
Aroclor-1016-4	5.829	5.734	5.934	436.300	500.000	-12.7
Aroclor-1016-5	6.121	6.027	6.227	444.620	500.000	-11.1
Aroclor-1260-1	7.238	7.147	7.347	397.080	500.000	-20.6
Aroclor-1260-2	7.491	7.401	7.601	392.490	500.000	-21.5
Aroclor-1260-3	7.848	7.760	7.960	400.340	500.000	-19.9
Aroclor-1260-4	8.071	7.985	8.185	410.090	500.000	-18.0
Aroclor-1260-5	8.391	8.303	8.503	410.200	500.000	-18.0
Decachlorobiphenyl	10.190	10.113	10.313	51.560	50.000	3.1
Tetrachloro-m-xylene	4.497	4.396	4.596	47.860	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: FIRS02

Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/16/2025

Lab Sample No.: AR1660CCC500 Data File : PP072973.D Time Analyzed: 15:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.870	4.777	4.977	486.410	500.000	-2.7
Aroclor-1016-2	4.888	4.795	4.995	500.200	500.000	0.0
Aroclor-1016-3	5.065	4.973	5.173	501.030	500.000	0.2
Aroclor-1016-4	5.107	5.015	5.215	509.210	500.000	1.8
Aroclor-1016-5	5.320	5.229	5.429	499.920	500.000	0.0
Aroclor-1260-1	6.351	6.264	6.464	477.270	500.000	-4.5
Aroclor-1260-2	6.540	6.453	6.653	456.590	500.000	-8.7
Aroclor-1260-3	6.691	6.606	6.806	481.590	500.000	-3.7
Aroclor-1260-4	7.162	7.077	7.277	471.910	500.000	-5.6
Aroclor-1260-5	7.404	7.319	7.519	458.110	500.000	-8.4
Decachlorobiphenyl	8.799	8.723	8.923	58.210	50.000	16.4
Tetrachloro-m-xylene	3.791	3.691	3.891	51.020	50.000	2.0