

CALIBRATION VERIFICATION SUMMARY

Contract: FIRS02

Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/16/2025

Lab Sample No.: AR1660CCC500 Data File : PP072984.D Time Analyzed: 19:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.647	5.552	5.752	422.260	500.000	-15.5
Aroclor-1016-2	5.668	5.574	5.774	451.220	500.000	-9.8
Aroclor-1016-3	5.730	5.636	5.836	459.870	500.000	-8.0
Aroclor-1016-4	5.828	5.734	5.934	475.900	500.000	-4.8
Aroclor-1016-5	6.119	6.027	6.227	457.290	500.000	-8.5
Aroclor-1260-1	7.236	7.147	7.347	425.990	500.000	-14.8
Aroclor-1260-2	7.490	7.401	7.601	413.520	500.000	-17.3
Aroclor-1260-3	7.847	7.760	7.960	422.080	500.000	-15.6
Aroclor-1260-4	8.071	7.985	8.185	437.620	500.000	-12.5
Aroclor-1260-5	8.389	8.303	8.503	438.940	500.000	-12.2
Decachlorobiphenyl	10.187	10.113	10.313	54.380	50.000	8.8
Tetrachloro-m-xylene	4.495	4.396	4.596	47.560	50.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Contract: FIRS02

Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/16/2025

Lab Sample No.: AR1660CCC500 Data File : PP072984.D Time Analyzed: 19:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.871	4.777	4.977	493.000	500.000	-1.4
Aroclor-1016-2	4.888	4.795	4.995	515.350	500.000	3.1
Aroclor-1016-3	5.065	4.973	5.173	517.980	500.000	3.6
Aroclor-1016-4	5.107	5.015	5.215	525.960	500.000	5.2
Aroclor-1016-5	5.320	5.229	5.429	528.430	500.000	5.7
Aroclor-1260-1	6.351	6.264	6.464	524.870	500.000	5.0
Aroclor-1260-2	6.540	6.453	6.653	505.380	500.000	1.1
Aroclor-1260-3	6.692	6.606	6.806	528.820	500.000	5.8
Aroclor-1260-4	7.162	7.077	7.277	515.740	500.000	3.1
Aroclor-1260-5	7.404	7.319	7.519	508.570	500.000	1.7
Decachlorobiphenyl	8.800	8.723	8.923	65.200	50.000	30.4
Tetrachloro-m-xylene	3.791	3.691	3.891	51.520	50.000	3.0