

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                |                           |                   |
|-----------------|--------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q2320</u>                   | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>First Environment, Inc.</u> | <b>Datafile :</b>         | <u>PP072964.D</u> |

| Lab Sample ID            | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|--------------------------|-----------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|                          |           |       |        |       |     |     |      | Qual | Low     | High      |     |
| PB168483BS<br>(Column 1) | AR1016    | 166.6 | 125    | ug/kg | 75  |     |      |      | 40 (71) | 140 (120) |     |
|                          | AR1260    | 166.6 | 122    | ug/kg | 73  |     |      |      | 40 (65) | 140 (130) |     |
| PB168483BS<br>(Column 2) | AR1016    | 166.6 | 132    | ug/kg | 79  |     |      |      | 40 (71) | 140 (120) |     |
|                          | AR1260    | 166.6 | 143    | ug/kg | 86  |     |      |      | 40 (65) | 140 (130) |     |