

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2320

Client: First Environment, Inc.

Analytical Method: SW8260-Low

Datafile : VN087052.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0617WBSD01	Vinyl chloride	20	19.5	ug/L	98	10		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	20.7	ug/L	104	5		70 (74)	130 (110)	20 (20)
	2-Butanone	100	83.8	ug/L	84	5		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	20.0	ug/L	100	4		70 (77)	130 (113)	20 (15)
	Chloroform	20	19.3	ug/L	97	4		70 (79)	130 (113)	20 (20)
	Benzene	20	20.4	ug/L	102	2		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	20.1	ug/L	101	4		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	21.2	ug/L	106	1		70 (77)	130 (113)	20 (15)
	Tetrachloroethene	20	21.2	ug/L	106	3		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	21.4	ug/L	107	1		70 (82)	130 (109)	20 (15)