

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: FIRS02
 Lab Code: CHEM Case No.: Q2320 SAS No.: Q2320 SDG NO.: Q2320
 Lab File ID: VN087045.D Date Analyzed: 06/17/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	178887	8.23	312096	9.11	278021	11.87
UPPER LIMIT	357774	8.73	624192	9.606	556042	12.365
LOWER LIMIT	89443.5	7.73	156048	8.606	139011	11.365
EPA SAMPLE NO.						
WC	195874	8.23	390146	9.11	350367	11.87
VN0617WBL01	200915	8.23	403651	9.11	368689	11.87
VN0617WBS01	157438	8.23	281588	9.11	248700	11.87
VN0617WBSD01	171983	8.23	303445	9.11	262555	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.