

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2329 SAS No.: Q2329 SDG No.: Q2329

Instrument ID: MSVOA_N Calibration Date/Time: 06/18/2025 19:42

Lab File ID: VN087098.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.532		6.61	50
Chloromethane	0.644	0.586	0.1	-9.01	50
Vinyl Chloride	0.664	0.742		11.75	50
Bromomethane	0.372	0.385		3.49	50
Chloroethane	0.429	0.473		10.26	50
Trichlorofluoromethane	0.868	0.960		10.6	50
1,1,2-Trichlorotrifluoroethane	0.545	0.597		9.54	50
1,1-Dichloroethene	0.557	0.608		9.16	50
Acetone	0.355	0.325		-8.45	50
Carbon Disulfide	1.539	1.739		12.99	50
Methyl tert-butyl Ether	2.015	2.066		2.53	50
Methyl Acetate	1.035	1.067		3.09	50
Methylene Chloride	0.665	0.685		3.01	50
trans-1,2-Dichloroethene	0.619	0.648		4.68	50
1,1-Dichloroethane	1.120	1.175	0.1	4.91	50
Cyclohexane	1.086	1.032		-4.97	50
2-Butanone	0.577	0.534		-7.45	50
Carbon Tetrachloride	0.433	0.464		7.16	50
cis-1,2-Dichloroethene	0.740	0.782		5.68	50
Bromochloromethane	0.550	0.510		-7.27	50
Chloroform	1.118	1.163		4.03	50
1,1,1-Trichloroethane	0.951	0.980		3.05	50
Methylcyclohexane	0.605	0.542		-10.41	50
Benzene	1.444	1.508		4.43	50
1,2-Dichloroethane	0.438	0.451		2.97	50
Trichloroethene	0.342	0.363		6.14	50
1,2-Dichloropropane	0.351	0.365		3.99	50
Bromodichloromethane	0.480	0.499		3.96	50
4-Methyl-2-Pentanone	0.543	0.534		-1.66	50
Toluene	0.882	0.940		6.58	50
t-1,3-Dichloropropene	0.537	0.552		2.79	50
cis-1,3-Dichloropropene	0.574	0.600		4.53	50
1,1,2-Trichloroethane	0.340	0.357		5	50
2-Hexanone	0.350	0.358		2.29	50
Dibromochloromethane	0.354	0.390		10.17	50
1,2-Dibromoethane	0.348	0.364		4.6	50
Tetrachloroethene	0.316	0.324		2.53	50
Chlorobenzene	1.103	1.153	0.3	4.53	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2329 SAS No.: Q2329 SDG No.: Q2329

Instrument ID: MSVOA_N Calibration Date/Time: 06/18/2025 19:42

Lab File ID: VN087098.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	1.939		2.05	50
m/p-Xylenes	0.727	0.770		5.91	50
o-Xylene	0.696	0.743		6.75	50
Styrene	1.192	1.271		6.63	50
Bromoform	0.263	0.294	0.1	11.79	50
Isopropylbenzene	3.643	3.579		-1.76	50
1,1,2,2-Tetrachloroethane	1.234	1.250	0.3	1.3	50
1,3-Dichlorobenzene	1.644	1.670		1.58	50
1,4-Dichlorobenzene	1.676	1.718		2.51	50
1,2-Dichlorobenzene	1.579	1.620		2.6	50
1,2-Dibromo-3-Chloropropane	0.295	0.271		-8.14	50
1,2,4-Trichlorobenzene	1.008	0.875		-13.19	50
1,2,3-Trichlorobenzene	1.002	0.856		-14.57	50
1,2-Dichloroethane-d4	0.669	0.687		2.69	50
Dibromofluoromethane	0.296	0.331		11.82	50
Toluene-d8	1.173	1.212		3.33	50
4-Bromofluorobenzene	0.436	0.455		4.36	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.