

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087076.D
 Acq On : 18 Jun 2025 11:46
 Operator : JC\MD
 Sample : VN0618WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0618WBL01

Quant Time: Jun 19 01:24:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	270298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	541331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	486295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	227513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	190055	52.516	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.040%
35) Dibromofluoromethane	8.177	113	162627	50.693	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.380%
50) Toluene-d8	10.570	98	664885	52.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.700%
62) 4-Bromofluorobenzene	12.847	95	232316	49.237	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087076.D
Acq On : 18 Jun 2025 11:46
Operator : JC\MD
Sample : VN0618WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01

Quant Time: Jun 19 01:24:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 110
Delta R.T. -0.007 min
Lab File: VN087076.D
Acq: 18 Jun 2025 11:46




Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01


Tgt Ion:168 Resp: 270298
Ion Ratio Lower Upper
168 100
99 57.4 49.1 73.7



#33
1,2-Dichloroethane-d4
Concen: 52.516 ug/l
RT: 8.582 min Scan# 1127
Delta R.T. -0.006 min
Lab File: VN087076.D
Acq: 18 Jun 2025 11:46

Tgt Ion: 65 Resp: 190055
Ion Ratio Lower Upper
65 100
67 55.3 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN087076.D

Acq: 18 Jun 2025 11:46

Instrument :

MSVOA_N

ClientSampleId :

VN0618WBL01





Tgt Ion:114 Resp: 541331

Ion Ratio Lower Upper

114 100

63 19.6 0.0 39.6

88 15.6 0.0 30.2



#35

Dibromofluoromethane

Concen: 50.693 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN087076.D

Acq: 18 Jun 2025 11:46




Tgt Ion:113 Resp: 162627


Ion Ratio Lower Upper

113 100

111 103.1 84.2 126.2

192 17.4 14.2 21.4









#50
Toluene-d8
Concen: 52.354 ug/l
RT: 10.570 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN087076.D
Acq: 18 Jun 2025 11:46




Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01


Tgt Ion: 98 Resp: 664885
Ion Ratio Lower Upper
98 100
100 66.3 53.4 80.0



#62
4-Bromofluorobenzene
Concen: 49.237 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN087076.D
Acq: 18 Jun 2025 11:46

Tgt Ion: 95 Resp: 232316
Ion Ratio Lower Upper
95 100
174 72.0 0.0 141.8
176 68.5 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN087076.D
Acq: 18 Jun 2025 11:46

Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01

Tgt Ion:117 Resp: 486295
Ion Ratio Lower Upper
117 100
82 54.4 44.6 67.0
119 31.9 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.006 min
Lab File: VN087076.D
Acq: 18 Jun 2025 11:46

Tgt Ion:152 Resp: 227513
Ion Ratio Lower Upper
152 100
115 60.0 30.1 90.5
150 157.2 0.0 345.0

