

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134155
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247
CCC	VP134156
Internal Standard/PEM	
ICV/I.BLK	VP134248
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086861.D	06 Jun 2025 07:59		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN086862.D	06 Jun 2025 12:44	Method failed for com.#13	JC\MD	Ok,M
3	VSTDIC005	VSTDIC005	VN086863.D	06 Jun 2025 13:17		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN086864.D	06 Jun 2025 13:40		JC\MD	Ok,M
5	VSTDIC050	VSTDIC050	VN086865.D	06 Jun 2025 14:03		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN086866.D	06 Jun 2025 14:26		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN086867.D	06 Jun 2025 14:49		JC\MD	Ok,M
8	IBLK	IBLK	VN086868.D	06 Jun 2025 15:12		JC\MD	Ok
9	VSTDICV050	ICVVN060625	VN086869.D	06 Jun 2025 15:54		JC\MD	Ok,M
10	VN0606WBL01	VN0606WBL01	VN086870.D	06 Jun 2025 16:47		JC\MD	Ok
11	VN0606WBL02	VN0606WBL02	VN086871.D	06 Jun 2025 17:10		JC\MD	Ok
12	VN0606WBS01	VN0606WBS01	VN086872.D	06 Jun 2025 17:33		JC\MD	Ok,M
13	VN0606WBSD01	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56		JC\MD	Ok,M
14	Q2254-01	BP-VPB-182-GW-810-8	VN086874.D	06 Jun 2025 18:19	vial A pH<2 endccc out of tune	JC\MD	Not Ok
15	Q2237-02	TW-WTS-10	VN086875.D	06 Jun 2025 18:42	vial A pH<2	JC\MD	Ok
16	Q2216-02	3887	VN086876.D	06 Jun 2025 19:05	vial A pH<2	JC\MD	Ok
17	Q2216-03	3888	VN086877.D	06 Jun 2025 19:28	vial A pH<2	JC\MD	Ok
18	Q2216-04	3864	VN086878.D	06 Jun 2025 19:51	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134155		
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC	VP134156		
Internal Standard/PEM			
ICV/I.BLK	VP134248		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2216-05	3865	VN086879.D	06 Jun 2025 20:13	vial A pH<2 Out of Tune	JC\MD	Not Ok
20	Q2216-06	3851	VN086880.D	06 Jun 2025 20:36	vial A pH<2 Out of Tune	JC\MD	Not Ok
21	Q2206-04	TP-1	VN086881.D	06 Jun 2025 20:59	vial A pH<2 Out of Tune	JC\MD	Not Ok
22	Q2242-04	TP09-MHJ	VN086882.D	06 Jun 2025 21:21	vial A pH<2 Out of Tune	JC\MD	Not Ok
23	Q2192-01	SB-1	VN086883.D	06 Jun 2025 21:44	vial A pH<2 Out of Tune	JC\MD	Not Ok
24	Q2198-02	B-202-SB02	VN086884.D	06 Jun 2025 22:07	vial A pH<2 Out of Tune	JC\MD	Not Ok
25	Q2198-04	B-207-SB02	VN086885.D	06 Jun 2025 22:29	vial A pH<2 Out of Tune	JC\MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VN086886.D	06 Jun 2025 22:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 3:15:59 PM
Supervise By	Maresh Dadoda	Supervise On	6/18/2025 6:19:51 PM
SubDirectory	VN061725	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134375
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134376,VP134377

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087044.D	17 Jun 2025 08:51		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087045.D	17 Jun 2025 09:25	pH#Lot#V12668	JC\MD	Ok,M
3	VN0617MBL01	VN0617MBL01	VN087046.D	17 Jun 2025 09:59		JC\MD	Ok
4	VN0617WBL01	VN0617WBL01	VN087047.D	17 Jun 2025 10:20		JC\MD	Ok
5	VN0617MBS01	VN0617MBS01	VN087048.D	17 Jun 2025 10:41		JC\MD	Ok,M
6	Q2307-03	LINDEN-SAA-WATER	VN087049.D	17 Jun 2025 11:16	vial A pH<2	JC\MD	Ok
7	Q2286-01ME	BP-VPB-182-TB-20250	VN087050.D	17 Jun 2025 11:37		JC\MD	Ok
8	VN0617WBS01	VN0617WBS01	VN087051.D	17 Jun 2025 11:59		JC\MD	Ok,M
9	VN0617WBSD01	VN0617WBSD01	VN087052.D	17 Jun 2025 12:35		JC\MD	Ok,M
10	Q2299-16DL	RE134D4-20250610DL	VN087053.D	17 Jun 2025 12:56	vial B pH<2	JC\MD	Ok
11	Q2299-18DL	TT190D1-20250611DL	VN087054.D	17 Jun 2025 13:18	vial B pH<2	JC\MD	Ok
12	Q2299-17DL	RE134D3-20250610DL	VN087055.D	17 Jun 2025 13:39	vial B pH<2	JC\MD	Ok
13	PB168498TB	PB168498TB	VN087056.D	17 Jun 2025 14:00	vial A pH#5.0	JC\MD	Ok
14	Q2311-04	TP03-MH2MH3-WC	VN087057.D	17 Jun 2025 14:22	vial A pH#6.0	JC\MD	Ok
15	Q2311-08	TP04-MH4-WC	VN087058.D	17 Jun 2025 14:43	vial A pH 7	JC\MD	Ok
16	Q2312-04	TP-1	VN087059.D	17 Jun 2025 15:04	vial A pH 5	JC\MD	Ok
17	Q2319-04	MH-B	VN087060.D	17 Jun 2025 15:26	vial A pH 5	JC\MD	Ok
18	Q2320-01	WC	VN087061.D	17 Jun 2025 15:47	vial A pH#5.0	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 3:15:59 PM
Supervise By	Mahesh Dadoda	Supervise On	6/18/2025 6:19:51 PM
SubDirectory	VN061725	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134375		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134376,VP134377		

19	Q2325-04	TP-8	VN087062.D	17 Jun 2025 16:08	vial A pH 5	JC\MD	Ok,M
20	Q2329-01	TT192D1-20250611	VN087063.D	17 Jun 2025 16:30	vial A pH<2	JC\MD	Ok,M
21	Q2329-02	TT188S1-20250612	VN087064.D	17 Jun 2025 16:51	vial A pH<2	JC\MD	Ok
22	Q2329-03	TT119S1-20250612	VN087065.D	17 Jun 2025 17:13	vial A pH<2	JC\MD	Ok
23	Q2329-04	FB01-20250613	VN087066.D	17 Jun 2025 17:34	vial A pH<2 FB; Hit of comp.#16,20	JC\MD	ReRun
24	Q2329-05	TT162S1-20250613	VN087067.D	17 Jun 2025 17:56	vial A pH<2	JC\MD	Ok
25	Q2329-06	BPOW6-7-20250613	VN087068.D	17 Jun 2025 18:17	vial A pH<2	JC\MD	Ok
26	Q2299-05	TT191D1-20250609	VN087069.D	17 Jun 2025 18:39	vial A pH<2	JC\MD	Ok
27	Q2299-03MS	TT191D1-20250609MS	VN087070.D	17 Jun 2025 19:00	vial A pH<2 Recovery failed	JC\MD	Not Ok
28	Q2299-04MSD	TT191D1-20250609MS	VN087071.D	17 Jun 2025 19:22	vial A pH<2	JC\MD	Ok,M
29	VSTDCCC050	VSTDCCC050EC	VN087072.D	17 Jun 2025 19:43		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061825

Review By	John Carlone	Review On	6/19/2025 9:02:07 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:14:25 AM
SubDirectory	VN061825	HP Acquire Method	MSVOA_N HP Processing Method 82n060625w.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134412
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134414,VP134415 VP133935

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087073.D	18 Jun 2025 08:43		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087074.D	18 Jun 2025 10:21	pH#Lot#V12668	JC\MD	Ok,M
3	VN0618MBL01	VN0618MBL01	VN087075.D	18 Jun 2025 11:01		JC\MD	Ok
4	VN0618WBL01	VN0618WBL01	VN087076.D	18 Jun 2025 11:46		JC\MD	Ok
5	VN0618WBS01	VN0618WBS01	VN087077.D	18 Jun 2025 12:07		JC\MD	Ok,M
6	VN0618WBSD01	VN0618WBSD01	VN087078.D	18 Jun 2025 12:29		JC\MD	Ok,M
7	PB168515TB	PB168515TB	VN087079.D	18 Jun 2025 12:51		JC\MD	Ok
8	Q2299-02	RE117D2-20250609	VN087080.D	18 Jun 2025 13:13	vial A pH<2	JC\MD	Ok
9	Q2339-04	MH-C	VN087081.D	18 Jun 2025 13:34	vial A pH#5.0	JC\MD	Ok
10	Q2340-04	TP05-MH17A-WC	VN087082.D	18 Jun 2025 13:56	vial A pH#5.0	JC\MD	Ok
11	Q2341-04	TP-9	VN087083.D	18 Jun 2025 14:18	vial A pH#5.0	JC\MD	Ok
12	Q2341-08	EP-3	VN087084.D	18 Jun 2025 14:40	vial A pH#5.0	JC\MD	Ok
13	Q2347-04	TP-10	VN087085.D	18 Jun 2025 15:02	vial A pH#5.0	JC\MD	Ok
14	Q2329-07	BPOW6-8-20250613	VN087086.D	18 Jun 2025 15:24	vial A pH<2	JC\MD	Ok
15	Q2329-08	BPOW6-11-20250613	VN087087.D	18 Jun 2025 15:46	vial A pH<2	JC\MD	Ok
16	Q2329-09	TT149S1-20250613	VN087088.D	18 Jun 2025 16:08	vial A pH<2	JC\MD	Ok
17	Q2329-13	DUP03-20250613	VN087089.D	18 Jun 2025 16:30	vial A pH<2	JC\MD	Ok
18	Q2329-14	DUP04-20250613	VN087090.D	18 Jun 2025 16:52	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061825

Review By	John Carlone	Review On	6/19/2025 9:02:07 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:14:25 AM
SubDirectory	VN061825	HP Acquire Method	MSVOA_N HP Processing Method 82n060625w.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134412		
CCC	VP134414,VP134415		
Internal Standard/PEM	VP133935		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2329-15	EB01-20250613	VN087091.D	18 Jun 2025 17:14	vial A pH<2 EB;Hit of comp.#16,20	JC\MD	ReRun
20	Q2329-16	TB01-20250612	VN087092.D	18 Jun 2025 17:36	vial A pH<2 TB	JC\MD	Ok
21	Q2329-17	TB02-20250612	VN087093.D	18 Jun 2025 17:57	vial A pH<2 TB	JC\MD	Ok
22	Q2329-18	TT163S1-20250613	VN087094.D	18 Jun 2025 18:18	vial A pH<2	JC\MD	Ok
23	Q2329-04	FB01-20250613	VN087095.D	18 Jun 2025 18:39	vial B pH<2 FB	JC\MD	Ok
24	Q2299-03MS	TT191D1-20250609MS	VN087096.D	18 Jun 2025 19:00	vial B pH<2	JC\MD	Ok,M
25	VIBLK	VIBLK	VN087097.D	18 Jun 2025 19:21		JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN087098.D	18 Jun 2025 19:42		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061925

Review By	John Carlone	Review On	6/20/2025 8:50:20 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:24:10 AM
SubDirectory	VN061925	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134425
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134426,VP134427

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087099.D	19 Jun 2025 08:58		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087100.D	19 Jun 2025 09:31	pH#Lot#V12668	JC\MD	Ok,M
3	VN0619MBL01	VN0619MBL01	VN087101.D	19 Jun 2025 10:05		JC\MD	Ok
4	VN0619WBL01	VN0619WBL01	VN087102.D	19 Jun 2025 10:26		JC\MD	Ok
5	Q2329-15	EB01-20250613	VN087103.D	19 Jun 2025 10:47	vial B pH<2 EB	JC\MD	Ok
6	VN0619WBS01	VN0619WBS01	VN087104.D	19 Jun 2025 12:22		JC\MD	Ok,M
7	VN0619WBSD01	VN0619WBSD01	VN087105.D	19 Jun 2025 12:56	Recovery fail	JC\MD	Not Ok
8	Q2333-01	MW-06	VN087106.D	19 Jun 2025 13:17	vial A pH<2	JC\MD	Ok,M
9	Q2333-02	MW-08	VN087107.D	19 Jun 2025 13:38	vial A pH<2 Need 20x	JC\MD	Dilution
10	Q2333-03	MW-10	VN087108.D	19 Jun 2025 13:59	vial A pH<2 E flag of previous sample	JC\MD	ReRun
11	Q2333-04	MW-11	VN087109.D	19 Jun 2025 14:21	vial A pH<2	JC\MD	Ok
12	Q2333-05	MW-13	VN087110.D	19 Jun 2025 14:42	vial A pH<2	JC\MD	Ok
13	Q2333-06	MW-12	VN087111.D	19 Jun 2025 15:03	vial A pH<2 Need 10x	JC\MD	Dilution
14	IBLK	IBLK	VN087112.D	19 Jun 2025 15:24		JC\MD	Ok
15	IBLK	IBLK	VN087113.D	19 Jun 2025 15:45		JC\MD	Ok
16	Q2333-02DL	MW-08DL	VN087114.D	19 Jun 2025 16:07	vial B pH<2	JC\MD	Ok
17	Q2333-03	MW-10	VN087115.D	19 Jun 2025 16:28	vial B pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061925

Review By	John Carlone	Review On	6/20/2025 8:50:20 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:24:10 AM
SubDirectory	VN061925	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134425		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134426,VP134427		

18	Q2333-06DL	MW-12DL	VN087116.D	19 Jun 2025 16:49	vial B pH<2	JC\MD	Ok,M
19	Q2329-10	TT172S1-20250613	VN087117.D	19 Jun 2025 17:10	vial A pH<2	JC\MD	Ok
20	Q2329-11MS	TT172S1-20250613MS	VN087118.D	19 Jun 2025 17:31	vial A pH<2	JC\MD	Ok,M
21	Q2329-12MSD	TT172S1-20250613MS	VN087119.D	19 Jun 2025 17:52	vial A pH<2	JC\MD	Ok,M
22	VSTDCCC050	VSTDCCC050EC	VN087120.D	19 Jun 2025 18:13		JC\MD	Ok,M

M : Manual Integration