

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2329 SAS No.: Q2329 SDG NO.: Q2329
Lab File ID: VN087074.D Date Analyzed: 06/18/2025
Instrument ID: MSVOA_N Time Analyzed: 10:21
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	164001	13.788				
UPPER LIMIT	328002	14.288				
LOWER LIMIT	82000.5	13.288				
EPA SAMPLE NO.						
FB01-20250613	153113	13.79				
BPOW6-8-20250613	166828	13.79				
BPOW6-11-20250613	143673	13.79				
TT149S1-20250613	210865	13.79				
DUP03-20250613	235183	13.79				
DUP04-20250613	162129	13.79				
TB01-20250612	158047	13.79				
TB02-20250612	154445	13.79				
TT163S1-20250613	141945	13.79				
VN0618WBL01	227513	13.79				
VN0618WBS01	167760	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.