

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	06/12/25
Project:	RFK Bridge RMB-Randall Island	Date Received:	06/13/25
Client Sample ID:	MW-13	SDG No.:	Q2333
Lab Sample ID:	Q2333-05	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	970 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024998.D	1	06/17/25 09:25	06/18/25 17:30	PB168509

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.20	U	0.77	5.20	ug/L
83-32-9	Acenaphthene	5.20	U	0.57	5.20	ug/L
86-73-7	Fluorene	5.20	U	0.65	5.20	ug/L
85-01-8	Phenanthrene	5.20	U	0.52	5.20	ug/L
120-12-7	Anthracene	5.20	U	0.63	5.20	ug/L
206-44-0	Fluoranthene	5.20	U	0.85	5.20	ug/L
129-00-0	Pyrene	5.20	U	0.52	5.20	ug/L
56-55-3	Benzo(a)anthracene	5.20	U	0.46	5.20	ug/L
218-01-9	Chrysene	5.20	U	0.45	5.20	ug/L
205-99-2	Benzo(b)fluoranthene	5.20	U	0.51	5.20	ug/L
207-08-9	Benzo(k)fluoranthene	5.20	U	0.49	5.20	ug/L
50-32-8	Benzo(a)pyrene	5.20	U	0.57	5.20	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.20	U	0.61	5.20	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.20	U	0.69	5.20	ug/L
191-24-2	Benzo(g,h,i)perylene	5.20	U	0.71	5.20	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	87.3		67 - 132	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	84.8		52 - 132	85%	SPK: 100
1718-51-0	Terphenyl-d14	81.3		42 - 152	81%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	278000	7.607			
1146-65-2	Naphthalene-d8	1100000	10.378			
15067-26-2	Acenaphthene-d10	687000	14.254			
1517-22-2	Phenanthrene-d10	1380000	17.066			
1719-03-5	Chrysene-d12	1450000	21.495			
1520-96-3	Perylene-d12	1600000	24.759			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	06/12/25
Project:	RFK Bridge RMB-Randall Island	Date Received:	06/13/25
Client Sample ID:	MW-13	SDG No.:	Q2333
Lab Sample ID:	Q2333-05	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	970 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024998.D	1	06/17/25 09:25	06/18/25 17:30	PB168509

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products