

Surrogate Summary

SW-846

SDG No.: Q2333

Client: LiRo Engineers, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168509BL	PB168509BL	Nitrobenzene-d5	100	84.1	84		67	132
		2-Fluorobiphenyl	100	81.5	81		52	132
		Terphenyl-d14	100	84.2	84		42	152
PB168509BS	PB168509BS	Nitrobenzene-d5	100	84.7	85		67	132
		2-Fluorobiphenyl	100	85.5	86		52	132
		Terphenyl-d14	100	88.9	89		42	152
PB168509BSD	PB168509BSD	Nitrobenzene-d5	100	79.6	80		67	132
		2-Fluorobiphenyl	100	77.8	78		52	132
		Terphenyl-d14	100	85.9	86		42	152
Q2333-01	MW-06	Nitrobenzene-d5	100	89.1	89		67	132
		2-Fluorobiphenyl	100	84.0	84		52	132
		Terphenyl-d14	100	70.2	70		42	152
Q2333-02	MW-08	Nitrobenzene-d5	100	87.7	88		67	132
		2-Fluorobiphenyl	100	79.7	80		52	132
		Terphenyl-d14	100	81.7	82		42	152
Q2333-03	MW-10	Nitrobenzene-d5	100	87.3	87		67	132
		2-Fluorobiphenyl	100	81.9	82		52	132
		Terphenyl-d14	100	69.4	69		42	152
Q2333-04	MW-11	Nitrobenzene-d5	100	91.5	91		67	132
		2-Fluorobiphenyl	100	84.9	85		52	132
		Terphenyl-d14	100	68.3	68		42	152
Q2333-05	MW-13	Nitrobenzene-d5	100	87.3	87		67	132
		2-Fluorobiphenyl	100	84.8	85		52	132
		Terphenyl-d14	100	81.3	81		42	152
Q2333-06	MW-12	Nitrobenzene-d5	100	78.8	79		67	132
		2-Fluorobiphenyl	100	74.6	75		52	132
		Terphenyl-d14	100	75.2	75		42	152