

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2333

Client: LiRo Engineers, Inc.

Analytical Method: 8270E

DataFile: BP024989.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168509BS	Acenaphthylene	50	49.0	ug/L	98				79	103	
	Acenaphthene	50	48.5	ug/L	97				59	113	
	Fluorene	50	48.5	ug/L	97				64	107	
	Phenanthrene	50	47.9	ug/L	96				62	109	
	Anthracene	50	48.9	ug/L	98				65	110	
	Fluoranthene	50	47.8	ug/L	96				64	110	
	Pyrene	50	50.5	ug/L	101				71	103	
	Benzo(a)anthracene	50	49.9	ug/L	100				62	107	
	Chrysene	50	49.8	ug/L	100				61	108	
	Benzo(b)fluoranthene	50	48.3	ug/L	97				77	113	
	Benzo(k)fluoranthene	50	49.9	ug/L	100				77	105	
	Benzo(a)pyrene	50	49.3	ug/L	99				72	131	
	Indeno(1,2,3-cd)pyrene	50	48.5	ug/L	97				72	105	
	Dibenz(a,h)anthracene	50	48.9	ug/L	98				78	115	
	Benzo(g,h,i)perylene	50	47.7	ug/L	95				75	118	