

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2333

Client: LiRo Engineers, Inc.

Analytical Method: 8270E DataFile: BP024990.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		RPD
							Qual	Qual	Low	High	
PB168509BSD	Acenaphthylene	50	45.4	ug/L	91	8			79	103	20
	Acenaphthene	50	45.0	ug/L	90	7			59	113	20
	Fluorene	50	44.9	ug/L	90	8			64	107	20
	Phenanthrene	50	46.2	ug/L	92	4			62	109	20
	Anthracene	50	46.6	ug/L	93	5			65	110	20
	Fluoranthene	50	45.7	ug/L	91	4			64	110	20
	Pyrene	50	50.1	ug/L	100	1			71	103	20
	Benzo(a)anthracene	50	47.6	ug/L	95	5			62	107	20
	Chrysene	50	46.0	ug/L	92	8			61	108	20
	Benzo(b)fluoranthene	50	47.7	ug/L	95	1			77	113	20
	Benzo(k)fluoranthene	50	46.5	ug/L	93	7			77	105	20
	Benzo(a)pyrene	50	47.8	ug/L	96	3			72	131	20
	Indeno(1,2,3-cd)pyrene	50	48.2	ug/L	96	1			72	105	20
	Dibenz(a,h)anthracene	50	48.6	ug/L	97	1			78	115	20
	Benzo(g,h,i)perylene	50	48.0	ug/L	96	1			75	118	20