

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2333 SAS No.: Q2333 SDG NO.: Q2333

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/18/2025

Lab File ID: BP024987.D Time Analyzed: 09:55

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	287169	7.607	1150640	10.38	726300	14.26
UPPER LIMIT	574338	8.107	2301280	10.878	1452600	14.76
LOWER LIMIT	143585	7.107	575320	9.878	363150	13.76
EPA SAMPLE NO.						
01 MW-08	284799	7.60	1178300	10.37	818800	14.25
02 MW-13	277507	7.61	1098580	10.38	686732	14.25
03 MW-12	259382	7.60	1022680	10.38	715185	14.26
04 PB168509BL	278843	7.61	1046610	10.38	632257	14.26
05 PB168509BS	309420	7.61	1193310	10.38	709072	14.25
06 PB168509BSD	299910	7.60	1166190	10.37	733128	14.25

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.