

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134155
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247
CCC	VP134156
Internal Standard/PEM	
ICV/I.BLK	VP134248
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086861.D	06 Jun 2025 07:59		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN086862.D	06 Jun 2025 12:44	Method failed for com.#13	JC\MD	Ok,M
3	VSTDIC005	VSTDIC005	VN086863.D	06 Jun 2025 13:17		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN086864.D	06 Jun 2025 13:40		JC\MD	Ok,M
5	VSTDIC050	VSTDIC050	VN086865.D	06 Jun 2025 14:03		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN086866.D	06 Jun 2025 14:26		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN086867.D	06 Jun 2025 14:49		JC\MD	Ok,M
8	IBLK	IBLK	VN086868.D	06 Jun 2025 15:12		JC\MD	Ok
9	VSTDICV050	ICVVN060625	VN086869.D	06 Jun 2025 15:54		JC\MD	Ok,M
10	VN0606WBL01	VN0606WBL01	VN086870.D	06 Jun 2025 16:47		JC\MD	Ok
11	VN0606WBL02	VN0606WBL02	VN086871.D	06 Jun 2025 17:10		JC\MD	Ok
12	VN0606WBS01	VN0606WBS01	VN086872.D	06 Jun 2025 17:33		JC\MD	Ok,M
13	VN0606WBSD01	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56		JC\MD	Ok,M
14	Q2254-01	BP-VPB-182-GW-810-8	VN086874.D	06 Jun 2025 18:19	vial A pH<2 endccc out of tune	JC\MD	Not Ok
15	Q2237-02	TW-WTS-10	VN086875.D	06 Jun 2025 18:42	vial A pH<2	JC\MD	Ok
16	Q2216-02	3887	VN086876.D	06 Jun 2025 19:05	vial A pH<2	JC\MD	Ok
17	Q2216-03	3888	VN086877.D	06 Jun 2025 19:28	vial A pH<2	JC\MD	Ok
18	Q2216-04	3864	VN086878.D	06 Jun 2025 19:51	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134155 VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC Internal Standard/PEM	VP134156		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134248		

19	Q2216-05	3865	VN086879.D	06 Jun 2025 20:13	vial A pH<2 Out of Tune	JC\MD	Not Ok
20	Q2216-06	3851	VN086880.D	06 Jun 2025 20:36	vial A pH<2 Out of Tune	JC\MD	Not Ok
21	Q2206-04	TP-1	VN086881.D	06 Jun 2025 20:59	vial A pH<2 Out of Tune	JC\MD	Not Ok
22	Q2242-04	TP09-MHJ	VN086882.D	06 Jun 2025 21:21	vial A pH<2 Out of Tune	JC\MD	Not Ok
23	Q2192-01	SB-1	VN086883.D	06 Jun 2025 21:44	vial A pH<2 Out of Tune	JC\MD	Not Ok
24	Q2198-02	B-202-SB02	VN086884.D	06 Jun 2025 22:07	vial A pH<2 Out of Tune	JC\MD	Not Ok
25	Q2198-04	B-207-SB02	VN086885.D	06 Jun 2025 22:29	vial A pH<2 Out of Tune	JC\MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VN086886.D	06 Jun 2025 22:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061925

Review By	John Carlone	Review On	6/20/2025 8:50:20 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:24:10 AM
SubDirectory	VN061925	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134425
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134426,VP134427

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087099.D	19 Jun 2025 08:58		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087100.D	19 Jun 2025 09:31	pH#Lot#V12668	JC\MD	Ok,M
3	VN0619MBL01	VN0619MBL01	VN087101.D	19 Jun 2025 10:05		JC\MD	Ok
4	VN0619WBL01	VN0619WBL01	VN087102.D	19 Jun 2025 10:26		JC\MD	Ok
5	Q2329-15	EB01-20250613	VN087103.D	19 Jun 2025 10:47	vial B pH<2 EB	JC\MD	Ok
6	VN0619WBS01	VN0619WBS01	VN087104.D	19 Jun 2025 12:22		JC\MD	Ok,M
7	VN0619WBSD01	VN0619WBSD01	VN087105.D	19 Jun 2025 12:56	Recovery fail	JC\MD	Not Ok
8	Q2333-01	MW-06	VN087106.D	19 Jun 2025 13:17	vial A pH<2	JC\MD	Ok,M
9	Q2333-02	MW-08	VN087107.D	19 Jun 2025 13:38	vial A pH<2 Need 20x	JC\MD	Dilution
10	Q2333-03	MW-10	VN087108.D	19 Jun 2025 13:59	vial A pH<2 E flag of previous sample	JC\MD	ReRun
11	Q2333-04	MW-11	VN087109.D	19 Jun 2025 14:21	vial A pH<2	JC\MD	Ok
12	Q2333-05	MW-13	VN087110.D	19 Jun 2025 14:42	vial A pH<2	JC\MD	Ok
13	Q2333-06	MW-12	VN087111.D	19 Jun 2025 15:03	vial A pH<2 Need 10x	JC\MD	Dilution
14	IBLK	IBLK	VN087112.D	19 Jun 2025 15:24		JC\MD	Ok
15	IBLK	IBLK	VN087113.D	19 Jun 2025 15:45		JC\MD	Ok
16	Q2333-02DL	MW-08DL	VN087114.D	19 Jun 2025 16:07	vial B pH<2	JC\MD	Ok
17	Q2333-03	MW-10	VN087115.D	19 Jun 2025 16:28	vial B pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061925

Review By	John Carlone	Review On	6/20/2025 8:50:20 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:24:10 AM
SubDirectory	VN061925	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134425		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134426,VP134427		

18	Q2333-06DL	MW-12DL	VN087116.D	19 Jun 2025 16:49	vial B pH<2	JC\MD	Ok,M
19	Q2329-10	TT172S1-20250613	VN087117.D	19 Jun 2025 17:10	vial A pH<2	JC\MD	Ok
20	Q2329-11MS	TT172S1-20250613MS	VN087118.D	19 Jun 2025 17:31	vial A pH<2	JC\MD	Ok,M
21	Q2329-12MSD	TT172S1-20250613MS	VN087119.D	19 Jun 2025 17:52	vial A pH<2	JC\MD	Ok,M
22	VSTDCCC050	VSTDCCC050EC	VN087120.D	19 Jun 2025 18:13		JC\MD	Ok,M

M : Manual Integration