

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2333

Client: LiRo Engineers, Inc.

Analytical Method: SW8260-Low

Datafile : VN087104.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0619WBS01	Methyl tert-butyl Ether	20	18.1	ug/L	91			78	114	
	Benzene	20	19.2	ug/L	96			82	109	
	Toluene	20	19.5	ug/L	98			82	110	
	Ethyl Benzene	20	18.6	ug/L	93			83	109	
	m/p-Xylenes	40	38.5	ug/L	96			82	110	
	o-Xylene	20	18.7	ug/L	94			83	109	
	Isopropylbenzene	20	18.1	ug/L	91			83	112	
	N-propylbenzene	20	17.7	ug/L	89			83	112	
	1,3,5-Trimethylbenzene	20	18.2	ug/L	91			85	112	
	tert-Butylbenzene	20	16.6	ug/L	83			83	112	
	1,2,4-Trimethylbenzene	20	18.1	ug/L	91			85	111	
	Sec-butylbenzene	20	16.4	ug/L	82			81	114	
	p-Isopropyltoluene	20	16.5	ug/L	83			78	116	
	n-Butylbenzene	20	15.1	ug/L	76			75	115	
	Naphthalene	20	16.4	ug/L	82			78	119	