

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
 Lab Code: CHEM Case No.: Q2333 SAS No.: Q2333 SDG NO.: Q2333
 Lab File ID: VN087100.D Date Analyzed: 06/19/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	126484	13.788				
UPPER LIMIT	252968	14.288				
LOWER LIMIT	63242	13.288				
EPA SAMPLE NO.						
TT172S1-20250613MS	145576	13.79				
TT172S1-20250613MSD	142178	13.79				
MW-06	131506	13.79				
MW-08	201938	13.79				
MW-08DL	162105	13.79				
MW-10	152345	13.79				
MW-11	174687	13.79				
MW-13	169301	13.79				
MW-12	153356	13.79				
MW-12DL	175112	13.79				
VN0619WBL01	156562	13.79				
VN0619WBS01	132931	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.