



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	06/12/25 13:45
Project:	Buff	Date Received:	06/13/25
Client Sample ID:	MW3	SDG No.:	Q2334
Lab Sample ID:	Q2334-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Alkalinity	122		1	1.00	2.00	mg/L		06/19/25 12:24	SM 2320 B-21
Nitrate	0.18	J	1	0.095	0.50	mg/L		06/13/25 17:54	300.0
Sulfate	4.70		1	0.46	3.00	mg/L		06/13/25 17:54	300.0
COD	14.4		1	1.50	10.0	mg/L		06/19/25 13:32	SM 5220 D-11

Comments: The alkalinity to pH 4.31=122 mg CaCO3/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	06/12/25 12:45
Project:	Buff	Date Received:	06/13/25
Client Sample ID:	MW4	SDG No.:	Q2334
Lab Sample ID:	Q2334-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Alkalinity	104		1	1.00	2.00	mg/L		06/19/25 12:34	SM 2320 B-21
Nitrate	0.43	J	1	0.095	0.50	mg/L		06/13/25 18:58	300.0
Sulfate	16.2		1	0.46	3.00	mg/L		06/13/25 18:58	300.0
COD	15.4		1	1.50	10.0	mg/L		06/19/25 13:34	SM 5220 D-11

Comments: The alkalinity to pH 4.30=104 mg CaCO₃/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	06/12/25 13:00
Project:	Buff	Date Received:	06/13/25
Client Sample ID:	GBTW1	SDG No.:	Q2334
Lab Sample ID:	Q2334-03	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Alkalinity	84.9		1	1.00	2.00	mg/L		06/19/25 12:38	SM 2320 B-21
Nitrate	0.095	U	1	0.095	0.50	mg/L		06/13/25 19:20	300.0
Sulfate	19.4		1	0.46	3.00	mg/L		06/13/25 19:20	300.0
COD	53.0		1	1.50	10.0	mg/L		06/19/25 13:35	SM 5220 D-11

Comments: The alkalinity to pH 4.33=84.9 mg CaCO3/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2334

Project: Buff

RunNo.: LB136158

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	05/22/2025
Chloride	mg/L	3.1	3	103	90-110	05/22/2025
Fluoride	mg/L	2.1	2	105	90-110	05/22/2025
Nitrite	mg/L	3.1	3	103	90-110	05/22/2025
Nitrate	mg/L	2.6	2.5	104	90-110	05/22/2025
Sulfate	mg/L	15.1	15	101	90-110	05/22/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	05/22/2025
Sample ID: CCV1						
Bromide	mg/L	9.9	10	99	90-110	06/13/2025
Chloride	mg/L	3	3	100	90-110	06/13/2025
Fluoride	mg/L	0.81	2	40	90-110	06/13/2025
Nitrite	mg/L	3	3	100	90-110	06/13/2025
Nitrate	mg/L	2.5	2.5	100	90-110	06/13/2025
Sulfate	mg/L	14.6	15	97	90-110	06/13/2025
Orthophosphate as P	mg/L	4.5	5	90	90-110	06/13/2025
Sample ID: CCV2						
Bromide	mg/L	10	10	100	90-110	06/13/2025
Chloride	mg/L	3	3	100	90-110	06/13/2025
Fluoride	mg/L	2	2	100	90-110	06/13/2025
Nitrite	mg/L	3	3	100	90-110	06/13/2025
Nitrate	mg/L	2.5	2.5	100	90-110	06/13/2025
Sulfate	mg/L	14.8	15	99	90-110	06/13/2025
Orthophosphate as P	mg/L	5	5	100	90-110	06/13/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2334

Project: Buff

RunNo.: LB136206

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	47.915	50	96	95-105	06/19/2025
Sample ID: CCV2 COD	mg/L	48.931	50	98	95-105	06/19/2025

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

RunNo.: LB136158

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	05/22/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	05/22/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	05/22/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	05/22/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	05/22/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	05/22/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	05/22/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/13/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/13/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/13/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/13/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/13/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/13/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/13/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/13/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/13/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/13/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/13/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/13/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/13/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/13/2025

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

RunNo.: LB136206

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	1.50	10	05/28/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/19/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/19/2025

Preparation Blank Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136158BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/13/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/13/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/13/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/13/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/13/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/13/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/13/2025
Sample ID: LB136205BLW							
Alkalinity	mg/L	< 1.0000	1.0000	U	1	2	06/19/2025
Sample ID: LB136206BL							
COD	mg/L	< 5.0000	5.0000	U	1.5	10.0	06/19/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q2334
Project:	Buff	Sample ID:	Q2334-01
Client ID:	MW3MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.1		0.37	U	10	1	101		06/13/2025
COD	mg/L	75-125	61.1		14.4		50.0	1	93		06/19/2025
Chloride	mg/L	80-120	4.20		1.20		3	1	100		06/13/2025
Fluoride	mg/L	80-120	2.10		0.11	U	2	1	105		06/13/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		06/13/2025
Nitrate	mg/L	80-120	2.60		0.18	J	2.5	1	97		06/13/2025
Sulfate	mg/L	80-120	19.2		4.70		15	1	97		06/13/2025
Orthophosphate as P	mg/L	80-120	5.30		0.34	U	5	1	106		06/13/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q2334
Project:	Buff	Sample ID:	Q2334-01
Client ID:	MW3MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.1		0.37	U	10	1	101		06/13/2025
COD	mg/L	75-125	60.1		14.4		50.0	1	91		06/19/2025
Chloride	mg/L	80-120	4.20		1.20		3	1	100		06/13/2025
Fluoride	mg/L	80-120	2.10		0.11	U	2	1	105		06/13/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		06/13/2025
Nitrate	mg/L	80-120	2.60		0.18	J	2.5	1	97		06/13/2025
Sulfate	mg/L	80-120	19.2		4.70		15	1	97		06/13/2025
Orthophosphate as P	mg/L	80-120	5.00		0.34	U	5	1	100		06/13/2025

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q2334
Project:	Buff	Sample ID:	Q2334-01
Client ID:	MW3DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Alkalinity	mg/L	+/-20	122		121		1	1		06/19/2025
COD	mg/L	+/-20	14.4		15.4		1	6.71		06/19/2025

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q2334
Project:	Buff	Sample ID:	Q2334-01
Client ID:	MW3MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Bromide	mg/L	+/-20	10.1		10.1		1	0		06/13/2025
Chloride	mg/L	+/-20	4.20		4.20		1	0		06/13/2025
Fluoride	mg/L	+/-20	2.10		2.10		1	0		06/13/2025
Nitrate	mg/L	+/-20	2.60		2.60		1	0		06/13/2025
Nitrite	mg/L	+/-20	3.00		3.00		1	0		06/13/2025
Sulfate	mg/L	+/-20	19.2		19.2		1	0		06/13/2025
Orthophosphate as P	mg/L	+/-20	5.30		5.00		1	6		06/13/2025
COD	mg/L	+/-20	61.1		60.1		1	1.65		06/19/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Run No.: LB136158

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136158BSW							
Bromide	mg/L	10	10.0		100	1	90-110	06/13/2025
Chloride	mg/L	3	3.00		100	1	90-110	06/13/2025
Fluoride	mg/L	2	2.00		100	1	90-110	06/13/2025
Nitrite	mg/L	3	3.00		100	1	90-110	06/13/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	06/13/2025
Sulfate	mg/L	15	15.0		100	1	90-110	06/13/2025
Orthophosphate as P	mg/L	5	4.90		98	1	90-110	06/13/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Run No.: LB136205

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136205BSW							
Alkalinity	mg/L	50	47.5		95	1	80-120	06/19/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Run No.: LB136206

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136206BS							
COD	mg/L	50	51.0		102	1	90-110	06/19/2025

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136158

Review By	Iwona	Review On	6/16/2025 1:26:39 PM
Supervise By	Sohil	Supervise On	6/16/2025 1:28:02 PM
SubDirectory	LB136158	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113186,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192		
ICV Standard	WP113193		
CCV Standard	WP113532		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113533		
Chk Standard	WP113194,WP113195		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	05/22/25 11:09	All standards, samples, and	NF/IZ	OK
2	STD2	STD2	CAL2	05/22/25 11:30	QC are filtered through	NF/IZ	OK
3	STD3	STD3	CAL3	05/22/25 11:52	0.45um, filter lot W3160	NF/IZ	OK
4	STD4	STD4	CAL4	05/22/25 12:13		NF/IZ	OK
5	STD5	STD5	CAL5	05/22/25 12:35		NF/IZ	OK
6	STD6	STD6	CAL6	05/22/25 12:56		NF/IZ	OK
7	STD7	STD7	CAL7	05/22/25 13:17		NF/IZ	OK
8	ICV1	ICV1	ICV	05/22/25 13:39		NF/IZ	OK
9	ICB1	ICB1	ICB	05/22/25 14:22		NF/IZ	OK
10	CCV1	CCV1	CCV	06/13/25 16:28		NF/IZ	OK
11	CCB1	CCB1	CCB	06/13/25 16:49		NF/IZ	OK
12	LB136158BLW	LB136158BLW	MB	06/13/25 17:11		NF/IZ	OK
13	LB136158BSW	LB136158BSW	LCS	06/13/25 17:32		NF/IZ	OK
14	Q2334-01	MW3	SAM	06/13/25 17:54		NF/IZ	OK
15	Q2334-01MS	MW3MS	MS	06/13/25 18:15	9.5ml of sample, 0.5mL W3092	NF/IZ	OK
16	Q2334-01MSD	MW3MSD	MSD	06/13/25 18:37	9.5ml of sample, 0.5mL W3092	NF/IZ	OK
17	Q2334-02	MW4	SAM	06/13/25 18:58		NF/IZ	OK
18	Q2334-03	GBTW1	SAM	06/13/25 19:20		NF/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136158

Review By	Iwona	Review On	6/16/2025 1:26:39 PM
Supervise By	Sohil	Supervise On	6/16/2025 1:28:02 PM
SubDirectory	LB136158	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113186,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192		
ICV Standard	WP113193		
CCV Standard	WP113532		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113533		
Chk Standard	WP113194,WP113195		

19	CCV2	CCV2	CCV	06/13/25 19:42		NF/IZ	OK
20	CCB2	CCB2	CCB	06/13/25 20:03		NF/IZ	OK

Instrument ID: TITRATOR

Daily Analysis Runlog For Sequence/QC Batch ID # LB136205

Review By	Iwona	Review On	6/20/2025 4:04:31 PM
Supervise By	Sohil	Supervise On	6/20/2025 4:09:21 PM
SubDirectory	LB136205	Test	Alkalinity
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113583		
Chk Standard	W3071,W3178,W3150		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136205BLW	LB136205BLW	MB	06/19/25 12:15	pH=3.98	Iwona	OK
2	LB136205BSW	LB136205BSW	LCS	06/19/25 12:18	pH=4.38	Iwona	OK
3	Q2334-01	MW3	SAM	06/19/25 12:24	pH=4.31	Iwona	OK
4	Q2334-01DUP	MW3DUP	DUP	06/19/25 12:29	pH=4.35	Iwona	OK
5	Q2334-02	MW4	SAM	06/19/25 12:34	pH=4.30	Iwona	OK
6	Q2334-03	GBTW1	SAM	06/19/25 12:38	pH=4.33	Iwona	OK
7	Q2344-01	MW-1	SAM	06/19/25 12:43	pH=4.31	Iwona	OK
8	Q2344-03	MW-3	SAM	06/19/25 12:49	pH=4.40	Iwona	OK
9	Q2344-05	MW-4	SAM	06/19/25 12:55	pH=4.37	Iwona	OK
10	Q2344-07	MW-2	SAM	06/19/25 13:02	pH=4.32	Iwona	OK

Instrument ID: SPECTROPHOTOMETER-2

Daily Analysis Runlog For Sequence/QC Batch ID # LB136206

Review By	Iwona	Review On	6/19/2025 2:30:45 PM
Supervise By	Sohil	Supervise On	6/19/2025 2:34:44 PM
SubDirectory	LB136206	Test	COD
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP113238,WP113237,WP113235,WP113234,WP113233,WP113240,WP113236,W3129,WP113586,WP113587,WP1		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	05/28/25 13:10		Iwona	OK
2	CAL2	CAL2	CAL	05/28/25 13:10		Iwona	OK
3	CAL3	CAL3	CAL	05/28/25 13:11		Iwona	OK
4	CAL4	CAL4	CAL	05/28/25 13:11		Iwona	OK
5	CAL5	CAL5	CAL	05/28/25 13:12		Iwona	OK
6	CAL6	CAL6	CAL	05/28/25 13:12		Iwona	OK
7	ICV	ICV	ICV	05/28/25 13:13		Iwona	OK
8	ICB	ICB	ICB	05/28/25 13:13		Iwona	OK
9	CCV1	CCV1	CCV	06/19/25 13:30		Iwona	OK
10	CCB1	CCB1	CCB	06/19/25 13:30		Iwona	OK
11	RL Check	RL Check	RL	06/19/25 13:31		Iwona	OK
12	LB136206BL	LB136206BL	MB	06/19/25 13:31		Iwona	OK
13	LB136206BS	LB136206BS	LCS	06/19/25 13:32		Iwona	OK
14	Q2334-01	MW3	SAM	06/19/25 13:32		Iwona	OK
15	Q2334-01DUP	MW3DUP	DUP	06/19/25 13:33		Iwona	OK
16	Q2334-01MS	MW3MS	MS	06/19/25 13:33		Iwona	OK
17	Q2334-01MSD	MW3MSD	MSD	06/19/25 13:34		Iwona	OK
18	Q2334-02	MW4	SAM	06/19/25 13:34		Iwona	OK

Instrument ID: SPECTROPHOTOMETER-2

Daily Analysis Runlog For Sequence/QC Batch ID # LB136206

Review By	Iwona	Review On	6/19/2025 2:30:45 PM
Supervise By	Sohil	Supervise On	6/19/2025 2:34:44 PM
SubDirectory	LB136206	Test	COD
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP113238,WP113237,WP113235,WP113234,WP113233,WP113240,WP113236,W3129,WP113586,WP113587,WP1		

19	Q2334-03	GBTW1	SAM	06/19/25 13:35		Iwona	OK
20	CCV2	CCV2	CCV	06/19/25 13:35		Iwona	OK
21	CCB2	CCB2	CCB	06/19/25 13:36		Iwona	OK

LAB CHRONICLE

OrderID:	Q2334	OrderDate:	6/13/2025 3:40:07 PM
Client:	G Environmental	Project:	Buff
Contact:	Gary Landis	Location:	D52,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2334-01	MW3	WATER			06/12/25 13:45			06/13/25
			Alkalinity	SM2320 B			06/19/25 12:24	
			Anions Group1	300.0			06/13/25 17:54	
			COD	SM5220 D			06/19/25 13:32	
Q2334-02	MW4	WATER			06/12/25 12:45			06/13/25
			Alkalinity	SM2320 B			06/19/25 12:34	
			Anions Group1	300.0			06/13/25 18:58	
			COD	SM5220 D			06/19/25 13:34	
Q2334-03	GBTW1	WATER			06/12/25 13:00			06/13/25
			Alkalinity	SM2320 B			06/19/25 12:38	
			Anions Group1	300.0			06/13/25 19:20	
			COD	SM5220 D			06/19/25 13:35	