

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Run No.: LB136158

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136158BSW							
Bromide	mg/L	10	10.0		100	1	90-110	06/13/2025
Chloride	mg/L	3	3.00		100	1	90-110	06/13/2025
Fluoride	mg/L	2	2.00		100	1	90-110	06/13/2025
Nitrite	mg/L	3	3.00		100	1	90-110	06/13/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	06/13/2025
Sulfate	mg/L	15	15.0		100	1	90-110	06/13/2025
Orthophosphate as P	mg/L	5	4.90		98	1	90-110	06/13/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Run No.: LB136205

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136205BSW							
Alkalinity	mg/L	50	47.5		95	1	80-120	06/19/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2334

Project: Buff

Run No.: LB136206

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136206BS							
COD	mg/L	50	51.0		102	1	90-110	06/19/2025