

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2334 SAS No.: Q2334 SDG No.: Q2334
 Instrument ID: MSVOA_X Calibration Date(s): 06/17/2025 06/17/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:19 17:18
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF005 = VX046718.D		RRF020 = VX046719.D		RRF050 = VX046720.D			
		RRF100 = VX046721.D		RRF150 = VX046722.D		RRF001 = VX046725.D			
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD	
Dichlorodifluoromethane	0.464	0.507	0.583	0.627	0.598	0.425	0.534	15.1	
Chloromethane	0.541	0.570	0.593	0.641	0.627	0.485	0.576	10	
Vinyl Chloride	0.569	0.632	0.634	0.687	0.644	0.521	0.614	9.7	
Bromomethane	0.371	0.367	0.369	0.341	0.280		0.346	11.1	
Chloroethane	0.350	0.388	0.377	0.405	0.381	0.332	0.372	7.1	
Trichlorofluoromethane	0.897	0.974	0.967	1.030	0.972	0.757	0.933	10.3	
1,1,2-Trichlorotrifluoroethane	0.563	0.608	0.576	0.623	0.594	0.470	0.572	9.6	
1,1-Dichloroethene	0.527	0.574	0.569	0.609	0.583	0.451	0.552	10.2	
Acetone	0.219	0.222	0.220	0.238	0.234	0.251	0.231	5.6	
Carbon Disulfide	1.503	1.644	1.599	1.735	1.656	1.869	1.668	7.5	
Methyl tert-butyl Ether	1.628	1.803	1.752	1.898	1.808	1.427	1.720	9.8	
Methyl Acetate	0.577	0.590	0.581	0.650	0.647	0.470	0.586	11.2	
Methylene Chloride	0.655	0.655	0.610	0.666	0.624	0.637	0.641	3.3	
trans-1,2-Dichloroethene	0.569	0.627	0.589	0.637	0.595	0.529	0.591	6.7	
1,1-Dichloroethane	1.054	1.153	1.088	1.170	1.114	0.972	1.092	6.6	
Cyclohexane	0.983	1.086	1.013	1.074	1.021		1.036	4.2	
2-Butanone	0.319	0.325	0.338	0.371	0.353	0.251	0.326	12.7	
Carbon Tetrachloride	0.509	0.537	0.515	0.548	0.531	0.470	0.518	5.4	
cis-1,2-Dichloroethene	0.681	0.731	0.686	0.741	0.704	0.623	0.694	6.1	
Bromochloromethane	0.527	0.459	0.485	0.519	0.500	0.480	0.495	5.2	
Chloroform	1.102	1.190	1.114	1.181	1.109	0.857	1.092	11.1	
1,1,1-Trichloroethane	0.931	1.012	0.956	1.046	0.990	0.812	0.958	8.6	
Methylcyclohexane	0.631	0.642	0.629	0.672	0.647	0.550	0.628	6.6	
Benzene	1.431	1.477	1.417	1.509	1.427	1.218	1.413	7.2	
1,2-Dichloroethane	0.508	0.523	0.495	0.528	0.497	0.429	0.497	7.2	
Trichloroethene	0.355	0.374	0.356	0.389	0.366	0.320	0.360	6.5	
1,2-Dichloropropane	0.347	0.372	0.346	0.374	0.357	0.305	0.350	7.2	
Bromodichloromethane	0.510	0.545	0.522	0.562	0.540	0.404	0.514	11.1	
4-Methyl-2-Pentanone	0.408	0.414	0.426	0.460	0.439	0.322	0.411	11.6	
Toluene	0.884	0.927	0.888	0.927	0.881	0.750	0.876	7.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2334 SAS No.: Q2334 SDG No.: Q2334
Instrument ID: MSVOA_X Calibration Date(s): 06/17/2025 06/17/2025
Heated Purge: (Y/N) N Calibration Time(s): 11:19 17:18
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF005 = VX046718.D		RRF020 = VX046719.D		RRF050 = VX046720.D			
		RRF100 = VX046721.D		RRF150 = VX046722.D		RRF001 = VX046725.D			
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD	
t-1,3-Dichloropropene	0.438	0.484	0.488	0.555	0.540	0.355	0.477	15.3	
cis-1,3-Dichloropropene	0.516	0.555	0.548	0.603	0.589	0.436	0.541	11.1	
1,1,2-Trichloroethane	0.330	0.341	0.326	0.347	0.329	0.287	0.327	6.4	
2-Hexanone	0.285	0.285	0.297	0.320	0.305	0.214	0.284	13	
Dibromochloromethane	0.380	0.402	0.388	0.419	0.402	0.318	0.385	9.3	
1,2-Dibromoethane	0.333	0.348	0.335	0.363	0.346	0.267	0.332	10.1	
Tetrachloroethene	0.345	0.353	0.340	0.360	0.339	0.341	0.346	2.4	
Chlorobenzene	1.114	1.148	1.091	1.165	1.100	1.005	1.104	5.1	
Ethyl Benzene	1.905	2.030	1.933	2.062	1.945	1.696	1.929	6.7	
m/p-Xylenes	0.705	0.764	0.724	0.765	0.718	0.635	0.719	6.7	
o-Xylene	0.686	0.729	0.692	0.739	0.698	0.498	0.673	13.1	
Styrene	1.144	1.256	1.208	1.267	1.203	0.993	1.179	8.6	
Bromoform	0.268	0.295	0.287	0.315	0.304	0.226	0.282	11.3	
Isopropylbenzene	3.593	3.914	3.723	4.004	3.814	3.048	3.682	9.3	
1,1,2,2-Tetrachloroethane	1.037	1.075	1.044	1.124	1.074	0.816	1.028	10.6	
1,3-Dichlorobenzene	1.703	1.787	1.678	1.786	1.719	1.694	1.728	2.7	
1,4-Dichlorobenzene	1.743	1.830	1.653	1.789	1.702	1.932	1.775	5.6	
1,2-Dichlorobenzene	1.604	1.701	1.592	1.703	1.628	1.460	1.615	5.5	
1,2-Dibromo-3-Chloropropane	0.196	0.205	0.211	0.235	0.237	0.134	0.203	18.6	
1,2,4-Trichlorobenzene	1.040	1.138	1.127	1.253	1.160	1.170	1.148	6	
1,2,3-Trichlorobenzene	1.062	1.129	1.082	1.207	1.153	0.957	1.098	7.9	
1,2-Dichloroethane-d4	0.801	0.567	0.649	0.726	0.714		0.692	12.7	
Dibromofluoromethane	0.362	0.267	0.320	0.354	0.351		0.331	11.9	
Toluene-d8	1.342	0.973	1.144	1.250	1.234		1.189	11.7	
4-Bromofluorobenzene	0.513	0.354	0.426	0.466	0.456		0.443	13.3	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.