

Surrogate Summary

SDG No.: **Q2334**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2334-01	MW3	1,2-Dichloroethane-d4	50	48.3	97	70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98	70 (75)	130 (124)
		Toluene-d8	50	50.6	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102	70 (77)	130 (121)
Q2334-02	MW4	1,2-Dichloroethane-d4	50	47.8	96	70 (74)	130 (125)
		Dibromofluoromethane	50	48.4	97	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.4	99	70 (77)	130 (121)
Q2334-03	GBTW1	1,2-Dichloroethane-d4	50	48.2	96	70 (74)	130 (125)
		Dibromofluoromethane	50	49.1	98	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
VX0619WBL01	VX0619WBL01	1,2-Dichloroethane-d4	50	48.3	97	70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.4	99	70 (77)	130 (121)
VX0619WBS01	VX0619WBS01	1,2-Dichloroethane-d4	50	49.8	100	70 (74)	130 (125)
		Dibromofluoromethane	50	52.0	104	70 (75)	130 (124)
		Toluene-d8	50	52.3	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.1	108	70 (77)	130 (121)
VX0619WBSD01	VX0619WBSD01	1,2-Dichloroethane-d4	50	51.5	103	70 (74)	130 (125)
		Dibromofluoromethane	50	52.8	106	70 (75)	130 (124)
		Toluene-d8	50	52.7	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.9	110	70 (77)	130 (121)

