

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2334

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX046763.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBSD01	Dichlorodifluoromethane	20	19.8	ug/L	99	3		40 (69)	160 (116)	20 (19)
	Chloromethane	20	19.5	ug/L	98	4		40 (65)	160 (116)	20 (21)
	Vinyl chloride	20	19.7	ug/L	99	4		70 (65)	130 (117)	20 (19)
	Bromomethane	20	22.2	ug/L	111	5		40 (58)	160 (125)	20 (20)
	Chloroethane	20	20.6	ug/L	103	3		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	19.3	ug/L	97	5		40 (73)	160 (115)	20 (16)
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99	6		70 (80)	130 (112)	20 (15)
	1,1-Dichloroethene	20	19.5	ug/L	98	5		70 (74)	130 (110)	20 (20)
	Acetone	100	81.5	ug/L	82	10		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	18.0	ug/L	90	6		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	18.7	ug/L	94	10		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	17.8	ug/L	89	11		70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	19.3	ug/L	97	6		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	19.2	ug/L	96	4		70 (75)	130 (108)	20 (16)
	1,1-Dichloroethane	20	19.4	ug/L	97	4		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	19.7	ug/L	99	5		70 (75)	130 (110)	20 (20)
	2-Butanone	100	87.0	ug/L	87	10		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	19.1	ug/L	96	6		70 (77)	130 (113)	20 (15)
	cis-1,2-Dichloroethene	20	19.4	ug/L	97	4		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	23.1	ug/L	116	11		70 (70)	130 (124)	20 (20)
	Chloroform	20	20.5	ug/L	103	8		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	19.3	ug/L	97	9		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	19.0	ug/L	95	2		70 (72)	130 (115)	20 (20)
	Benzene	20	20.0	ug/L	100	5		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	20.1	ug/L	101	8		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.3	ug/L	97	5		70 (77)	130 (113)	20 (15)
	1,2-Dichloropropane	20	20.3	ug/L	102	10		70 (83)	130 (111)	20 (16)
	Bromodichloromethane	20	20.1	ug/L	101	6		70 (83)	130 (110)	20 (16)
	4-Methyl-2-Pentanone	100	91.8	ug/L	92	11		40 (74)	160 (118)	20 (25)
	Toluene	20	20.3	ug/L	102	7		70 (82)	130 (110)	20 (16)
	t-1,3-Dichloropropene	20	19.5	ug/L	98	9		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	19.9	ug/L	100	9		70 (82)	130 (110)	20 (16)
	1,1,2-Trichloroethane	20	20.4	ug/L	102	9		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	88.8	ug/L	89	12		40 (73)	160 (117)	20 (25)
	Dibromochloromethane	20	19.9	ug/L	100	8		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	20.1	ug/L	101	12		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	18.8	ug/L	94	5		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	19.4	ug/L	97	7		70 (82)	130 (109)	20 (15)
	Ethyl Benzene	20	19.1	ug/L	96	6		70 (83)	130 (109)	20 (16)
	m/p-Xylenes	40	38.8	ug/L	97	4		70 (82)	130 (110)	20 (15)
	o-Xylene	20	19.8	ug/L	99	6		70 (83)	130 (109)	20 (20)
	Styrene	20	19.7	ug/L	99	7		70 (80)	130 (111)	20 (17)
	Bromoform	20	18.4	ug/L	92	9		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	18.6	ug/L	93	2		70 (83)	130 (112)	20 (29)
	1,1,2,2-Tetrachloroethane	20	18.4	ug/L	92	7		70 (76)	130 (118)	20 (20)
	1,3-Dichlorobenzene	20	18.6	ug/L	93	4		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	17.8	ug/L	89	1		70 (82)	130 (107)	20 (15)
	1,2-Dichlorobenzene	20	19.0	ug/L	95	4		70 (82)	130 (109)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2334

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX046763.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBSD01	1,2-Dibromo-3-Chloropropane	20	16.1	ug/L	81	6		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	18.3	ug/L	92	7		70 (75)	130 (113)	20 (29)
	1,2,3-Trichlorobenzene	20	18.2	ug/L	91	3		70 (76)	130 (114)	20 (29)