

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2334 SAS No.: Q2334 SDG NO.: Q2334
 Lab File ID: VX046758.D Date Analyzed: 06/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	96867	12.018				
UPPER LIMIT	193734	12.518				
LOWER LIMIT	48433.5	11.518				
EPA SAMPLE NO.						
MW3	92623	12.02				
MW4	136063	12.02				
GBTW1	125261	12.02				
VX0619WBL01	93040	12.02				
VX0619WBS01	104652	12.02				
VX0619WBSD01	94851	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.