

Continuing Calibration Report for SequenceID : FE061825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054494.D

Aliphatic C9-C12	7330477.000	60.000	3.072	6.717	122174.617	125255.224	2.459
Aliphatic C12-C16	5115155.000	40.000	6.718	10.164	127878.875	131151.686	2.495
Aliphatic C16-C21	7880814.000	60.000	10.165	13.536	131346.900	133925.941	1.926
Aliphatic C21-C28	10214402.000	80.000	13.537	17.204	127680.025	127068.779	-0.481
Aliphatic C28-C40	16126543.000	120.000	17.205	22.063	134387.858	128245.359	-4.790
Aliphatic EPH	46667391.000	360.000	3.072	22.063	129631.642	128755.230	-0.681