

LAB CHRONICLE

OrderID:	Q2344	OrderDate:	6/17/2025 11:52:00 AM
Client:	Lockwood, Kessler & Bartlett, Inc.	Project:	Ansonia Landfill 2025
Contact:	John Gerlach	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2344-01	MW-1	Water	Metals Group4	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-02	MW-1	Water	Dissolved Metals Group5	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-03	MW-3	Water	Metals Group4	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-04	MW-3	Water	Dissolved Metals Group5	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-05	MW-4	Water	Metals Group4	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-06	MW-4	Water	Dissolved Metals Group5	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-07	MW-2	Water	Metals Group4	6010D	06/16/25	06/19/25	06/23/25	06/17/25
Q2344-08	MW-2	Water	Dissolved Metals Group5	6010D	06/16/25	06/19/25	06/23/25	06/17/25



A
B
C
D
E
F
G
H

Order ID: Q2344

Project ID: Ansonia Landfill 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : MW-1								
Q2344-01	MW-1	Water	Barium	47.3	J	7.28	50.0	ug/L
Q2344-01	MW-1	Water	Cadmium	0.80	J	0.25	3.00	ug/L
Q2344-01	MW-1	Water	Chromium	3.48	J	1.06	5.00	ug/L
Q2344-01	MW-1	Water	Cobalt	8.07	J	1.13	15.0	ug/L
Q2344-01	MW-1	Water	Copper	3.91	J	2.30	10.0	ug/L
Q2344-01	MW-1	Water	Lead	6.99		1.15	6.00	ug/L
Q2344-01	MW-1	Water	Nickel	5.72	J	1.53	20.0	ug/L
Q2344-01	MW-1	Water	Potassium	4310		459	1000	ug/L
Q2344-01	MW-1	Water	Sodium	57400		434	1000	ug/L
Q2344-01	MW-1	Water	Zinc	111		1.75	20.0	ug/L
Client ID : MW-1								
Q2344-02	MW-1	Water	Iron	359		11.7	50.0	ug/L
Q2344-02	MW-1	Water	Manganese	74.2		2.97	10.0	ug/L
Client ID : MW-3								
Q2344-03	MW-3	Water	Barium	130		7.28	50.0	ug/L
Q2344-03	MW-3	Water	Chromium	1.18	J	1.06	5.00	ug/L
Q2344-03	MW-3	Water	Cobalt	11.2	J	1.13	15.0	ug/L
Q2344-03	MW-3	Water	Copper	2.69	J	2.30	10.0	ug/L
Q2344-03	MW-3	Water	Lead	9.16		1.15	6.00	ug/L
Q2344-03	MW-3	Water	Potassium	13000		459	1000	ug/L
Q2344-03	MW-3	Water	Sodium	129000		434	1000	ug/L
Q2344-03	MW-3	Water	Zinc	11.0	J	1.75	20.0	ug/L
Client ID : MW-3								
Q2344-04	MW-3	Water	Manganese	13.0		2.97	10.0	ug/L
Client ID : MW-4								
Q2344-05	MW-4	Water	Antimony	4.32	J	3.38	25.0	ug/L
Q2344-05	MW-4	Water	Barium	132		7.28	50.0	ug/L
Q2344-05	MW-4	Water	Chromium	1.79	J	1.06	5.00	ug/L
Q2344-05	MW-4	Water	Cobalt	11.1	J	1.13	15.0	ug/L
Q2344-05	MW-4	Water	Lead	7.81		1.15	6.00	ug/L
Q2344-05	MW-4	Water	Potassium	13400		459	1000	ug/L
Q2344-05	MW-4	Water	Sodium	132000		434	1000	ug/L
Q2344-05	MW-4	Water	Zinc	9.95	J	1.75	20.0	ug/L
Client ID : MW-4								

Hit Summary Sheet
SW-846

SDG No.: Q2344

Order ID: Q2344

Client: Lockwood, Kessler & Bartlett, Inc.

Project ID: Ansonia Landfill 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2344-06	MW-4	Water	Manganese	13.3		2.97	10.0	ug/L
Client ID : MW-2								
Q2344-07	MW-2	Water	Antimony	5.08	J	3.38	25.0	ug/L
Q2344-07	MW-2	Water	Barium	166		7.28	50.0	ug/L
Q2344-07	MW-2	Water	Chromium	4.01	J	1.06	5.00	ug/L
Q2344-07	MW-2	Water	Cobalt	12.6	J	1.13	15.0	ug/L
Q2344-07	MW-2	Water	Copper	28.2		2.30	10.0	ug/L
Q2344-07	MW-2	Water	Lead	9.67		1.15	6.00	ug/L
Q2344-07	MW-2	Water	Nickel	26.2		1.53	20.0	ug/L
Q2344-07	MW-2	Water	Potassium	23200		459	1000	ug/L
Q2344-07	MW-2	Water	Sodium	201000		434	1000	ug/L
Q2344-07	MW-2	Water	Thallium	3.38	J	2.19	20.0	ug/L
Q2344-07	MW-2	Water	Zinc	95.0		1.75	20.0	ug/L
Client ID : MW-2								
Q2344-08	MW-2	Water	Iron	6110		11.7	50.0	ug/L
Q2344-08	MW-2	Water	Manganese	884		2.97	10.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-1	SDG No.:	Q2344
Lab Sample ID:	Q2344-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	3.38	U	1	3.38	25.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-38-2	Arsenic	2.56	U	1	2.56	10.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-39-3	Barium	47.3	J	1	7.28	50.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-41-7	Beryllium	0.28	U	1	0.28	3.00	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-43-9	Cadmium	0.80	J	1	0.25	3.00	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-47-3	Chromium	3.48	J	1	1.06	5.00	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-48-4	Cobalt	8.07	J	1	1.13	15.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-50-8	Copper	3.91	J	1	2.30	10.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7439-92-1	Lead	6.99		1	1.15	6.00	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-02-0	Nickel	5.72	J	1	1.53	20.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-09-7	Potassium	4310		1	459	1000	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7782-49-2	Selenium	4.82	U	1	4.82	10.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-22-4	Silver	0.81	U	1	0.81	5.00	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-23-5	Sodium	57400		1	434	1000	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-28-0	Thallium	2.19	U	1	2.19	20.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-62-2	Vanadium	3.13	U	1	3.13	20.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010
7440-66-6	Zinc	111		1	1.75	20.0	ug/L	06/19/25 10:10	06/23/25 18:16	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-1	SDG No.:	Q2344
Lab Sample ID:	Q2344-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	359		1	11.7	50.0	ug/L	06/19/25 10:10	06/23/25 18:20	6010D	SW3010
7439-96-5	Manganese	74.2		1	2.97	10.0	ug/L	06/19/25 10:10	06/23/25 18:20	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-3	SDG No.:	Q2344
Lab Sample ID:	Q2344-03	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	3.38	U	1	3.38	25.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-38-2	Arsenic	2.56	U	1	2.56	10.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-39-3	Barium	130		1	7.28	50.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-41-7	Beryllium	0.28	U	1	0.28	3.00	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-43-9	Cadmium	0.25	U	1	0.25	3.00	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-47-3	Chromium	1.18	J	1	1.06	5.00	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-48-4	Cobalt	11.2	J	1	1.13	15.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-50-8	Copper	2.69	J	1	2.30	10.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7439-92-1	Lead	9.16		1	1.15	6.00	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-02-0	Nickel	1.53	U	1	1.53	20.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-09-7	Potassium	13000		1	459	1000	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7782-49-2	Selenium	4.82	U	1	4.82	10.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-22-4	Silver	0.81	U	1	0.81	5.00	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-23-5	Sodium	129000		1	434	1000	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-28-0	Thallium	2.19	U	1	2.19	20.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-62-2	Vanadium	3.13	U	1	3.13	20.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010
7440-66-6	Zinc	11.0	J	1	1.75	20.0	ug/L	06/19/25 10:10	06/23/25 18:25	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-3	SDG No.:	Q2344
Lab Sample ID:	Q2344-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	11.7	U	1	11.7	50.0	ug/L	06/19/25 10:10	06/23/25 18:30	6010D	SW3010
7439-96-5	Manganese	13.0		1	2.97	10.0	ug/L	06/19/25 10:10	06/23/25 18:30	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-4	SDG No.:	Q2344
Lab Sample ID:	Q2344-05	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	4.32	J	1	3.38	25.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-38-2	Arsenic	2.56	U	1	2.56	10.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-39-3	Barium	132		1	7.28	50.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-41-7	Beryllium	0.28	U	1	0.28	3.00	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-43-9	Cadmium	0.25	U	1	0.25	3.00	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-47-3	Chromium	1.79	J	1	1.06	5.00	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-48-4	Cobalt	11.1	J	1	1.13	15.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-50-8	Copper	2.30	U	1	2.30	10.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7439-92-1	Lead	7.81		1	1.15	6.00	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-02-0	Nickel	1.53	U	1	1.53	20.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-09-7	Potassium	13400		1	459	1000	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7782-49-2	Selenium	4.82	U	1	4.82	10.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-22-4	Silver	0.81	U	1	0.81	5.00	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-23-5	Sodium	132000		1	434	1000	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-28-0	Thallium	2.19	U	1	2.19	20.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-62-2	Vanadium	3.13	U	1	3.13	20.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010
7440-66-6	Zinc	9.95	J	1	1.75	20.0	ug/L	06/19/25 10:10	06/23/25 18:34	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-4	SDG No.:	Q2344
Lab Sample ID:	Q2344-06	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	11.7	U	1	11.7	50.0	ug/L	06/19/25 10:10	06/23/25 18:48	6010D	SW3010
7439-96-5	Manganese	13.3		1	2.97	10.0	ug/L	06/19/25 10:10	06/23/25 18:48	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-2	SDG No.:	Q2344
Lab Sample ID:	Q2344-07	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	5.08	J	1	3.38	25.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-38-2	Arsenic	2.56	U	1	2.56	10.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-39-3	Barium	166		1	7.28	50.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-41-7	Beryllium	0.28	U	1	0.28	3.00	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-43-9	Cadmium	0.25	U	1	0.25	3.00	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-47-3	Chromium	4.01	J	1	1.06	5.00	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-48-4	Cobalt	12.6	J	1	1.13	15.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-50-8	Copper	28.2		1	2.30	10.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7439-92-1	Lead	9.67		1	1.15	6.00	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-02-0	Nickel	26.2		1	1.53	20.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-09-7	Potassium	23200		1	459	1000	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7782-49-2	Selenium	4.82	U	1	4.82	10.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-22-4	Silver	0.81	U	1	0.81	5.00	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-23-5	Sodium	201000		1	434	1000	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-28-0	Thallium	3.38	J	1	2.19	20.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-62-2	Vanadium	3.13	U	1	3.13	20.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010
7440-66-6	Zinc	95.0		1	1.75	20.0	ug/L	06/19/25 10:10	06/23/25 18:52	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	06/16/25
Project:	Ansonia Landfill 2025	Date Received:	06/17/25
Client Sample ID:	MW-2	SDG No.:	Q2344
Lab Sample ID:	Q2344-08	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	6110	1	11.7		50.0	ug/L	06/19/25 10:10	06/23/25 19:19	6010D	SW3010
7439-96-5	Manganese	884	1	2.97		10.0	ug/L	06/19/25 10:10	06/23/25 19:19	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Antimony	4110	4000	103	90 - 110	P	06/23/2025	12:34	LB136236
	Arsenic	3910	4000	98	90 - 110	P	06/23/2025	12:34	LB136236
	Barium	7850	8000	98	90 - 110	P	06/23/2025	12:34	LB136236
	Beryllium	206	200	103	90 - 110	P	06/23/2025	12:34	LB136236
	Cadmium	2000	2000	100	90 - 110	P	06/23/2025	12:34	LB136236
	Chromium	806	800	101	90 - 110	P	06/23/2025	12:34	LB136236
	Cobalt	2040	2000	102	90 - 110	P	06/23/2025	12:34	LB136236
	Copper	1030	1000	103	90 - 110	P	06/23/2025	12:34	LB136236
	Iron	3920	4000	98	90 - 110	P	06/23/2025	12:34	LB136236
	Lead	3950	4000	99	90 - 110	P	06/23/2025	12:34	LB136236
	Manganese	1960	2000	98	90 - 110	P	06/23/2025	12:34	LB136236
	Nickel	2030	2000	102	90 - 110	P	06/23/2025	12:34	LB136236
	Potassium	19000	20000	95	90 - 110	P	06/23/2025	12:34	LB136236
	Selenium	3980	4000	100	90 - 110	P	06/23/2025	12:34	LB136236
	Silver	1070	1000	107	90 - 110	P	06/23/2025	12:34	LB136236
	Sodium	20000	20000	100	90 - 110	P	06/23/2025	12:34	LB136236
	Thallium	4050	4000	101	90 - 110	P	06/23/2025	12:34	LB136236
	Vanadium	1990	2000	100	90 - 110	P	06/23/2025	12:34	LB136236
	Zinc	2000	2000	100	90 - 110	P	06/23/2025	12:34	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Antimony	54.0	50.0	108	80 - 120	P	06/23/2025	13:09	LB136236
	Arsenic	18.0	20.0	90	80 - 120	P	06/23/2025	13:09	LB136236
	Barium	104	100	104	80 - 120	P	06/23/2025	13:09	LB136236
	Beryllium	6.48	6.0	108	80 - 120	P	06/23/2025	13:09	LB136236
	Cadmium	5.92	6.0	99	80 - 120	P	06/23/2025	13:09	LB136236
	Chromium	10.8	10.0	108	80 - 120	P	06/23/2025	13:09	LB136236
	Cobalt	31.6	30.0	105	80 - 120	P	06/23/2025	13:09	LB136236
	Copper	23.3	20.0	117	80 - 120	P	06/23/2025	13:09	LB136236
	Iron	102	100	102	80 - 120	P	06/23/2025	13:09	LB136236
	Lead	14.0	12.0	117	80 - 120	P	06/23/2025	13:09	LB136236
	Manganese	20.2	20.0	101	80 - 120	P	06/23/2025	13:09	LB136236
	Nickel	41.6	40.0	104	80 - 120	P	06/23/2025	13:09	LB136236
	Potassium	1690	2000	84	80 - 120	P	06/23/2025	13:09	LB136236
	Selenium	20.4	20.0	102	80 - 120	P	06/23/2025	13:09	LB136236
	Silver	10.8	10.0	108	80 - 120	P	06/23/2025	13:09	LB136236
	Sodium	1890	2000	94	80 - 120	P	06/23/2025	13:09	LB136236
	Thallium	41.5	40.0	104	80 - 120	P	06/23/2025	13:09	LB136236
	Vanadium	35.7	40.0	89	80 - 120	P	06/23/2025	13:09	LB136236
	Zinc	41.1	40.0	103	80 - 120	P	06/23/2025	13:09	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Antimony	4920	5000	98	90 - 110	P	06/23/2025	13:56	LB136236
	Arsenic	5140	5000	103	90 - 110	P	06/23/2025	13:56	LB136236
	Barium	9940	10000	99	90 - 110	P	06/23/2025	13:56	LB136236
	Beryllium	250	250	100	90 - 110	P	06/23/2025	13:56	LB136236
	Cadmium	2580	2500	103	90 - 110	P	06/23/2025	13:56	LB136236
	Chromium	988	1000	99	90 - 110	P	06/23/2025	13:56	LB136236
	Cobalt	2580	2500	103	90 - 110	P	06/23/2025	13:56	LB136236
	Copper	1240	1250	100	90 - 110	P	06/23/2025	13:56	LB136236
	Iron	4900	5000	98	90 - 110	P	06/23/2025	13:56	LB136236
	Lead	5160	5000	103	90 - 110	P	06/23/2025	13:56	LB136236
	Manganese	2450	2500	98	90 - 110	P	06/23/2025	13:56	LB136236
	Nickel	2570	2500	103	90 - 110	P	06/23/2025	13:56	LB136236
	Potassium	23600	25000	94	90 - 110	P	06/23/2025	13:56	LB136236
	Selenium	5020	5000	100	90 - 110	P	06/23/2025	13:56	LB136236
	Silver	1230	1250	98	90 - 110	P	06/23/2025	13:56	LB136236
	Sodium	24500	25000	98	90 - 110	P	06/23/2025	13:56	LB136236
	Thallium	5120	5000	102	90 - 110	P	06/23/2025	13:56	LB136236
	Vanadium	2440	2500	98	90 - 110	P	06/23/2025	13:56	LB136236
	Zinc	2450	2500	98	90 - 110	P	06/23/2025	13:56	LB136236
CCV02	Antimony	4900	5000	98	90 - 110	P	06/23/2025	14:54	LB136236
	Arsenic	5120	5000	102	90 - 110	P	06/23/2025	14:54	LB136236
	Barium	10000	10000	100	90 - 110	P	06/23/2025	14:54	LB136236
	Beryllium	250	250	100	90 - 110	P	06/23/2025	14:54	LB136236
	Cadmium	2570	2500	103	90 - 110	P	06/23/2025	14:54	LB136236
	Chromium	989	1000	99	90 - 110	P	06/23/2025	14:54	LB136236
	Cobalt	2560	2500	102	90 - 110	P	06/23/2025	14:54	LB136236
	Copper	1240	1250	99	90 - 110	P	06/23/2025	14:54	LB136236
	Iron	4910	5000	98	90 - 110	P	06/23/2025	14:54	LB136236
	Lead	5120	5000	102	90 - 110	P	06/23/2025	14:54	LB136236
	Manganese	2470	2500	99	90 - 110	P	06/23/2025	14:54	LB136236
	Nickel	2550	2500	102	90 - 110	P	06/23/2025	14:54	LB136236
	Potassium	23300	25000	93	90 - 110	P	06/23/2025	14:54	LB136236
	Selenium	5000	5000	100	90 - 110	P	06/23/2025	14:54	LB136236
	Silver	1230	1250	98	90 - 110	P	06/23/2025	14:54	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q2344

Contract: LOCK01 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Sodium	24100	25000	96	90 - 110	P	06/23/2025	14:54	LB136236
	Thallium	5060	5000	101	90 - 110	P	06/23/2025	14:54	LB136236
	Vanadium	2450	2500	98	90 - 110	P	06/23/2025	14:54	LB136236
	Zinc	2440	2500	98	90 - 110	P	06/23/2025	14:54	LB136236
CCV03	Antimony	4960	5000	99	90 - 110	P	06/23/2025	15:47	LB136236
	Arsenic	5210	5000	104	90 - 110	P	06/23/2025	15:47	LB136236
	Barium	10100	10000	101	90 - 110	P	06/23/2025	15:47	LB136236
	Beryllium	247	250	99	90 - 110	P	06/23/2025	15:47	LB136236
	Cadmium	2590	2500	104	90 - 110	P	06/23/2025	15:47	LB136236
	Chromium	1010	1000	101	90 - 110	P	06/23/2025	15:47	LB136236
	Cobalt	2590	2500	104	90 - 110	P	06/23/2025	15:47	LB136236
	Copper	1220	1250	98	90 - 110	P	06/23/2025	15:47	LB136236
	Iron	4960	5000	99	90 - 110	P	06/23/2025	15:47	LB136236
	Lead	5190	5000	104	90 - 110	P	06/23/2025	15:47	LB136236
	Manganese	2500	2500	100	90 - 110	P	06/23/2025	15:47	LB136236
	Nickel	2570	2500	103	90 - 110	P	06/23/2025	15:47	LB136236
	Potassium	24100	25000	96	90 - 110	P	06/23/2025	15:47	LB136236
	Selenium	5060	5000	101	90 - 110	P	06/23/2025	15:47	LB136236
	Silver	1260	1250	101	90 - 110	P	06/23/2025	15:47	LB136236
	Sodium	24700	25000	99	90 - 110	P	06/23/2025	15:47	LB136236
	Thallium	5120	5000	102	90 - 110	P	06/23/2025	15:47	LB136236
	Vanadium	2490	2500	100	90 - 110	P	06/23/2025	15:47	LB136236
	Zinc	2490	2500	99	90 - 110	P	06/23/2025	15:47	LB136236
CCV04	Antimony	4960	5000	99	90 - 110	P	06/23/2025	16:51	LB136236
	Arsenic	5240	5000	105	90 - 110	P	06/23/2025	16:51	LB136236
	Barium	10200	10000	102	90 - 110	P	06/23/2025	16:51	LB136236
	Beryllium	259	250	104	90 - 110	P	06/23/2025	16:51	LB136236
	Cadmium	2600	2500	104	90 - 110	P	06/23/2025	16:51	LB136236
	Chromium	1010	1000	101	90 - 110	P	06/23/2025	16:51	LB136236
	Cobalt	2610	2500	104	90 - 110	P	06/23/2025	16:51	LB136236
	Copper	1280	1250	103	90 - 110	P	06/23/2025	16:51	LB136236
	Iron	4900	5000	98	90 - 110	P	06/23/2025	16:51	LB136236
	Lead	5230	5000	105	90 - 110	P	06/23/2025	16:51	LB136236
	Manganese	2490	2500	100	90 - 110	P	06/23/2025	16:51	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q2344

Contract: LOCK01 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Nickel	2590	2500	104	90 - 110	P	06/23/2025	16:51	LB136236
	Potassium	23800	25000	95	90 - 110	P	06/23/2025	16:51	LB136236
	Selenium	5070	5000	101	90 - 110	P	06/23/2025	16:51	LB136236
	Silver	1260	1250	101	90 - 110	P	06/23/2025	16:51	LB136236
	Sodium	24800	25000	99	90 - 110	P	06/23/2025	16:51	LB136236
	Thallium	5160	5000	103	90 - 110	P	06/23/2025	16:51	LB136236
	Vanadium	2490	2500	100	90 - 110	P	06/23/2025	16:51	LB136236
	Zinc	2480	2500	99	90 - 110	P	06/23/2025	16:51	LB136236
CCV05	Antimony	4940	5000	99	90 - 110	P	06/23/2025	17:45	LB136236
	Arsenic	5200	5000	104	90 - 110	P	06/23/2025	17:45	LB136236
	Barium	10200	10000	102	90 - 110	P	06/23/2025	17:45	LB136236
	Beryllium	260	250	104	90 - 110	P	06/23/2025	17:45	LB136236
	Cadmium	2590	2500	104	90 - 110	P	06/23/2025	17:45	LB136236
	Chromium	1000	1000	100	90 - 110	P	06/23/2025	17:45	LB136236
	Cobalt	2600	2500	104	90 - 110	P	06/23/2025	17:45	LB136236
	Copper	1290	1250	104	90 - 110	P	06/23/2025	17:45	LB136236
	Iron	4880	5000	98	90 - 110	P	06/23/2025	17:45	LB136236
	Lead	5210	5000	104	90 - 110	P	06/23/2025	17:45	LB136236
	Manganese	2470	2500	99	90 - 110	P	06/23/2025	17:45	LB136236
	Nickel	2580	2500	103	90 - 110	P	06/23/2025	17:45	LB136236
	Potassium	23800	25000	95	90 - 110	P	06/23/2025	17:45	LB136236
	Selenium	5050	5000	101	90 - 110	P	06/23/2025	17:45	LB136236
	Silver	1260	1250	101	90 - 110	P	06/23/2025	17:45	LB136236
	Sodium	25000	25000	100	90 - 110	P	06/23/2025	17:45	LB136236
	Thallium	5140	5000	103	90 - 110	P	06/23/2025	17:45	LB136236
	Vanadium	2470	2500	99	90 - 110	P	06/23/2025	17:45	LB136236
	Zinc	2450	2500	98	90 - 110	P	06/23/2025	17:45	LB136236
	Antimony	4980	5000	100	90 - 110	P	06/23/2025	18:39	LB136236
CCV06	Arsenic	5190	5000	104	90 - 110	P	06/23/2025	18:39	LB136236
	Barium	10100	10000	102	90 - 110	P	06/23/2025	18:39	LB136236
	Beryllium	255	250	102	90 - 110	P	06/23/2025	18:39	LB136236
	Cadmium	2600	2500	104	90 - 110	P	06/23/2025	18:39	LB136236
	Chromium	998	1000	100	90 - 110	P	06/23/2025	18:39	LB136236
	Cobalt	2610	2500	104	90 - 110	P	06/23/2025	18:39	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q2344
 Contract: LOCK01 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Copper	1270	1250	102	90 - 110	P	06/23/2025	18:39	LB136236
	Iron	4860	5000	97	90 - 110	P	06/23/2025	18:39	LB136236
	Lead	5210	5000	104	90 - 110	P	06/23/2025	18:39	LB136236
	Manganese	2450	2500	98	90 - 110	P	06/23/2025	18:39	LB136236
	Nickel	2600	2500	104	90 - 110	P	06/23/2025	18:39	LB136236
	Potassium	24000	25000	96	90 - 110	P	06/23/2025	18:39	LB136236
	Selenium	5060	5000	101	90 - 110	P	06/23/2025	18:39	LB136236
	Silver	1240	1250	100	90 - 110	P	06/23/2025	18:39	LB136236
	Sodium	25000	25000	100	90 - 110	P	06/23/2025	18:39	LB136236
	Thallium	5170	5000	103	90 - 110	P	06/23/2025	18:39	LB136236
	Vanadium	2450	2500	98	90 - 110	P	06/23/2025	18:39	LB136236
	Zinc	2510	2500	101	90 - 110	P	06/23/2025	18:39	LB136236
CCV07	Antimony	5010	5000	100	90 - 110	P	06/23/2025	19:24	LB136236
	Arsenic	5220	5000	104	90 - 110	P	06/23/2025	19:24	LB136236
	Barium	10200	10000	102	90 - 110	P	06/23/2025	19:24	LB136236
	Beryllium	256	250	103	90 - 110	P	06/23/2025	19:24	LB136236
	Cadmium	2610	2500	104	90 - 110	P	06/23/2025	19:24	LB136236
	Chromium	996	1000	100	90 - 110	P	06/23/2025	19:24	LB136236
	Cobalt	2620	2500	105	90 - 110	P	06/23/2025	19:24	LB136236
	Copper	1280	1250	102	90 - 110	P	06/23/2025	19:24	LB136236
	Iron	4880	5000	98	90 - 110	P	06/23/2025	19:24	LB136236
	Lead	5240	5000	105	90 - 110	P	06/23/2025	19:24	LB136236
	Manganese	2440	2500	98	90 - 110	P	06/23/2025	19:24	LB136236
	Nickel	2600	2500	104	90 - 110	P	06/23/2025	19:24	LB136236
	Potassium	24200	25000	97	90 - 110	P	06/23/2025	19:24	LB136236
	Selenium	5080	5000	102	90 - 110	P	06/23/2025	19:24	LB136236
	Silver	1250	1250	100	90 - 110	P	06/23/2025	19:24	LB136236
	Sodium	25300	25000	101	90 - 110	P	06/23/2025	19:24	LB136236
	Thallium	5210	5000	104	90 - 110	P	06/23/2025	19:24	LB136236
	Vanadium	2460	2500	98	90 - 110	P	06/23/2025	19:24	LB136236
	Zinc	2490	2500	100	90 - 110	P	06/23/2025	19:24	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Antimony	4120	4000	103	90 - 110	P	06/26/2025	17:07	LB136305
	Arsenic	3860	4000	96	90 - 110	P	06/26/2025	17:07	LB136305
	Barium	7770	8000	97	90 - 110	P	06/26/2025	17:07	LB136305
	Beryllium	202	200	101	90 - 110	P	06/26/2025	17:07	LB136305
	Cadmium	1960	2000	98	90 - 110	P	06/26/2025	17:07	LB136305
	Chromium	783	800	98	90 - 110	P	06/26/2025	17:07	LB136305
	Cobalt	1950	2000	98	90 - 110	P	06/26/2025	17:07	LB136305
	Copper	1020	1000	102	90 - 110	P	06/26/2025	17:07	LB136305
	Iron	3810	4000	95	90 - 110	P	06/26/2025	17:07	LB136305
	Lead	3870	4000	97	90 - 110	P	06/26/2025	17:07	LB136305
	Manganese	1930	2000	96	90 - 110	P	06/26/2025	17:07	LB136305
	Nickel	1940	2000	97	90 - 110	P	06/26/2025	17:07	LB136305
	Potassium	19900	20000	100	90 - 110	P	06/26/2025	17:07	LB136305
	Selenium	3980	4000	99	90 - 110	P	06/26/2025	17:07	LB136305
	Silver	997	1000	100	90 - 110	P	06/26/2025	17:07	LB136305
	Sodium	20100	20000	100	90 - 110	P	06/26/2025	17:07	LB136305
	Thallium	3870	4000	97	90 - 110	P	06/26/2025	17:07	LB136305
	Vanadium	1940	2000	97	90 - 110	P	06/26/2025	17:07	LB136305
	Zinc	1950	2000	97	90 - 110	P	06/26/2025	17:07	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Antimony	52.2	50.0	104	80 - 120	P	06/26/2025	17:19	LB136305
	Arsenic	16.6	20.0	83	80 - 120	P	06/26/2025	17:19	LB136305
	Barium	99.4	100	99	80 - 120	P	06/26/2025	17:19	LB136305
	Beryllium	6.32	6.0	105	80 - 120	P	06/26/2025	17:19	LB136305
	Cadmium	5.76	6.0	96	80 - 120	P	06/26/2025	17:19	LB136305
	Chromium	11.5	10.0	115	80 - 120	P	06/26/2025	17:19	LB136305
	Cobalt	30.1	30.0	100	80 - 120	P	06/26/2025	17:19	LB136305
	Copper	22.6	20.0	113	80 - 120	P	06/26/2025	17:19	LB136305
	Iron	112	100	112	80 - 120	P	06/26/2025	17:19	LB136305
	Lead	13.5	12.0	112	80 - 120	P	06/26/2025	17:19	LB136305
	Manganese	21.1	20.0	105	80 - 120	P	06/26/2025	17:19	LB136305
	Nickel	39.5	40.0	99	80 - 120	P	06/26/2025	17:19	LB136305
	Potassium	1630	2000	81	80 - 120	P	06/26/2025	17:19	LB136305
	Selenium	21.2	20.0	106	80 - 120	P	06/26/2025	17:19	LB136305
	Silver	10.2	10.0	102	80 - 120	P	06/26/2025	17:19	LB136305
	Sodium	1860	2000	93	80 - 120	P	06/26/2025	17:19	LB136305
	Thallium	41.7	40.0	104	80 - 120	P	06/26/2025	17:19	LB136305
	Vanadium	37.6	40.0	94	80 - 120	P	06/26/2025	17:19	LB136305
	Zinc	41.5	40.0	104	80 - 120	P	06/26/2025	17:19	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q2344

Contract: LOCK01 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Antimony	5040	5000	101	90 - 110	P	06/26/2025	17:55	LB136305
	Arsenic	5000	5000	100	90 - 110	P	06/26/2025	17:55	LB136305
	Barium	10100	10000	101	90 - 110	P	06/26/2025	17:55	LB136305
	Beryllium	251	250	100	90 - 110	P	06/26/2025	17:55	LB136305
	Cadmium	2480	2500	99	90 - 110	P	06/26/2025	17:55	LB136305
	Chromium	1010	1000	101	90 - 110	P	06/26/2025	17:55	LB136305
	Cobalt	2470	2500	99	90 - 110	P	06/26/2025	17:55	LB136305
	Copper	1260	1250	101	90 - 110	P	06/26/2025	17:55	LB136305
	Iron	4790	5000	96	90 - 110	P	06/26/2025	17:55	LB136305
	Lead	4910	5000	98	90 - 110	P	06/26/2025	17:55	LB136305
	Manganese	2490	2500	100	90 - 110	P	06/26/2025	17:55	LB136305
	Nickel	2460	2500	99	90 - 110	P	06/26/2025	17:55	LB136305
	Potassium	25800	25000	103	90 - 110	P	06/26/2025	17:55	LB136305
	Selenium	4940	5000	99	90 - 110	P	06/26/2025	17:55	LB136305
	Silver	1250	1250	100	90 - 110	P	06/26/2025	17:55	LB136305
	Sodium	25600	25000	102	90 - 110	P	06/26/2025	17:55	LB136305
	Thallium	4980	5000	100	90 - 110	P	06/26/2025	17:55	LB136305
	Vanadium	2540	2500	102	90 - 110	P	06/26/2025	17:55	LB136305
	Zinc	2550	2500	102	90 - 110	P	06/26/2025	17:55	LB136305
CCV02	Antimony	5030	5000	101	90 - 110	P	06/26/2025	18:48	LB136305
	Arsenic	5030	5000	100	90 - 110	P	06/26/2025	18:48	LB136305
	Barium	10100	10000	101	90 - 110	P	06/26/2025	18:48	LB136305
	Beryllium	247	250	99	90 - 110	P	06/26/2025	18:48	LB136305
	Cadmium	2470	2500	99	90 - 110	P	06/26/2025	18:48	LB136305
	Chromium	1040	1000	104	90 - 110	P	06/26/2025	18:48	LB136305
	Cobalt	2470	2500	99	90 - 110	P	06/26/2025	18:48	LB136305
	Copper	1240	1250	99	90 - 110	P	06/26/2025	18:48	LB136305
	Iron	4840	5000	97	90 - 110	P	06/26/2025	18:48	LB136305
	Lead	4900	5000	98	90 - 110	P	06/26/2025	18:48	LB136305
	Manganese	2500	2500	100	90 - 110	P	06/26/2025	18:48	LB136305
	Nickel	2470	2500	99	90 - 110	P	06/26/2025	18:48	LB136305
	Potassium	25500	25000	102	90 - 110	P	06/26/2025	18:48	LB136305
	Selenium	4890	5000	98	90 - 110	P	06/26/2025	18:48	LB136305
	Silver	1240	1250	100	90 - 110	P	06/26/2025	18:48	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q2344

Contract: LOCK01 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Sodium	25500	25000	102	90 - 110	P	06/26/2025	18:48	LB136305
	Thallium	4950	5000	99	90 - 110	P	06/26/2025	18:48	LB136305
	Vanadium	2550	2500	102	90 - 110	P	06/26/2025	18:48	LB136305
	Zinc	2590	2500	104	90 - 110	P	06/26/2025	18:48	LB136305
CCV03	Antimony	4990	5000	100	90 - 110	P	06/26/2025	19:27	LB136305
	Arsenic	4980	5000	100	90 - 110	P	06/26/2025	19:27	LB136305
	Barium	10100	10000	101	90 - 110	P	06/26/2025	19:27	LB136305
	Beryllium	244	250	97	90 - 110	P	06/26/2025	19:27	LB136305
	Cadmium	2440	2500	98	90 - 110	P	06/26/2025	19:27	LB136305
	Chromium	1060	1000	106	90 - 110	P	06/26/2025	19:27	LB136305
	Cobalt	2440	2500	98	90 - 110	P	06/26/2025	19:27	LB136305
	Copper	1230	1250	99	90 - 110	P	06/26/2025	19:27	LB136305
	Iron	4860	5000	97	90 - 110	P	06/26/2025	19:27	LB136305
	Lead	4840	5000	97	90 - 110	P	06/26/2025	19:27	LB136305
	Manganese	2480	2500	99	90 - 110	P	06/26/2025	19:27	LB136305
	Nickel	2450	2500	98	90 - 110	P	06/26/2025	19:27	LB136305
	Potassium	25700	25000	103	90 - 110	P	06/26/2025	19:27	LB136305
	Selenium	4830	5000	97	90 - 110	P	06/26/2025	19:27	LB136305
	Silver	1240	1250	99	90 - 110	P	06/26/2025	19:27	LB136305
	Sodium	25500	25000	102	90 - 110	P	06/26/2025	19:27	LB136305
	Thallium	4910	5000	98	90 - 110	P	06/26/2025	19:27	LB136305
	Vanadium	2530	2500	101	90 - 110	P	06/26/2025	19:27	LB136305
	Zinc	2590	2500	103	90 - 110	P	06/26/2025	19:27	LB136305
CCV04	Antimony	5040	5000	101	90 - 110	P	06/26/2025	20:02	LB136305
	Arsenic	5060	5000	101	90 - 110	P	06/26/2025	20:02	LB136305
	Barium	10200	10000	102	90 - 110	P	06/26/2025	20:02	LB136305
	Beryllium	242	250	97	90 - 110	P	06/26/2025	20:02	LB136305
	Cadmium	2480	2500	99	90 - 110	P	06/26/2025	20:02	LB136305
	Chromium	1050	1000	105	90 - 110	P	06/26/2025	20:02	LB136305
	Cobalt	2480	2500	99	90 - 110	P	06/26/2025	20:02	LB136305
	Copper	1220	1250	98	90 - 110	P	06/26/2025	20:02	LB136305
	Iron	4780	5000	96	90 - 110	P	06/26/2025	20:02	LB136305
	Lead	4900	5000	98	90 - 110	P	06/26/2025	20:02	LB136305
	Manganese	2490	2500	100	90 - 110	P	06/26/2025	20:02	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Nickel	2470	2500	99	90 - 110	P	06/26/2025	20:02	LB136305
	Potassium	25200	25000	101	90 - 110	P	06/26/2025	20:02	LB136305
	Selenium	4880	5000	98	90 - 110	P	06/26/2025	20:02	LB136305
	Silver	1240	1250	99	90 - 110	P	06/26/2025	20:02	LB136305
	Sodium	25000	25000	100	90 - 110	P	06/26/2025	20:02	LB136305
	Thallium	4940	5000	99	90 - 110	P	06/26/2025	20:02	LB136305
	Vanadium	2560	2500	102	90 - 110	P	06/26/2025	20:02	LB136305
	Zinc	2600	2500	104	90 - 110	P	06/26/2025	20:02	LB136305
CCV05	Antimony	5100	5000	102	90 - 110	P	06/26/2025	20:46	LB136305
	Arsenic	5170	5000	103	90 - 110	P	06/26/2025	20:46	LB136305
	Barium	10200	10000	102	90 - 110	P	06/26/2025	20:46	LB136305
	Beryllium	252	250	101	90 - 110	P	06/26/2025	20:46	LB136305
	Cadmium	2500	2500	100	90 - 110	P	06/26/2025	20:46	LB136305
	Chromium	1050	1000	105	90 - 110	P	06/26/2025	20:46	LB136305
	Cobalt	2480	2500	99	90 - 110	P	06/26/2025	20:46	LB136305
	Copper	1270	1250	102	90 - 110	P	06/26/2025	20:46	LB136305
	Iron	4810	5000	96	90 - 110	P	06/26/2025	20:46	LB136305
	Lead	4950	5000	99	90 - 110	P	06/26/2025	20:46	LB136305
	Manganese	2490	2500	99	90 - 110	P	06/26/2025	20:46	LB136305
	Nickel	2480	2500	99	90 - 110	P	06/26/2025	20:46	LB136305
	Potassium	27200	25000	109	90 - 110	P	06/26/2025	20:46	LB136305
	Selenium	4990	5000	100	90 - 110	P	06/26/2025	20:46	LB136305
	Silver	1240	1250	100	90 - 110	P	06/26/2025	20:46	LB136305
	Sodium	26300	25000	105	90 - 110	P	06/26/2025	20:46	LB136305
	Thallium	4970	5000	100	90 - 110	P	06/26/2025	20:46	LB136305
	Vanadium	2550	2500	102	90 - 110	P	06/26/2025	20:46	LB136305
	Zinc	2640	2500	106	90 - 110	P	06/26/2025	20:46	LB136305
	Antimony	5050	5000	101	90 - 110	P	06/26/2025	21:39	LB136305
CCV06	Arsenic	5070	5000	102	90 - 110	P	06/26/2025	21:39	LB136305
	Barium	10100	10000	101	90 - 110	P	06/26/2025	21:39	LB136305
	Beryllium	244	250	98	90 - 110	P	06/26/2025	21:39	LB136305
	Cadmium	2470	2500	99	90 - 110	P	06/26/2025	21:39	LB136305
	Chromium	1070	1000	107	90 - 110	P	06/26/2025	21:39	LB136305
	Cobalt	2460	2500	98	90 - 110	P	06/26/2025	21:39	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q2344
Contract: LOCK01 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Copper	1240	1250	99	90 - 110	P	06/26/2025	21:39	LB136305
	Iron	4790	5000	96	90 - 110	P	06/26/2025	21:39	LB136305
	Lead	4880	5000	98	90 - 110	P	06/26/2025	21:39	LB136305
	Manganese	2480	2500	99	90 - 110	P	06/26/2025	21:39	LB136305
	Nickel	2470	2500	99	90 - 110	P	06/26/2025	21:39	LB136305
	Potassium	25500	25000	102	90 - 110	P	06/26/2025	21:39	LB136305
	Selenium	4880	5000	98	90 - 110	P	06/26/2025	21:39	LB136305
	Silver	1240	1250	99	90 - 110	P	06/26/2025	21:39	LB136305
	Sodium	25500	25000	102	90 - 110	P	06/26/2025	21:39	LB136305
	Thallium	4910	5000	98	90 - 110	P	06/26/2025	21:39	LB136305
	Vanadium	2560	2500	102	90 - 110	P	06/26/2025	21:39	LB136305
	Zinc	2610	2500	104	90 - 110	P	06/26/2025	21:39	LB136305
CCV07	Antimony	5010	5000	100	90 - 110	P	06/26/2025	22:31	LB136305
	Arsenic	5020	5000	100	90 - 110	P	06/26/2025	22:31	LB136305
	Barium	10000	10000	100	90 - 110	P	06/26/2025	22:31	LB136305
	Beryllium	240	250	96	90 - 110	P	06/26/2025	22:31	LB136305
	Cadmium	2450	2500	98	90 - 110	P	06/26/2025	22:31	LB136305
	Chromium	1060	1000	106	90 - 110	P	06/26/2025	22:31	LB136305
	Cobalt	2450	2500	98	90 - 110	P	06/26/2025	22:31	LB136305
	Copper	1210	1250	97	90 - 110	P	06/26/2025	22:31	LB136305
	Iron	4750	5000	95	90 - 110	P	06/26/2025	22:31	LB136305
	Lead	4840	5000	97	90 - 110	P	06/26/2025	22:31	LB136305
	Manganese	2450	2500	98	90 - 110	P	06/26/2025	22:31	LB136305
	Nickel	2450	2500	98	90 - 110	P	06/26/2025	22:31	LB136305
	Potassium	25100	25000	101	90 - 110	P	06/26/2025	22:31	LB136305
	Selenium	4790	5000	96	90 - 110	P	06/26/2025	22:31	LB136305
	Silver	1230	1250	98	90 - 110	P	06/26/2025	22:31	LB136305
	Sodium	25200	25000	101	90 - 110	P	06/26/2025	22:31	LB136305
	Thallium	4880	5000	98	90 - 110	P	06/26/2025	22:31	LB136305
	Vanadium	2520	2500	101	90 - 110	P	06/26/2025	22:31	LB136305
	Zinc	2600	2500	104	90 - 110	P	06/26/2025	22:31	LB136305
	Antimony	5010	5000	100	90 - 110	P	06/26/2025	22:49	LB136305
CCV08	Arsenic	5020	5000	100	90 - 110	P	06/26/2025	22:49	LB136305
	Barium	10200	10000	102	90 - 110	P	06/26/2025	22:49	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Beryllium	243	250	97	90 - 110	P	06/26/2025	22:49	LB136305
	Cadmium	2450	2500	98	90 - 110	P	06/26/2025	22:49	LB136305
	Chromium	1060	1000	106	90 - 110	P	06/26/2025	22:49	LB136305
	Cobalt	2450	2500	98	90 - 110	P	06/26/2025	22:49	LB136305
	Copper	1240	1250	99	90 - 110	P	06/26/2025	22:49	LB136305
	Iron	4800	5000	96	90 - 110	P	06/26/2025	22:49	LB136305
	Lead	4840	5000	97	90 - 110	P	06/26/2025	22:49	LB136305
	Manganese	2470	2500	99	90 - 110	P	06/26/2025	22:49	LB136305
	Nickel	2450	2500	98	90 - 110	P	06/26/2025	22:49	LB136305
	Potassium	25400	25000	102	90 - 110	P	06/26/2025	22:49	LB136305
	Selenium	4800	5000	96	90 - 110	P	06/26/2025	22:49	LB136305
	Silver	1230	1250	99	90 - 110	P	06/26/2025	22:49	LB136305
	Sodium	25500	25000	102	90 - 110	P	06/26/2025	22:49	LB136305
	Thallium	4880	5000	98	90 - 110	P	06/26/2025	22:49	LB136305
	Vanadium	2550	2500	102	90 - 110	P	06/26/2025	22:49	LB136305
	Zinc	2610	2500	104	90 - 110	P	06/26/2025	22:49	LB136305

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Antimony	52.9	50.0	106	65 - 135	P	06/23/2025	13:23	LB136236
	Arsenic	18.0	20.0	90	65 - 135	P	06/23/2025	13:23	LB136236
	Barium	100	100	100	65 - 135	P	06/23/2025	13:23	LB136236
	Beryllium	6.35	6.0	106	65 - 135	P	06/23/2025	13:23	LB136236
	Cadmium	5.84	6.0	97	65 - 135	P	06/23/2025	13:23	LB136236
	Chromium	10.1	10.0	101	65 - 135	P	06/23/2025	13:23	LB136236
	Cobalt	31.2	30.0	104	65 - 135	P	06/23/2025	13:23	LB136236
	Copper	23.3	20.0	116	65 - 135	P	06/23/2025	13:23	LB136236
	Iron	94.8	100	95	65 - 135	P	06/23/2025	13:23	LB136236
	Lead	11.8	12.0	98	65 - 135	P	06/23/2025	13:23	LB136236
	Manganese	20.4	20.0	102	65 - 135	P	06/23/2025	13:23	LB136236
	Nickel	40.7	40.0	102	65 - 135	P	06/23/2025	13:23	LB136236
	Potassium	1670	2000	83	65 - 135	P	06/23/2025	13:23	LB136236
	Selenium	17.5	20.0	88	65 - 135	P	06/23/2025	13:23	LB136236
	Silver	9.75	10.0	98	65 - 135	P	06/23/2025	13:23	LB136236
	Sodium	1870	2000	93	65 - 135	P	06/23/2025	13:23	LB136236
	Thallium	40.9	40.0	102	65 - 135	P	06/23/2025	13:23	LB136236
	Vanadium	34.3	40.0	86	65 - 135	P	06/23/2025	13:23	LB136236
	Zinc	40.3	40.0	101	65 - 135	P	06/23/2025	13:23	LB136236
CRI01	Antimony	53.0	50.0	106	65 - 135	P	06/26/2025	17:28	LB136305
	Arsenic	18.2	20.0	91	65 - 135	P	06/26/2025	17:28	LB136305
	Barium	100	100	100	65 - 135	P	06/26/2025	17:28	LB136305
	Beryllium	6.39	6.0	106	65 - 135	P	06/26/2025	17:28	LB136305
	Cadmium	5.73	6.0	96	65 - 135	P	06/26/2025	17:28	LB136305
	Chromium	11.3	10.0	113	65 - 135	P	06/26/2025	17:28	LB136305
	Cobalt	29.8	30.0	99	65 - 135	P	06/26/2025	17:28	LB136305
	Copper	22.8	20.0	114	65 - 135	P	06/26/2025	17:28	LB136305
	Iron	105	100	105	65 - 135	P	06/26/2025	17:28	LB136305
	Lead	11.9	12.0	99	65 - 135	P	06/26/2025	17:28	LB136305
	Manganese	21.6	20.0	108	65 - 135	P	06/26/2025	17:28	LB136305
	Nickel	39.3	40.0	98	65 - 135	P	06/26/2025	17:28	LB136305
	Potassium	1660	2000	83	65 - 135	P	06/26/2025	17:28	LB136305
	Selenium	20.5	20.0	102	65 - 135	P	06/26/2025	17:28	LB136305
	Silver	8.68	10.0	87	65 - 135	P	06/26/2025	17:28	LB136305
	Sodium	1900	2000	95	65 - 135	P	06/26/2025	17:28	LB136305
	Thallium	40.5	40.0	101	65 - 135	P	06/26/2025	17:28	LB136305

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Vanadium	33.1	40.0	83	65 - 135	P	06/26/2025	17:28	LB136305
	Zinc	40.9	40.0	102	65 - 135	P	06/26/2025	17:28	LB136305



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	13:19	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	13:19	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	13:19	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	13:19	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	13:19	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	13:19	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	13:19	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	13:19	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	13:19	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	13:19	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	13:19	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	13:19	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	13:19	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	13:19	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	13:19	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.		SDG No.: Q2344							
Contract: LOCK01	Lab Code: CHEM	Case No.: Q2344	SAS No.: Q2344						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	14:05	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	14:05	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	14:05	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	14:05	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	14:05	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	14:05	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	14:05	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	14:05	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	14:05	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	14:05	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	14:05	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
CCB02	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	14:58	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	14:58	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	14:58	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	14:58	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	14:58	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	14:58	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	14:58	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	14:58	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	14:58	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	14:58	LB136236
CCB03	Sodium	868	+/-1000	U	2000	P	06/23/2025	14:58	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
CCB03	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	15:52	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Lockwood, Kessler & Bartlett, Inc.</u>		SDG No.: <u>Q2344</u>							
Contract: <u>LOCK01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2344</u>	SAS No.: <u>Q2344</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	15:52	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	15:52	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	15:52	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	15:52	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	15:52	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	15:52	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	15:52	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	15:52	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	15:52	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	15:52	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
CCB04	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	16:56	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	16:56	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	16:56	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	16:56	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	16:56	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	16:56	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	16:56	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	16:56	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	16:56	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	16:56	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	16:56	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
CCB05	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	17:49	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	17:49	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.		SDG No.: Q2344							
Contract: LOCK01	Lab Code: CHEM	Case No.: Q2344	SAS No.: Q2344						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Barium	14.6	+/-50	U	100	P	06/23/2025	17:49	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	17:49	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	17:49	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	17:49	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	17:49	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	17:49	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	17:49	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	17:49	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	17:49	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	17:49	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	17:49	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	17:49	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	17:49	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
CCB06	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	18:43	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	18:43	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	18:43	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	18:43	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	18:43	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	18:43	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	18:43	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	18:43	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	18:43	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	18:43	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	18:43	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
CCB07	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	19:28	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	19:28	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	19:28	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	19:28	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	19:28	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	19:28	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	19:28	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	19:28	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	19:28	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	19:28	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	19:28	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	19:28	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	19:28	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	19:28	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	19:28	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	17:23	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	17:23	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	17:23	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	17:23	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	17:23	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	17:23	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	17:23	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	17:23	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	17:23	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	17:23	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	17:23	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	17:23	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.		SDG No.: Q2344							
Contract: LOCK01	Lab Code: CHEM	Case No.: Q2344	SAS No.: Q2344						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	17:59	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	17:59	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	17:59	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	17:59	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	17:59	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	17:59	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	17:59	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	17:59	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	17:59	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	17:59	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	17:59	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
CCB02	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	18:52	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	18:52	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	18:52	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	18:52	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	18:52	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	18:52	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	18:52	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	18:52	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	18:52	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	18:52	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	18:52	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
CCB03	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	19:31	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.		SDG No.: Q2344							
Contract: LOCK01	Lab Code: CHEM	Case No.: Q2344	SAS No.: Q2344						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	19:31	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	19:31	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	19:31	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	19:31	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	19:31	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	19:31	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	19:31	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	19:31	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	19:31	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	19:31	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
CCB04	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	20:07	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	20:07	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	20:07	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	20:07	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	20:07	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	20:07	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	20:07	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	20:07	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	20:07	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
CCB05	Silver	1.62	+/-5	U	10.0	P	06/26/2025	20:07	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	20:07	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Zinc	16.4	+/-20	J	40.0	P	06/26/2025	20:07	LB136305
	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	20:51	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	20:51	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.		SDG No.: Q2344							
Contract: LOCK01	Lab Code: CHEM	Case No.: Q2344	SAS No.: Q2344						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Barium	14.6	+/-50	U	100	P	06/26/2025	20:51	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	20:51	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	20:51	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	20:51	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	20:51	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	20:51	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	20:51	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	20:51	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	20:51	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	20:51	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
CCB06	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	21:43	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	21:43	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	21:43	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	21:43	LB136305
	Chromium	2.37	+/-5	J	10.0	P	06/26/2025	21:43	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	21:43	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	21:43	LB136305
	Lead	3.22	+/-6	J	12.0	P	06/26/2025	21:43	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	21:43	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	21:43	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	21:43	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
CCB07	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	22:35	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	22:35	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Lockwood, Kessler & Bartlett, Inc.</u>		SDG No.: <u>Q2344</u>							
Contract: <u>LOCK01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2344</u>	SAS No.: <u>Q2344</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	22:35	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	22:35	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	22:35	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	22:35	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	22:35	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	22:35	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	22:35	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	22:35	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	22:35	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
CCB08	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	22:53	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	22:53	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	22:53	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	22:53	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	22:53	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	22:53	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	22:53	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	22:53	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	22:53	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	22:53	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	22:53	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	22:53	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	22:53	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	22:53	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	22:53	LB136305

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Instrument: P5

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168551BL	WATER			Batch Number:	PB168551		Prep Date:	06/19/2025	
	Antimony	3.38	<12.5	U	25.0	P	06/26/2025	18:13	LB136305
	Arsenic	2.56	<5	U	10.0	P	06/26/2025	18:13	LB136305
	Barium	7.28	<25	U	50.0	P	06/26/2025	18:13	LB136305
	Beryllium	0.28	<1.5	U	3.00	P	06/26/2025	18:13	LB136305
	Cadmium	0.25	<1.5	U	3.00	P	06/26/2025	18:13	LB136305
	Chromium	1.06	<2.5	U	5.00	P	06/26/2025	18:13	LB136305
	Cobalt	1.13	<7.5	U	15.0	P	06/26/2025	18:13	LB136305
	Copper	2.30	<5	U	10.0	P	06/26/2025	18:13	LB136305
	Iron	11.7	<25	U	50.0	P	06/26/2025	18:13	LB136305
	Lead	1.15	<3	U	6.00	P	06/26/2025	18:13	LB136305
	Manganese	2.97	<5	U	10.0	P	06/26/2025	18:13	LB136305
	Nickel	1.53	<10	U	20.0	P	06/26/2025	18:13	LB136305
	Potassium	459	<500	U	1000	P	06/26/2025	18:13	LB136305
	Selenium	4.82	<5	U	10.0	P	06/26/2025	18:13	LB136305
	Silver	0.81	<2.5	U	5.00	P	06/26/2025	18:13	LB136305
	Sodium	434	<500	U	1000	P	06/26/2025	18:13	LB136305
	Thallium	2.19	<10	U	20.0	P	06/26/2025	18:13	LB136305
	Vanadium	3.13	<10	U	20.0	P	06/26/2025	18:13	LB136305
	Zinc	1.75	<10	U	20.0	P	06/26/2025	18:13	LB136305

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Antimony	19.1			-50	50	06/23/2025	13:34	LB136236
	Arsenic	-7.00			-20	20	06/23/2025	13:34	LB136236
	Barium	7.03	6.0	117	-94	106	06/23/2025	13:34	LB136236
	Beryllium	1.14			-6	6	06/23/2025	13:34	LB136236
	Cadmium	-1.06	1.0	106	-5	7	06/23/2025	13:34	LB136236
	Chromium	50.3	52.0	97	42	62	06/23/2025	13:34	LB136236
	Cobalt	27.1			-30	30	06/23/2025	13:34	LB136236
	Copper	-9.01	2.0	450	-18	22	06/23/2025	13:34	LB136236
	Iron	99900	100000	100	85600	116500	06/23/2025	13:34	LB136236
	Lead	-2.88			-12	12	06/23/2025	13:34	LB136236
	Manganese	6.23	7.0	89	-13	27	06/23/2025	13:34	LB136236
	Nickel	4.60	2.0	230	-38	42	06/23/2025	13:34	LB136236
	Potassium	8.86			0	0	06/23/2025	13:34	LB136236
	Selenium	-1.75			-20	20	06/23/2025	13:34	LB136236
	Silver	-0.54			-10	10	06/23/2025	13:34	LB136236
	Sodium	200			0	0	06/23/2025	13:34	LB136236
	Thallium	6.79			-40	40	06/23/2025	13:34	LB136236
	Vanadium	6.95			-40	40	06/23/2025	13:34	LB136236
	Zinc	11.2			-40	40	06/23/2025	13:34	LB136236
ICSAB01	Antimony	618	620	100	525	711	06/23/2025	13:43	LB136236
	Arsenic	94.5	100	94	88.4	120	06/23/2025	13:43	LB136236
	Barium	505	540	94	437	637	06/23/2025	13:43	LB136236
	Beryllium	513	500	103	420	570	06/23/2025	13:43	LB136236
	Cadmium	1030	970	106	826	1120	06/23/2025	13:43	LB136236
	Chromium	543	540	101	460	624	06/23/2025	13:43	LB136236
	Cobalt	531	480	111	404	548	06/23/2025	13:43	LB136236
	Copper	478	510	94	434	588	06/23/2025	13:43	LB136236
	Iron	96200	99000	97	84400	114500	06/23/2025	13:43	LB136236
	Lead	45.9	49.0	94	37	61	06/23/2025	13:43	LB136236
	Manganese	486	510	95	430	584	06/23/2025	13:43	LB136236
	Nickel	1010	950	106	810	1100	06/23/2025	13:43	LB136236
	Potassium	5.91			0	0	06/23/2025	13:43	LB136236
	Selenium	52.0	46.0	113	26	66	06/23/2025	13:43	LB136236
	Silver	209	200	104	170	232	06/23/2025	13:43	LB136236
	Sodium	192			0	0	06/23/2025	13:43	LB136236
	Thallium	98.9	110	90	68	148	06/23/2025	13:43	LB136236
	Vanadium	490	490	100	417	565	06/23/2025	13:43	LB136236
	Zinc	1010	950	106	809	1095	06/23/2025	13:43	LB136236
ICSA01	Antimony	17.7			-50	50	06/26/2025	17:37	LB136305
	Arsenic	-0.86			-20	20	06/26/2025	17:37	LB136305
	Barium	5.56	6.0	93	-94	106	06/26/2025	17:37	LB136305
	Beryllium	1.03			-6	6	06/26/2025	17:37	LB136305

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-1.09	1.0	109	-5	7	06/26/2025	17:37	LB136305
	Chromium	50.8	52.0	98	42	62	06/26/2025	17:37	LB136305
	Cobalt	27.8			-30	30	06/26/2025	17:37	LB136305
	Copper	-5.23	2.0	262	-18	22	06/26/2025	17:37	LB136305
	Iron	95600	100000	96	85600	116500	06/26/2025	17:37	LB136305
	Lead	2.20			-12	12	06/26/2025	17:37	LB136305
	Manganese	5.18	7.0	74	-13	27	06/26/2025	17:37	LB136305
	Nickel	3.31	2.0	166	-38	42	06/26/2025	17:37	LB136305
	Potassium	4.58			0	0	06/26/2025	17:37	LB136305
	Selenium	6.56			-20	20	06/26/2025	17:37	LB136305
	Silver	2.97			-10	10	06/26/2025	17:37	LB136305
	Sodium	157			0	0	06/26/2025	17:37	LB136305
	Thallium	16.0			-40	40	06/26/2025	17:37	LB136305
	Vanadium	3.94			-40	40	06/26/2025	17:37	LB136305
	Zinc	12.0			-40	40	06/26/2025	17:37	LB136305
ICSAB01	Antimony	608	620	98	525	711	06/26/2025	17:42	LB136305
	Arsenic	92.6	100	93	88.4	120	06/26/2025	17:42	LB136305
	Barium	499	540	92	437	637	06/26/2025	17:42	LB136305
	Beryllium	494	500	99	420	570	06/26/2025	17:42	LB136305
	Cadmium	970	970	100	826	1120	06/26/2025	17:42	LB136305
	Chromium	541	540	100	460	624	06/26/2025	17:42	LB136305
	Cobalt	503	480	105	404	548	06/26/2025	17:42	LB136305
	Copper	468	510	92	434	588	06/26/2025	17:42	LB136305
	Iron	96300	99000	97	84400	114500	06/26/2025	17:42	LB136305
	Lead	49.4	49.0	101	37	61	06/26/2025	17:42	LB136305
	Manganese	484	510	95	430	584	06/26/2025	17:42	LB136305
	Nickel	947	950	100	810	1100	06/26/2025	17:42	LB136305
	Potassium	4.86			0	0	06/26/2025	17:42	LB136305
	Selenium	52.3	46.0	114	26	66	06/26/2025	17:42	LB136305
	Silver	211	200	106	170	232	06/26/2025	17:42	LB136305
	Sodium	163			0	0	06/26/2025	17:42	LB136305
	Thallium	95.2	110	86	68	148	06/26/2025	17:42	LB136305
	Vanadium	470	490	96	417	565	06/26/2025	17:42	LB136305
	Zinc	1010	950	106	809	1095	06/26/2025	17:42	LB136305



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Lockwood, Kessler & Bartlett, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q2344</u>
contract: <u>LOCK01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2344</u> sas no.: <u>Q2344</u>
matrix: <u>Water</u>	sample id: <u>Q2344-07</u>	client id: <u>MW-2MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2344-07MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	ug/L	75 - 125	414		5.08	J	400	102		P
Arsenic	ug/L	75 - 125	427		10.0	U	400	107		P
Barium	ug/L	75 - 125	264		166		100	98		P
Beryllium	ug/L	75 - 125	86.5		3.00	U	100	86		P
Cadmium	ug/L	75 - 125	106		3.00	U	100	106		P
Chromium	ug/L	75 - 125	210		4.01	J	200	103		P
Cobalt	ug/L	75 - 125	118		12.6	J	100	105		P
Copper	ug/L	75 - 125	147		28.2		150	79		P
Iron	ug/L	75 - 125	7150		5700		1500	97		P
Lead	ug/L	75 - 125	513		9.67		500	101		P
Manganese	ug/L	75 - 125	938		849		100	88		P
Nickel	ug/L	75 - 125	287		26.2		250	104		P
Potassium	ug/L	75 - 125	28100		23200		5000	97		P
Selenium	ug/L	75 - 125	986		10.0	U	1000	99		P
Silver	ug/L	75 - 125	37.0		5.00	U	37.5	99		P
Sodium	ug/L	75 - 125	199000		201000		1500	-150		P
Thallium	ug/L	75 - 125	1010		3.38	J	1000	100		P
Vanadium	ug/L	75 - 125	150		20.0	U	150	100		P
Zinc	ug/L	75 - 125	203		95.0		100	108		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Lockwood, Kessler & Bartlett, Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q2344</u>
contract:	<u>LOCK01</u>	lab code:	<u>CHEM</u>	case no.:	<u>Q2344</u>
matrix:	<u>Water</u>	sample id:	<u>Q2344-07</u>	client id:	<u>MW-2MSD</u>
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>Q2344-07MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	ug/L	75 - 125	419		5.08	J	400	103		P
Arsenic	ug/L	75 - 125	429		10.0	U	400	107		P
Barium	ug/L	75 - 125	262		166		100	96		P
Beryllium	ug/L	75 - 125	101		3.00	U	100	101		P
Cadmium	ug/L	75 - 125	107		3.00	U	100	107		P
Chromium	ug/L	75 - 125	207		4.01	J	200	101		P
Cobalt	ug/L	75 - 125	118		12.6	J	100	106		P
Copper	ug/L	75 - 125	174		28.2		150	97		P
Iron	ug/L	75 - 125	7090		5700		1500	93		P
Lead	ug/L	75 - 125	517		9.67		500	102		P
Manganese	ug/L	75 - 125	932		849		100	83		P
Nickel	ug/L	75 - 125	291		26.2		250	106		P
Potassium	ug/L	75 - 125	28100		23200		5000	97		P
Selenium	ug/L	75 - 125	997		10.0	U	1000	100		P
Silver	ug/L	75 - 125	36.1		5.00	U	37.5	96		P
Sodium	ug/L	75 - 125	196000		201000		1500	-334		P
Thallium	ug/L	75 - 125	1020		3.38	J	1000	101		P
Vanadium	ug/L	75 - 125	152		20.0	U	150	101		P
Zinc	ug/L	75 - 125	195		95.0		100	100		P



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals
- 5b -

Client:

Lockwood, Kessler & Bartlett, Inc.

SDG No.:

Q2344

Contract:

LOCK01

Lab Code:

CHEM

Case No.:

Q2344

SAS No.:

Q2344

Matrix:

Level:

LOW

Client ID:

Sample ID:

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
Matrix: Water **Sample ID:** Q2344-07 **Client ID:** MW-2DUP
Percent Solids for Sample: NA **Duplicate ID** Q2344-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	5.08	J	3.86	J	27		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	166		162		2		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	0.26	J	200.0		P
Chromium	ug/L	20	4.01	J	4.25	J	6		P
Cobalt	ug/L	20	12.6	J	12.5	J	1		P
Copper	ug/L	20	28.2		27.4		3		P
Iron	ug/L	20	5700		5540		3		P
Lead	ug/L	20	9.67		9.89		2		P
Manganese	ug/L	20	849		830		2		P
Nickel	ug/L	20	26.2		25.6		2		P
Potassium	ug/L	20	23200		22800		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	201000		199000		1		P
Thallium	ug/L	20	3.38	J	3.17	J	6		P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	95.0		93.2		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
Matrix: Water **Sample ID:** Q2344-07MS **Client ID:** MW-2MSD
Percent Solids for Sample: NA **Duplicate ID** Q2344-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	414		419		1		P
Arsenic	ug/L	20	427		429		0		P
Barium	ug/L	20	264		262		1		P
Beryllium	ug/L	20	86.5		101		15		P
Cadmium	ug/L	20	106		107		1		P
Chromium	ug/L	20	210		207		1		P
Cobalt	ug/L	20	118		118		0		P
Copper	ug/L	20	147		174		17		P
Iron	ug/L	20	7150		7090		1		P
Lead	ug/L	20	513		517		1		P
Manganese	ug/L	20	938		932		1		P
Nickel	ug/L	20	287		291		1		P
Potassium	ug/L	20	28100		28100		0		P
Selenium	ug/L	20	986		997		1		P
Silver	ug/L	20	37.0		36.1		2		P
Sodium	ug/L	20	199000		196000		2		P
Thallium	ug/L	20	1010		1020		1		P
Vanadium	ug/L	20	150		152		1		P
Zinc	ug/L	20	203		195		4		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168551BS							
Antimony	ug/L	400	400		100	80 - 120	P
Arsenic	ug/L	400	361		90	80 - 120	P
Barium	ug/L	100	99.9		100	80 - 120	P
Beryllium	ug/L	100	99.8		100	80 - 120	P
Cadmium	ug/L	100	93.2		93	80 - 120	P
Chromium	ug/L	200	199		100	80 - 120	P
Cobalt	ug/L	100	95.3		95	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Iron	ug/L	1500	1500		100	80 - 120	P
Lead	ug/L	500	465		93	80 - 120	P
Manganese	ug/L	100	103		103	80 - 120	P
Nickel	ug/L	250	238		95	80 - 120	P
Potassium	ug/L	5000	4340		87	80 - 120	P
Selenium	ug/L	1000	962		96	80 - 120	P
Silver	ug/L	37.5	34.7		92	80 - 120	P
Sodium	ug/L	1500	1430		95	80 - 120	P
Thallium	ug/L	1000	986		99	80 - 120	P
Vanadium	ug/L	150	146		97	80 - 120	P
Zinc	ug/L	100	103		103	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

MW-2L

Lab Name: Chemtech Consulting Group

Contract: LOCK01

Lab Code: CHEM **Lb No.:** lb136236

Lab Sample ID : Q2344-07L **SDG No.:** Q2344

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
		C		C			
Antimony	5.08	J	125	U	100.0		P
Arsenic	10.0	U	50.0	U			P
Barium	166		165	J	1		P
Beryllium	3.00	U	15.0	U			P
Cadmium	3.00	U	15.0	U			P
Chromium	4.01	J	25.0	U	100.0		P
Cobalt	12.6	J	35.3	J	179		P
Copper	28.2		32.7	J	16		P
Iron	5700		5690		0		P
Lead	9.67		17.6	J	82		P
Manganese	849		862		2		P
Nickel	26.2		25.0	J	5		P
Potassium	23200		20300		13		P
Selenium	10.0	U	50.0	U			P
Silver	5.00	U	25.0	U			P
Sodium	201000		231000		15		P
Thallium	3.38	J	100	U	100.0		P
Vanadium	20.0	U	100	U			P
Zinc	95.0		90.9	J	4		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000720	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0001020
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	-0.0004810
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>Lockwood, Kessler & Bartlett, Inc.</u>	SDG No.:	<u>Q2344</u>
Contract:	<u>LOCK01</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>Q2344</u>
		SAS No.:	<u>Q2344</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB168551							
PB168551BL	PB168551BL	MB	WATER	06/19/2025	50.0	25.0	
PB168551BS	PB168551BS	LCS	WATER	06/19/2025	50.0	25.0	
Q2344-01	MW-1	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-02	MW-1	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-03	MW-3	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-04	MW-3	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-05	MW-4	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-06	MW-4	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-07	MW-2	SAM	WATER	06/19/2025	50.0	25.0	
Q2344-07DUP	MW-2DUP	DUP	WATER	06/19/2025	50.0	25.0	
Q2344-07MS	MW-2MS	MS	WATER	06/19/2025	50.0	25.0	
Q2344-07MSD	MW-2MSD	MSD	WATER	06/19/2025	50.0	25.0	
Q2344-08	MW-2	SAM	WATER	06/19/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Lockwood, Kessler & Bartlett, Inc.

Contract: LOCK01

Lab code: CHEM

Case no.: Q2344

Sas no.: Q2344

Sdg no.: Q2344

Instrument id number:

Method:

Run number: LB136236

Start date: 06/23/2025

End date: 06/23/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1153	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1158	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1202	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1207	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1211	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1215	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1234	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1309	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1319	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1323	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1334	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1343	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1356	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1405	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1454	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1458	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1547	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1552	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1651	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1656	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1745	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1749	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-01	MW-1	1	1816	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-02	MW-1	1	1820	Fe,Mn
Q2344-03	MW-3	1	1825	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-04	MW-3	1	1830	Fe,Mn
Q2344-05	MW-4	1	1834	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1839	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1843	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-06	MW-4	1	1848	Fe,Mn
Q2344-07	MW-2	1	1852	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-07DUP	MW-2DUP	1	1857	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-07L	MW-2L	5	1902	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-07MS	MW-2MS	1	1906	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-07MSD	MW-2MSD	1	1910	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q2344-08	MW-2	1	1919	Fe,Mn
CCV07	CCV07	1	1924	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	1928	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Lockwood, Kessler & Bartlett, Inc.

Contract: LOCK01

Lab code: CHEM **Case no.:** Q2344 **Sas no.:** Q2344

Sdg no.: Q2344

Instrument id number: **Method:**

Run number: LB136305

Start date: 06/26/2025 **End date:** 06/26/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1641	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1645	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1650	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1654	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1658	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1703	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1707	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1719	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1723	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1728	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1737	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1742	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1755	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1759	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB168551BL	PB168551BL	1	1813	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB168551BS	PB168551BS	1	1817	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1848	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1852	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1927	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1931	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	2002	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	2007	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	2046	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	2051	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	2139	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	2143	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	2231	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	2235	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV08	CCV08	1	2249	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB08	CCB08	1	2253	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn