



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	13:19	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	13:19	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	13:19	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	13:19	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	13:19	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	13:19	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	13:19	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	13:19	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	13:19	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	13:19	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	13:19	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	13:19	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	13:19	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	13:19	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	13:19	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	13:19	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	14:05	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	14:05	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	14:05	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	14:05	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	14:05	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	14:05	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	14:05	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	14:05	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	14:05	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	14:05	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	14:05	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	14:05	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	14:05	LB136236
CCB02	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	14:58	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	14:58	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	14:58	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	14:58	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	14:58	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	14:58	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	14:58	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	14:58	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	14:58	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	14:58	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	14:58	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	14:58	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	14:58	LB136236
CCB03	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	15:52	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	15:52	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	15:52	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	15:52	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	15:52	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	15:52	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	15:52	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	15:52	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	15:52	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	15:52	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	15:52	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	15:52	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	15:52	LB136236
CCB04	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	16:56	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	16:56	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	16:56	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	16:56	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	16:56	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	16:56	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	16:56	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	16:56	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	16:56	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	16:56	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	16:56	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	16:56	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	16:56	LB136236
CCB05	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	17:49	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	17:49	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Barium	14.6	+/-50	U	100	P	06/23/2025	17:49	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	17:49	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	17:49	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	17:49	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	17:49	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	17:49	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	17:49	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	17:49	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	17:49	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	17:49	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	17:49	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	17:49	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	17:49	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	17:49	LB136236
CCB06	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	18:43	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	18:43	LB136236
	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	18:43	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	18:43	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	18:43	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	18:43	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	18:43	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	18:43	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	18:43	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	18:43	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	18:43	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	18:43	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	18:43	LB136236
CCB07	Antimony	6.76	+/-25	U	50.0	P	06/23/2025	19:28	LB136236
	Arsenic	5.12	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Barium	14.6	+/-50	U	100	P	06/23/2025	19:28	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Beryllium	0.56	+/-3	U	6.00	P	06/23/2025	19:28	LB136236
	Cadmium	0.50	+/-3	U	6.00	P	06/23/2025	19:28	LB136236
	Chromium	2.12	+/-5	U	10.0	P	06/23/2025	19:28	LB136236
	Cobalt	2.26	+/-15	U	30.0	P	06/23/2025	19:28	LB136236
	Copper	4.60	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Iron	23.4	+/-50	U	100	P	06/23/2025	19:28	LB136236
	Lead	2.30	+/-6	U	12.0	P	06/23/2025	19:28	LB136236
	Manganese	5.94	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Nickel	3.06	+/-20	U	40.0	P	06/23/2025	19:28	LB136236
	Potassium	918	+/-1000	U	2000	P	06/23/2025	19:28	LB136236
	Selenium	9.64	+/-10	U	20.0	P	06/23/2025	19:28	LB136236
	Silver	1.62	+/-5	U	10.0	P	06/23/2025	19:28	LB136236
	Sodium	868	+/-1000	U	2000	P	06/23/2025	19:28	LB136236
	Thallium	4.38	+/-20	U	40.0	P	06/23/2025	19:28	LB136236
	Vanadium	6.26	+/-20	U	40.0	P	06/23/2025	19:28	LB136236
	Zinc	3.50	+/-20	U	40.0	P	06/23/2025	19:28	LB136236

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	17:23	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	17:23	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	17:23	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	17:23	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	17:23	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	17:23	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	17:23	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	17:23	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	17:23	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	17:23	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	17:23	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	17:23	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	17:59	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	17:59	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	17:59	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	17:59	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	17:59	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	17:59	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	17:59	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	17:59	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	17:59	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	17:59	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	17:59	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
CCB02	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	18:52	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	18:52	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	18:52	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	18:52	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	18:52	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	18:52	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	18:52	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	18:52	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	18:52	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	18:52	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	18:52	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
CCB03	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	19:31	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	19:31	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	19:31	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	19:31	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	19:31	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	19:31	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	19:31	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	19:31	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	19:31	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	19:31	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	19:31	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
CCB04	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	20:07	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	20:07	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	20:07	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	20:07	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	20:07	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	20:07	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	20:07	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	20:07	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	20:07	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
CCB05	Silver	1.62	+/-5	U	10.0	P	06/26/2025	20:07	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	20:07	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Zinc	16.4	+/-20	J	40.0	P	06/26/2025	20:07	LB136305
	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	20:51	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	20:51	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Barium	14.6	+/-50	U	100	P	06/26/2025	20:51	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	20:51	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	20:51	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	20:51	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	20:51	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	20:51	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	20:51	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	20:51	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	20:51	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	20:51	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
CCB06	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	21:43	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	21:43	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	21:43	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	21:43	LB136305
	Chromium	2.37	+/-5	J	10.0	P	06/26/2025	21:43	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	21:43	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	21:43	LB136305
	Lead	3.22	+/-6	J	12.0	P	06/26/2025	21:43	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	21:43	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	21:43	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	21:43	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	21:43	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	21:43	LB136305
CCB07	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	22:35	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	22:35	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q2344

Contract: LOCK01

Lab Code: CHEM

Case No.: Q2344

SAS No.: Q2344

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	22:35	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	22:35	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	22:35	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	22:35	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	22:35	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	22:35	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	22:35	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	22:35	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	22:35	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	22:35	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	22:35	LB136305
CCB08	Antimony	6.76	+/-25	U	50.0	P	06/26/2025	22:53	LB136305
	Arsenic	5.12	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Barium	14.6	+/-50	U	100	P	06/26/2025	22:53	LB136305
	Beryllium	0.56	+/-3	U	6.00	P	06/26/2025	22:53	LB136305
	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	22:53	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	22:53	LB136305
	Cobalt	2.26	+/-15	U	30.0	P	06/26/2025	22:53	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Iron	23.4	+/-50	U	100	P	06/26/2025	22:53	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	22:53	LB136305
	Manganese	5.94	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	22:53	LB136305
	Potassium	918	+/-1000	U	2000	P	06/26/2025	22:53	LB136305
	Selenium	9.64	+/-10	U	20.0	P	06/26/2025	22:53	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	22:53	LB136305
	Sodium	868	+/-1000	U	2000	P	06/26/2025	22:53	LB136305
	Thallium	4.38	+/-20	U	40.0	P	06/26/2025	22:53	LB136305
	Vanadium	6.26	+/-20	U	40.0	P	06/26/2025	22:53	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	22:53	LB136305