

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
Matrix: Water **Sample ID:** Q2344-07 **Client ID:** MW-2DUP
Percent Solids for Sample: NA **Duplicate ID** Q2344-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	5.08	J	3.86	J	27		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	166		162		2		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	0.26	J	200.0		P
Chromium	ug/L	20	4.01	J	4.25	J	6		P
Cobalt	ug/L	20	12.6	J	12.5	J	1		P
Copper	ug/L	20	28.2		27.4		3		P
Iron	ug/L	20	5700		5540		3		P
Lead	ug/L	20	9.67		9.89		2		P
Manganese	ug/L	20	849		830		2		P
Nickel	ug/L	20	26.2		25.6		2		P
Potassium	ug/L	20	23200		22800		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	201000		199000		1		P
Thallium	ug/L	20	3.38	J	3.17	J	6		P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	95.0		93.2		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344
Matrix: Water **Sample ID:** Q2344-07MS **Client ID:** MW-2MSD
Percent Solids for Sample: NA **Duplicate ID** Q2344-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	414		419		1		P
Arsenic	ug/L	20	427		429		0		P
Barium	ug/L	20	264		262		1		P
Beryllium	ug/L	20	86.5		101		15		P
Cadmium	ug/L	20	106		107		1		P
Chromium	ug/L	20	210		207		1		P
Cobalt	ug/L	20	118		118		0		P
Copper	ug/L	20	147		174		17		P
Iron	ug/L	20	7150		7090		1		P
Lead	ug/L	20	513		517		1		P
Manganese	ug/L	20	938		932		1		P
Nickel	ug/L	20	287		291		1		P
Potassium	ug/L	20	28100		28100		0		P
Selenium	ug/L	20	986		997		1		P
Silver	ug/L	20	37.0		36.1		2		P
Sodium	ug/L	20	199000		196000		2		P
Thallium	ug/L	20	1010		1020		1		P
Vanadium	ug/L	20	150		152		1		P
Zinc	ug/L	20	203		195		4		P