

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q2344
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q2344 **SAS No.:** Q2344

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168551BS							
Antimony	ug/L	400	400		100	80 - 120	P
Arsenic	ug/L	400	361		90	80 - 120	P
Barium	ug/L	100	99.9		100	80 - 120	P
Beryllium	ug/L	100	99.8		100	80 - 120	P
Cadmium	ug/L	100	93.2		93	80 - 120	P
Chromium	ug/L	200	199		100	80 - 120	P
Cobalt	ug/L	100	95.3		95	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Iron	ug/L	1500	1500		100	80 - 120	P
Lead	ug/L	500	465		93	80 - 120	P
Manganese	ug/L	100	103		103	80 - 120	P
Nickel	ug/L	250	238		95	80 - 120	P
Potassium	ug/L	5000	4340		87	80 - 120	P
Selenium	ug/L	1000	962		96	80 - 120	P
Silver	ug/L	37.5	34.7		92	80 - 120	P
Sodium	ug/L	1500	1430		95	80 - 120	P
Thallium	ug/L	1000	986		99	80 - 120	P
Vanadium	ug/L	150	146		97	80 - 120	P
Zinc	ug/L	100	103		103	80 - 120	P