

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LOCK01

Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344 SDG No.: Q2344

Instrument ID: MSVOA_X Calibration Date/Time: 06/18/2025 09:33

Lab File ID: VX046728.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methyl Iodide	50.000	45.95	8.099998	20