



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
Data File : VX046727.D
Acq On : 18 Jun 2025 08:59
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Title : SW846 8260
Last Update : Wed Jun 18 03:09:16 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	9538	PASS
75	95	30	60	50.8	25608	PASS
95	95	100	100	100.0	50435	PASS
96	95	5	9	6.8	3435	PASS
173	174	0.00	2	0.6	216	PASS
174	95	50	100	74.0	37317	PASS
175	174	5	9	7.5	2809	PASS
176	174	95	101	96.5	36000	PASS
177	176	5	9	6.3	2277	PASS