

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LOCK01
 Lab Code: CHEM Case No.: Q2344 SAS No.: Q2344 SDG NO.: Q2344
 Lab File ID: VX046728.D Date Analyzed: 06/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:33
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	104970	12.018				
UPPER LIMIT	209940	12.518				
LOWER LIMIT	52485	11.518				
EPA SAMPLE NO.						
MW-1	143592	12.02				
MW-3	138013	12.02				
MW-4	130790	12.02				
MW-2	148475	12.02				
TRIP-BLANK	93543	12.02				
VX0618WBL01	121135	12.02				
VX0618WBS01	121898	12.02				
VX0618WBSD01	112227	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.