

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062625\
 Data File : BN037401.D
 Acq On : 26 Jun 2025 19:42
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 27 03:00:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 26 16:06:33 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2	1,4-Dioxane	0.379	0.416	-9.8	79	0.00
3	n-Nitrosodimethylamine	0.375	0.371	1.1	69	0.00
4 S	2-Fluorophenol	0.771	0.791	-2.6	69	0.00
5 S	Phenol-d6	0.799	0.773	3.3	67	0.00
6	bis(2-Chloroethyl)ether	0.711	0.579	18.6	56	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	69	0.00
8 S	Nitrobenzene-d5	0.321	0.299	6.9	66	0.00
9	Naphthalene	1.030	1.011	1.8	70	-0.01
10	Hexachlorobutadiene	0.409	0.434	-6.1	72	0.00
11 SURR	2-Methylnaphthalene-d10	0.619	0.588	5.0	70	0.00
12	2-Methylnaphthalene	0.709	0.678	4.4	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.235	0.263	-11.9	78	0.00
15 S	2-Fluorobiphenyl	1.727	1.679	2.8	69	0.00
16	Acenaphthylene	1.659	1.570	5.4	71	-0.01
17	Acenaphthene	1.083	1.045	3.5	71	0.00
18	Fluorene	1.530	1.498	2.1	72	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
20	4,6-Dinitro-2-methylphenol	0.103	0.101	1.9	92	0.00
21	4-Bromophenyl-phenylether	0.282	0.271	3.9	78	0.00
22	Hexachlorobenzene	0.303	0.301	0.7	77	0.00
23	Atrazine	0.224	0.225	-0.4	81	0.00
24	Pentachlorophenol	0.164	0.159	3.0	79	0.00
25	Phenanthrene	1.127	1.057	6.2	75	0.00
26	Anthracene	1.036	0.943	9.0	75	0.00
27 SURR	Fluoranthene-d10	1.146	1.129	1.5	82	0.00
28	Fluoranthene	1.439	1.460	-1.5	83	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
30	Pyrene	1.514	1.445	4.6	83	0.00
31 S	Terphenyl-d14	0.853	0.857	-0.5	87	0.00
32	Benzo(a)anthracene	1.257	1.265	-0.6	95	0.00
33	Chrysene	1.556	1.535	1.3	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.498	0.559	-12.2	97	0.00
35 I	Perylene-d12	1.000	1.000	0.0	91	0.00
36	Indeno(1,2,3-cd)pyrene	1.686	1.599	5.2	92	0.00
37	Benzo(b)fluoranthene	1.410	1.330	5.7	92	0.00
38	Benzo(k)fluoranthene	1.507	1.476	2.1	92	0.00
39 C	Benzo(a)pyrene	1.254	1.175	6.3	90	0.00
40	Dibenzo(a,h)anthracene	1.296	1.172	9.6	88	0.00
41	Benzo(g,h,i)perylene	1.548	1.496	3.4	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4