

LAB CHRONICLE

OrderID:	Q2345	OrderDate:	6/17/2025 12:40:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	D61, VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2345-01	EB02-20250616	Water	SVOC-SIMGroup1	8270-Modified	06/16/25	06/20/25	06/26/25	06/17/25
Q2345-13	TT189D2-20250617	Water	SVOC-SIMGroup1	8270-Modified	06/17/25	06/20/25	06/26/25	06/17/25
Q2345-13DL	TT189D2-20250617DL	Water	SVOC-SIMGroup1	8270-Modified	06/17/25	06/20/25	06/27/25	06/17/25
Q2345-14	TT150S1-20250617	Water	SVOC-SIMGroup1	8270-Modified	06/17/25	06/20/25	06/26/25	06/17/25
Q2345-15	RW8-MW01D1-20250617	Water	SVOC-SIMGroup1	8270-Modified	06/17/25	06/20/25	06/26/25	06/17/25
Q2345-16	TT192D2-20250617	Water	SVOC-SIMGroup1	8270-Modified	06/17/25	06/20/25	06/26/25	06/17/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q2345
Client: AECOM Technical Services, Inc.

Sample ID	Client ID		Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : TT189D2-20250617									
Q2345-13	TT189D2-20250617	WATER	1,4-Dioxane	5.300	E	0.07	0.2	0.2	ug/L
			Total Svoc :			5.30			
			Total Concentration:			5.30			
Client ID : TT189D2-20250617DL									
Q2345-13DL	TT189D2-20250617DL	WATER	1,4-Dioxane	5.600	D	0.13	0.4	0.4	ug/L
			Total Svoc :			5.60			
			Total Concentration:			5.60			
Client ID : TT150S1-20250617									
Q2345-14	TT150S1-20250617	WATER	1,4-Dioxane	0.240		0.07	0.2	0.2	ug/L
			Total Svoc :			0.24			
			Total Concentration:			0.24			
Client ID : RW8-MW01D1-20250617									
Q2345-15	RW8-MW01D1-2025061	WATER	1,4-Dioxane	0.230		0.07	0.2	0.2	ug/L
			Total Svoc :			0.23			
			Total Concentration:			0.23			
Client ID : TT192D2-20250617									
Q2345-16	TT192D2-20250617	WATER	1,4-Dioxane	0.920		0.07	0.2	0.2	ug/L
			Total Svoc :			0.92			
			Total Concentration:			0.92			



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/16/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/17/25
Client Sample ID:	EB02-20250616		SDG No.:	Q2345
Lab Sample ID:	Q2345-01		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037395.D	1	06/20/25 09:32	06/26/25 16:06	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		99%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		75%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.44	*	53 - 106		111%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.47		58 - 132		118%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1750		7.568			
1146-65-2	Naphthalene-d8	3970		10.34			
15067-26-2	Acenaphthene-d10	2670		14.213			
1517-22-2	Phenanthrene-d10	5240		16.971			
1719-03-5	Chrysene-d12	4860		21.162			
1520-96-3	Perylene-d12	4950		23.351			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/17/25	
Client Sample ID:	TT189D2-20250617		SDG No.:	Q2345	
Lab Sample ID:	Q2345-13		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	990	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037396.D	1	06/20/25 09:32	06/26/25 16:42	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	5.30	E	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		82%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		101%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		69%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.60	*	53 - 106		150%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.44		58 - 132		110%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1680		7.568			
1146-65-2	Naphthalene-d8	3640		10.34			
15067-26-2	Acenaphthene-d10	2500		14.213			
1517-22-2	Phenanthrene-d10	5120		16.971			
1719-03-5	Chrysene-d12	5270		21.162			
1520-96-3	Perylene-d12	5670		23.348			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/17/25	
Client Sample ID:	TT189D2-20250617DL		SDG No.:	Q2345	
Lab Sample ID:	Q2345-13DL		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	990	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037419.D	2	06/20/25 09:32	06/27/25 11:21	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	5.60	D	0.13	0.40	0.40	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		89%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150		112%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		69%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.58	*	53 - 106		146%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.47		58 - 132		117%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1830		7.56			
1146-65-2	Naphthalene-d8	3780		10.34			
15067-26-2	Acenaphthene-d10	2820		14.213			
1517-22-2	Phenanthrene-d10	5810		16.971			
1719-03-5	Chrysene-d12	6100		21.162			
1520-96-3	Perylene-d12	6390		23.345			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/17/25
Client Sample ID:	TT150S1-20250617		SDG No.:	Q2345
Lab Sample ID:	Q2345-14		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037397.D	1	06/20/25 09:32	06/26/25 17:18	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.24		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		81%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		97%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.29		55 - 111		72%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.55	*	53 - 106		137%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.49		58 - 132		122%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1590	7.568				
1146-65-2	Naphthalene-d8	3610	10.34				
15067-26-2	Acenaphthene-d10	2490	14.213				
1517-22-2	Phenanthrene-d10	5000	16.971				
1719-03-5	Chrysene-d12	4480	21.162				
1520-96-3	Perylene-d12	4560	23.345				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/17/25
Client Sample ID:	RW8-MW01D1-20250617		SDG No.:	Q2345
Lab Sample ID:	Q2345-15		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037398.D	1	06/20/25 09:32	06/26/25 17:54	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.23		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		90%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		100%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		84%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.68	*	53 - 106		170%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.48		58 - 132		120%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1640	7.568				
1146-65-2	Naphthalene-d8	3480	10.34				
15067-26-2	Acenaphthene-d10	2380	14.213				
1517-22-2	Phenanthrene-d10	4640	16.971				
1719-03-5	Chrysene-d12	4430	21.162				
1520-96-3	Perylene-d12	4450	23.348				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/17/25	
Client Sample ID:	TT192D2-20250617		SDG No.:	Q2345	
Lab Sample ID:	Q2345-16		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	990	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037399.D	1	06/20/25 09:32	06/26/25 18:30	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.92		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		79%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.29		55 - 111		73%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.48	*	53 - 106		119%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.41		58 - 132		103%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1650	7.561				
1146-65-2	Naphthalene-d8	3570	10.341				
15067-26-2	Acenaphthene-d10	2380	14.213				
1517-22-2	Phenanthrene-d10	5030	16.972				
1719-03-5	Chrysene-d12	5330	21.162				
1520-96-3	Perylene-d12	5700	23.345				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2345

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168563BL	PB168563BL	2-Methylnaphthalene-d10	0.4	0.32	81		30	150
		Fluoranthene-d10	0.4	0.37	93		30	150
		Nitrobenzene-d5	0.4	0.30	75		55	111
		2-Fluorobiphenyl	0.4	0.33	81		53	106
		Terphenyl-d14	0.4	0.37	93		58	132
PB168563BS	PB168563BS	2-Methylnaphthalene-d10	0.4	0.49	123		30	150
		Fluoranthene-d10	0.4	0.35	86		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.39	98		58	132
PB168563BSD	PB168563BSD	2-Methylnaphthalene-d10	0.4	0.39	97		30	150
		Fluoranthene-d10	0.4	0.35	86		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.40	99		53	106
		Terphenyl-d14	0.4	0.38	94		58	132
Q2345-01	EB02-20250616	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.40	99		30	150
		Nitrobenzene-d5	0.4	0.30	75		55	111
		2-Fluorobiphenyl	0.4	0.44	111	*	53	106
		Terphenyl-d14	0.4	0.47	118		58	132
Q2345-13	TT189D2-20250617	2-Methylnaphthalene-d10	0.4	0.33	82		30	150
		Fluoranthene-d10	0.4	0.40	101		30	150
		Nitrobenzene-d5	0.4	0.28	69		55	111
		2-Fluorobiphenyl	0.4	0.60	150	*	53	106
		Terphenyl-d14	0.4	0.44	110		58	132
Q2345-13DL	TT189D2-20250617DL	2-Methylnaphthalene-d10	0.4	0.36	89		30	150
		Fluoranthene-d10	0.4	0.45	112		30	150
		Nitrobenzene-d5	0.4	0.28	69		55	111
		2-Fluorobiphenyl	0.4	0.58	146	*	53	106
		Terphenyl-d14	0.4	0.47	117		58	132
Q2345-14	TT150S1-20250617	2-Methylnaphthalene-d10	0.4	0.33	81		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
		2-Fluorobiphenyl	0.4	0.55	137	*	53	106
		Terphenyl-d14	0.4	0.49	122		58	132
Q2345-15	RW8-MW01D1-20250617	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.40	100		30	150
		Nitrobenzene-d5	0.4	0.34	84		55	111
		2-Fluorobiphenyl	0.4	0.68	170	*	53	106
		Terphenyl-d14	0.4	0.48	120		58	132
Q2345-16	TT192D2-20250617	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.29	73		55	111
		2-Fluorobiphenyl	0.4	0.48	119	*	53	106
		Terphenyl-d14	0.4	0.41	103		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2345

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN037363.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168563BS	1,4-Dioxane	0.4	0.32	ug/L	80				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2345

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN037364.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168563BSD	1,4-Dioxane	0.4	0.30	ug/L	75	6			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168563BL

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q2345

SAS No.: Q2345 SDG NO.: Q2345

Lab File ID: BN037361.D

Lab Sample ID: PB168563BL

Instrument ID: BNA_N

Date Extracted: 06/20/2025

Matrix: (soil/water) Water

Date Analyzed: 06/20/2025

Level: (low/med) LOW

Time Analyzed: 22:16

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168563BS	PB168563BS	BN037363.D	06/20/2025
PB168563BSD	PB168563BSD	BN037364.D	06/21/2025
EB02-20250616	Q2345-01	BN037395.D	06/26/2025
TT189D2-20250617	Q2345-13	BN037396.D	06/26/2025
TT150S1-20250617	Q2345-14	BN037397.D	06/26/2025
RW8-MW01D1-20250617	Q2345-15	BN037398.D	06/26/2025
TT192D2-20250617	Q2345-16	BN037399.D	06/26/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q2345 SDG NO.: Q2345

Lab File ID: BN037351.D

DFTPP Injection Date: 06/20/2025

Instrument ID: BNA_N

DFTPP Injection Time: 15:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.7) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	4.9
441	Present, but less than mass 443	85.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.4 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037353.D	06/20/2025	16:51
SSTDICC0.2	SSTDICC0.2	BN037354.D	06/20/2025	17:27
SSTDICCC0.4	SSTDICCC0.4	BN037355.D	06/20/2025	18:03
SSTDICC0.8	SSTDICC0.8	BN037356.D	06/20/2025	18:39
SSTDICC1.6	SSTDICC1.6	BN037357.D	06/20/2025	19:15
SSTDICC3.2	SSTDICC3.2	BN037358.D	06/20/2025	19:51
SSTDICC5.0	SSTDICC5.0	BN037359.D	06/20/2025	20:27
PB168563BL	PB168563BL	BN037361.D	06/20/2025	22:16
PB168563BS	PB168563BS	BN037363.D	06/20/2025	23:28
PB168563BSD	PB168563BSD	BN037364.D	06/21/2025	00:04
SSTDCCC0.4EC	SSTDCCC0.4	BN037365.D	06/21/2025	01:17

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q2345 SDG NO.: Q2345

Lab File ID: BN037385.D

DFTPP Injection Date: 06/26/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:01

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (0.9) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	4.9
441	Present, but less than mass 443	78.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.1 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037386.D	06/26/2025	10:41
SSTDICC0.2	SSTDICC0.2	BN037387.D	06/26/2025	11:17
SSTDICCC0.4	SSTDICCC0.4	BN037388.D	06/26/2025	11:53
SSTDICC0.8	SSTDICC0.8	BN037389.D	06/26/2025	12:29
SSTDICC1.6	SSTDICC1.6	BN037390.D	06/26/2025	13:05
SSTDICC3.2	SSTDICC3.2	BN037391.D	06/26/2025	13:41
SSTDICC5.0	SSTDICC5.0	BN037392.D	06/26/2025	14:17
EB02-20250616	Q2345-01	BN037395.D	06/26/2025	16:06
TT189D2-20250617	Q2345-13	BN037396.D	06/26/2025	16:42
TT150S1-20250617	Q2345-14	BN037397.D	06/26/2025	17:18
RW8-MW01D1-20250617	Q2345-15	BN037398.D	06/26/2025	17:54
TT192D2-20250617	Q2345-16	BN037399.D	06/26/2025	18:30
SSTDCCC0.4EC	SSTDCCC0.4	BN037401.D	06/26/2025	19:42

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q2345

SDG NO.: Q2345

Lab File ID: BN037416.D

DFTPP Injection Date: 06/27/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (1.1) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	5.1
441	Present, but less than mass 443	79.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.4 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037417.D	06/27/2025	10:09
TT189D2-20250617DL	Q2345-13DL	BN037419.D	06/27/2025	11:21
SSTDCCC0.4EC	SSTDCCC0.4	BN037423.D	06/27/2025	15:34

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/20/2025

Lab File ID: BN037355.D Time Analyzed: 18:03

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1912	7.568	4157	10.34	2811	14.21
UPPER LIMIT	3824	8.068	8314	10.84	5622	14.713
LOWER LIMIT	956	7.068	2078.5	9.84	1405.5	13.713
EPA SAMPLE NO.						
01 PB168563BL	1968	7.57	4045	10.35	2736	14.22
02 PB168563BS	1960	7.57	4204	10.34	2586	14.21
03 PB168563BSD	1885	7.57	4095	10.34	2623	14.21

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/20/2025

Lab File ID: BN037355.D Time Analyzed: 18:03

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5776	16.971	4813	21.171	4943	23.354
UPPER LIMIT	11552	17.471	9626	21.671	9886	23.854
LOWER LIMIT	2888	16.471	2406.5	20.671	2471.5	22.854
EPA SAMPLE NO.						
01 PB168563BL	4864	16.98	4288	21.17	3457	23.36
02 PB168563BS	4830	16.97	3875	21.17	2749	23.35
03 PB168563BSD	5035	16.97	4193	21.16	4470	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/26/2025

Lab File ID: BN037388.D Time Analyzed: 11:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2117	7.568	4565	10.34	3048	14.21
UPPER LIMIT	4234	8.068	9130	10.84	6096	14.713
LOWER LIMIT	1058.5	7.068	2282.5	9.84	1524	13.713
EPA SAMPLE NO.						
01 EB02-20250616	1748	7.57	3967	10.34	2665	14.21
02 TT189D2-20250617	1675	7.57	3636	10.34	2498	14.21
03 TT150S1-20250617	1586	7.57	3605	10.34	2491	14.21
04 RW8-MW01D1-20250617	1638	7.57	3476	10.34	2375	14.21
05 TT192D2-20250617	1647	7.56	3570	10.34	2375	14.21

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/26/2025

Lab File ID: BN037388.D Time Analyzed: 11:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6284	16.959	5612	21.162	5950	23.345
UPPER LIMIT	12568	17.459	11224	21.662	11900	23.845
LOWER LIMIT	3142	16.459	2806	20.662	2975	22.845
EPA SAMPLE NO.						
01 EB02-20250616	5237	16.97	4862	21.16	4950	23.35
02 TT189D2-20250617	5119	16.97	5273	21.16	5674	23.35
03 TT150S1-20250617	4999	16.97	4478	21.16	4562	23.35
04 RW8-MW01D1-20250617	4640	16.97	4434	21.16	4454	23.35
05 TT192D2-20250617	5031	16.97	5333	21.16	5697	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/27/2025

Lab File ID: BN037417.D Time Analyzed: 10:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1790	7.56	3949	10.34	2792	14.21
UPPER LIMIT	3580	8.06	7898	10.84	5584	14.713
LOWER LIMIT	895	7.06	1974.5	9.84	1396	13.713
EPA SAMPLE NO.						
01 TT189D2-20250617DL	1828	7.56	3778	10.34	2817	14.21

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/27/2025

Lab File ID: BN037417.D Time Analyzed: 10:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6217	16.959	6164	21.162	6622	23.345
UPPER LIMIT	12434	17.459	12328	21.662	13244	23.845
LOWER LIMIT	3108.5	16.459	3082	20.662	3311	22.845
EPA SAMPLE NO.						
01 TT189D2-20250617DL	5807	16.97	6103	21.16	6394	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB168563BL		SDG No.:	Q2345
Lab Sample ID:	PB168563BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037361.D	1	06/20/25 09:32	06/20/25 22:16	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		81%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		93%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		75%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		81%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.37		58 - 132		93%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1970	7.568				
1146-65-2	Naphthalene-d8	4050	10.351				
15067-26-2	Acenaphthene-d10	2740	14.224				
1517-22-2	Phenanthrene-d10	4860	16.984				
1719-03-5	Chrysene-d12	4290	21.171				
1520-96-3	Perylene-d12	3460	23.357				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB168563BS		SDG No.:	Q2345
Lab Sample ID:	PB168563BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037363.D	1	06/20/25 09:32	06/20/25 23:28	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.32		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.49		30 - 150		123%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		92%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		97%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		58 - 132		98%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1960	7.568				
1146-65-2	Naphthalene-d8	4200	10.34				
15067-26-2	Acenaphthene-d10	2590	14.213				
1517-22-2	Phenanthrene-d10	4830	16.971				
1719-03-5	Chrysene-d12	3880	21.171				
1520-96-3	Perylene-d12	2750	23.351				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB168563BSD		SDG No.:	Q2345
Lab Sample ID:	PB168563BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037364.D	1	06/20/25 09:32	06/21/25 00:04	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.30		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.39		30 - 150		97%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		99%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		94%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1890	7.568				
1146-65-2	Naphthalene-d8	4100	10.34				
15067-26-2	Acenaphthene-d10	2620	14.213				
1517-22-2	Phenanthrene-d10	5040	16.971				
1719-03-5	Chrysene-d12	4190	21.162				
1520-96-3	Perylene-d12	4470	23.354				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN062125.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 20 23:41:54 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037353.D 0.2 =BN037354.D 0.4 =BN037355.D 0.8 =BN037356.D 1.6 =BN037357.D 3.2 =BN037358.D 5 =BN037359.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.506	0.390	0.412	0.417	0.370	0.346	0.407	13.63	
3)	n-Nitrosodimet...	0.392	0.394	0.367	0.391	0.354	0.338	0.373	6.37	
4) S	2-Fluorophenol	0.854	0.818	0.751	0.781	0.834	0.786	0.771	0.799	4.65
5) S	Phenol-d6	0.781	0.766	0.766	0.785	0.891	0.878	0.896	0.823	7.45
6)	bis(2-Chloroet...	0.719	0.546	0.707	0.738	0.826	0.786	0.791	0.730	12.59
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.254	0.292	0.319	0.318	0.363	0.354	0.361	0.323	12.43
9)	Naphthalene	1.056	1.056	1.046	1.014	1.105	1.052	1.063	1.056	2.54
10)	Hexachlorobuta...	0.452	0.434	0.441	0.407	0.424	0.384	0.376	0.417	6.95
11) SURR2	Methylnaphth...	0.619	0.638	0.666	0.610	0.666	0.664	0.675	0.648	3.96
12)	2-Methylnaphth...	0.703	0.692	0.711	0.704	0.777	0.778	0.787	0.736	5.73
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.219	0.226	0.230	0.231	0.253	0.238	0.246	0.235	4.96
15) S	2-Fluorobiphenyl	1.705	1.675	1.777	1.714	1.897	1.752	1.778	1.757	4.15
16)	Acenaphthylene	1.646	1.636	1.597	1.595	1.797	1.717	1.786	1.682	5.06
17)	Acenaphthene	1.108	1.070	1.061	1.051	1.174	1.123	1.160	1.107	4.40
18)	Fluorene	1.499	1.470	1.506	1.490	1.660	1.605	1.660	1.556	5.34
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.070	0.079	0.097	0.110	0.107	0.114	0.096	18.63	
21)	4-Bromophenyl-...	0.264	0.267	0.279	0.284	0.305	0.295	0.299	0.285	5.55
22)	Hexachlorobenzene	0.322	0.319	0.314	0.304	0.324	0.296	0.292	0.310	4.11
23)	Atrazine	0.221	0.215	0.218	0.220	0.239	0.239	0.238	0.227	4.74
24)	Pentachlorophenol	0.131	0.137	0.157	0.169	0.161	0.170	0.154	10.69	
25)	Phenanthrene	1.108	1.075	1.104	1.139	1.242	1.221	1.222	1.158	5.88
26)	Anthracene	0.993	0.984	0.990	1.054	1.150	1.137	1.171	1.068	7.75
27) SURR	Fluoranthene-d10	1.097	1.070	1.161	1.166	1.235	1.151	1.158	1.148	4.62
28)	Fluoranthene	1.412	1.343	1.367	1.492	1.605	1.512	1.518	1.464	6.39
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.726	1.690	1.660	1.444	1.572	1.642	1.643	1.625	5.73
31) S	Terphenyl-d14	0.949	0.911	0.925	0.829	0.909	0.935	0.921	0.912	4.25
32)	Benzo(a)anthra...	1.309	1.168	1.216	1.278	1.431	1.372	1.429	1.315	7.76
33)	Chrysene	1.752	1.706	1.611	1.481	1.586	1.528	1.495	1.594	6.52
34)	Bis(2-ethylhex...	0.606	0.541	0.487	0.520	0.531	0.554	0.540	7.34	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN062125.M

36)	Indeno(1,2,3-c...	1.741	1.715	1.695	1.761	1.974	1.819	1.856	1.794	5.42
37)	Benzo(b)fluora...	1.428	1.380	1.444	1.392	1.541	1.532	1.587	1.472	5.50
38)	Benzo(k)fluora...	1.576	1.593	1.569	1.466	1.671	1.617	1.686	1.597	4.59
39) C	Benzo(a)pyrene	1.320	1.249	1.274	1.247	1.399	1.348	1.395	1.319	4.90
40)	Dibenzo(a,h)an...	1.179	1.185	1.236	1.355	1.561	1.446	1.478	1.348	11.32
41)	Benzo(g,h,i)pe...	1.620	1.554	1.589	1.560	1.720	1.577	1.598	1.603	3.53

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN062625.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 26 16:06:33 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037386.D 0.2 =BN037387.D 0.4 =BN037388.D 0.8 =BN037389.D 1.6 =BN037390.D 3.2 =BN037391.D 5 =BN037392.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.470	0.369	0.366	0.377	0.363	0.331	0.379	12.38	
3)	n-Nitrosodimet...	0.377	0.374	0.366	0.385	0.374	0.371	0.375	1.73	
4) S	2-Fluorophenol	0.771	0.778	0.804	0.697	0.754	0.773	0.818	0.771	5.09
5) S	Phenol-d6	0.663	0.706	0.812	0.737	0.839	0.886	0.951	0.799	12.86
6)	bis(2-Chloroet...	0.574	0.661	0.718	0.694	0.765	0.775	0.790	0.711	10.74
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.270	0.281	0.311	0.296	0.345	0.353	0.389	0.321	13.45
9)	Naphthalene	1.042	1.001	0.999	0.972	1.048	1.050	1.096	1.030	4.06
10)	Hexachlorobuta...	0.407	0.410	0.413	0.400	0.422	0.404	0.405	0.409	1.74
11) SURR	2-Methylnaphth...	0.532	0.567	0.576	0.568	0.628	0.656	0.807	0.619	14.98
12)	2-Methylnaphth...	0.635	0.665	0.684	0.662	0.746	0.767	0.803	0.709	8.89
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.194	0.207	0.239	0.225	0.251	0.256	0.271	0.235	11.68
15) S	2-Fluorobiphenyl	1.548	1.656	1.723	1.652	1.798	1.801	1.910	1.727	6.95
16)	Acenaphthylene	1.585	1.600	1.576	1.538	1.714	1.741	1.858	1.659	6.95
17)	Acenaphthene	1.030	1.027	1.045	1.009	1.123	1.147	1.203	1.083	6.85
18)	Fluorene	1.444	1.417	1.476	1.420	1.603	1.643	1.706	1.530	7.74
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.084	0.085	0.087	0.113	0.119	0.128	0.103	19.13	
21)	4-Bromophenyl-...	0.264	0.253	0.270	0.269	0.303	0.302	0.316	0.282	8.58
22)	Hexachlorobenzene	0.310	0.299	0.302	0.291	0.311	0.301	0.311	0.303	2.45
23)	Atrazine	0.200	0.208	0.213	0.204	0.232	0.242	0.266	0.224	10.81
24)	Pentachlorophenol	0.164	0.155	0.148	0.165	0.165	0.168	0.183	0.164	7.30
25)	Phenanthrene	1.056	1.067	1.080	1.038	1.172	1.190	1.287	1.127	8.12
26)	Anthracene	0.935	0.972	0.969	0.959	1.076	1.120	1.222	1.036	10.28
27) SURR	Fluoranthene-d10	1.073	1.137	1.063	1.028	1.121	1.154	1.447	1.146	12.22
28)	Fluoranthene	1.377	1.398	1.362	1.320	1.487	1.507	1.624	1.439	7.32
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.459	1.449	1.534	1.431	1.561	1.550	1.612	1.514	4.47
31) S	Terphenyl-d14	0.800	0.794	0.868	0.796	0.878	0.892	0.943	0.853	6.77
32)	Benzo(a)anthra...	1.149	1.202	1.167	1.152	1.318	1.364	1.446	1.257	9.46
33)	Chrysene	1.630	1.573	1.583	1.491	1.554	1.510	1.551	1.556	2.98
34)	Bis(2-ethylhex...	0.509	0.504	0.466	0.481	0.488	0.537	0.498	5.00	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN062625.M

36)	Indeno(1,2,3-c...	1.487	1.515	1.577	1.620	1.786	1.853	1.967	1.686	10.88
37)	Benzo(b)fluora...	1.329	1.305	1.318	1.320	1.464	1.504	1.632	1.410	8.93
38)	Benzo(k)fluora...	1.408	1.389	1.462	1.430	1.552	1.613	1.697	1.507	7.73
39) C	Benzo(a)pyrene	1.211	1.160	1.183	1.174	1.286	1.341	1.426	1.254	8.01
40)	Dibenzo(a,h)an...	1.050	1.138	1.211	1.229	1.394	1.485	1.561	1.296	14.53
41)	Benzo(g,h,i)pe...	1.425	1.473	1.477	1.462	1.617	1.658	1.725	1.548	7.51

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG No.: Q2345

Instrument ID: BNA_N Calibration Date/Time: 06/21/2025 01:17

Lab File ID: BN037365.D Init. Calib. Date(s): 06/20/2025 06/20/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 16:51 20:27

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.648	0.670		3.4	50.0
Fluoranthene-d10	1.148	1.189		3.6	50.0
2-Fluorophenol	0.799	0.740		-7.4	50.0
Phenol-d6	0.823	0.753		-8.5	50.0
Nitrobenzene-d5	0.323	0.318		-1.5	50.0
2-Fluorobiphenyl	1.757	1.777		1.1	50.0
2,4,6-Tribromophenol	0.235	0.218		-7.2	50.0
Terphenyl-d14	0.912	0.903		-1.0	50.0
1,4-Dioxane	0.407	0.399		-2.0	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG No.: Q2345

Instrument ID: BNA_N Calibration Date/Time: 06/26/2025 19:42

Lab File ID: BN037401.D Init. Calib. Date(s): 06/26/2025 06/26/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 10:41 14:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.619	0.588		-5.0	50.0
Fluoranthene-d10	1.146	1.129		-1.5	50.0
2-Fluorophenol	0.771	0.791		2.6	50.0
Phenol-d6	0.799	0.773		-3.3	50.0
Nitrobenzene-d5	0.321	0.299		-6.9	50.0
2-Fluorobiphenyl	1.727	1.679		-2.8	50.0
2,4,6-Tribromophenol	0.235	0.263		11.9	50.0
Terphenyl-d14	0.853	0.857		0.5	50.0
1,4-Dioxane	0.379	0.416		9.8	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG No.: Q2345

Instrument ID: BNA_N Calibration Date/Time: 06/27/2025 10:09

Lab File ID: BN037417.D Init. Calib. Date(s): 06/26/2025 06/26/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 10:41 14:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.619	0.578		-6.6	20.0
Fluoranthene-d10	1.146	1.126		-1.7	20.0
2-Fluorophenol	0.771	0.793		2.9	20.0
Phenol-d6	0.799	0.783		-2.0	20.0
Nitrobenzene-d5	0.321	0.297		-7.5	20.0
2-Fluorobiphenyl	1.727	1.669		-3.4	20.0
2,4,6-Tribromophenol	0.235	0.245		4.3	20.0
Terphenyl-d14	0.853	0.854		0.1	20.0
1,4-Dioxane	0.379	0.367		-3.2	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG No.: Q2345
 Instrument ID: BNA_N Calibration Date/Time: 06/27/2025 15:34
 Lab File ID: BN037423.D Init. Calib. Date(s): 06/26/2025 06/26/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 10:41 14:17
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.619	0.609		-1.6	50.0
Fluoranthene-d10	1.146	1.092		-4.7	50.0
2-Fluorophenol	0.771	0.857		11.2	50.0
Phenol-d6	0.799	0.861		7.8	50.0
Nitrobenzene-d5	0.321	0.290		-9.7	50.0
2-Fluorobiphenyl	1.727	1.640		-5.0	50.0
2,4,6-Tribromophenol	0.235	0.221		-6.0	50.0
Terphenyl-d14	0.853	0.853		0.0	50.0
1,4-Dioxane	0.379	0.383		1.1	50.0

All other compounds must meet a minimum RRF of 0.010.