

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG No.: Q2345

Instrument ID: MSVOA\_X Calibration Date/Time: 06/18/2025 20:01

Lab File ID: VX046756.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.534 | 0.495  |            | -7.3   | 50    |
| Chloromethane                  | 0.576 | 0.545  | 0.1        | -5.38  | 50    |
| Vinyl Chloride                 | 0.614 | 0.577  |            | -6.03  | 50    |
| Bromomethane                   | 0.346 | 0.354  |            | 2.31   | 50    |
| Chloroethane                   | 0.372 | 0.348  |            | -6.45  | 50    |
| Trichlorofluoromethane         | 0.933 | 0.862  |            | -7.61  | 50    |
| 1,1,2-Trichlorotrifluoroethane | 0.572 | 0.515  |            | -9.97  | 50    |
| 1,1-Dichloroethene             | 0.552 | 0.517  |            | -6.34  | 50    |
| Acetone                        | 0.231 | 0.176  |            | -23.81 | 50    |
| Carbon Disulfide               | 1.668 | 1.433  |            | -14.09 | 50    |
| Methyl tert-butyl Ether        | 1.720 | 1.510  |            | -12.21 | 50    |
| Methyl Acetate                 | 0.586 | 0.539  |            | -8.02  | 50    |
| Methylene Chloride             | 0.641 | 0.573  |            | -10.61 | 50    |
| trans-1,2-Dichloroethene       | 0.591 | 0.539  |            | -8.8   | 50    |
| 1,1-Dichloroethane             | 1.092 | 1.005  | 0.1        | -7.97  | 50    |
| Cyclohexane                    | 1.036 | 0.922  |            | -11    | 50    |
| 2-Butanone                     | 0.326 | 0.268  |            | -17.79 | 50    |
| Carbon Tetrachloride           | 0.518 | 0.455  |            | -12.16 | 50    |
| cis-1,2-Dichloroethene         | 0.694 | 0.632  |            | -8.93  | 50    |
| Bromochloromethane             | 0.495 | 0.505  |            | 2.02   | 50    |
| Chloroform                     | 1.092 | 1.017  |            | -6.87  | 50    |
| 1,1,1-Trichloroethane          | 0.958 | 0.867  |            | -9.5   | 50    |
| Methylcyclohexane              | 0.628 | 0.546  |            | -13.06 | 50    |
| Benzene                        | 1.413 | 1.295  |            | -8.35  | 50    |
| 1,2-Dichloroethane             | 0.497 | 0.448  |            | -9.86  | 50    |
| Trichloroethene                | 0.360 | 0.325  |            | -9.72  | 50    |
| 1,2-Dichloropropane            | 0.350 | 0.321  |            | -8.29  | 50    |
| Bromodichloromethane           | 0.514 | 0.475  |            | -7.59  | 50    |
| 4-Methyl-2-Pentanone           | 0.411 | 0.348  |            | -15.33 | 50    |
| Toluene                        | 0.876 | 0.801  |            | -8.56  | 50    |
| t-1,3-Dichloropropene          | 0.477 | 0.432  |            | -9.43  | 50    |
| cis-1,3-Dichloropropene        | 0.541 | 0.489  |            | -9.61  | 50    |
| 1,1,2-Trichloroethane          | 0.327 | 0.294  |            | -10.09 | 50    |
| 2-Hexanone                     | 0.284 | 0.237  |            | -16.55 | 50    |
| Dibromochloromethane           | 0.385 | 0.345  |            | -10.39 | 50    |
| 1,2-Dibromoethane              | 0.332 | 0.302  |            | -9.04  | 50    |
| Tetrachloroethene              | 0.346 | 0.299  |            | -13.58 | 50    |
| Chlorobenzene                  | 1.104 | 0.976  | 0.3        | -11.59 | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG No.: Q2345

Instrument ID: MSVOA\_X Calibration Date/Time: 06/18/2025 20:01

Lab File ID: VX046756.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| Ethyl Benzene               | 1.929 | 1.726  |            | -10.52 | 50    |
| m/p-Xylenes                 | 0.719 | 0.649  |            | -9.74  | 50    |
| o-Xylene                    | 0.673 | 0.627  |            | -6.84  | 50    |
| Styrene                     | 1.179 | 1.077  |            | -8.65  | 50    |
| Bromoform                   | 0.282 | 0.247  | 0.1        | -12.41 | 50    |
| Isopropylbenzene            | 3.682 | 3.343  |            | -9.21  | 50    |
| 1,1,2,2-Tetrachloroethane   | 1.028 | 0.894  | 0.3        | -13.03 | 50    |
| 1,3-Dichlorobenzene         | 1.728 | 1.510  |            | -12.62 | 50    |
| 1,4-Dichlorobenzene         | 1.775 | 1.497  |            | -15.66 | 50    |
| 1,2-Dichlorobenzene         | 1.615 | 1.425  |            | -11.77 | 50    |
| 1,2-Dibromo-3-Chloropropane | 0.203 | 0.162  |            | -20.2  | 50    |
| 1,2,4-Trichlorobenzene      | 1.148 | 1.012  |            | -11.85 | 50    |
| 1,2,3-Trichlorobenzene      | 1.098 | 0.970  |            | -11.66 | 50    |
| 1,2-Dichloroethane-d4       | 0.692 | 0.697  |            | 0.72   | 50    |
| Dibromofluoromethane        | 0.331 | 0.344  |            | 3.93   | 50    |
| Toluene-d8                  | 1.189 | 1.233  |            | 3.7    | 50    |
| 4-Bromofluorobenzene        | 0.443 | 0.468  |            | 5.64   | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.