

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061825\  
 Data File : VX046739.D  
 Acq On : 18 Jun 2025 13:58  
 Operator : JC/MD  
 Sample : Q2345-07  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BPOW6-5-20250616

Quant Time: Jun 19 01:24:11 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061725W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 18 03:09:16 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.568  | 168  | 138024   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.769  | 114  | 229333   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 204785   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.018 | 152  | 98514    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 92104    | 48.244   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 96.480% |
| 35) Dibromofluoromethane    | 5.403  | 113   | 73789    | 48.631   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 97.260% |
| 50) Toluene-d8              | 8.653  | 98    | 268168   | 49.192   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 98.380% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 101037   | 49.720   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =    | 99.440% |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061825\  
Data File : VX046739.D  
Acq On : 18 Jun 2025 13:58  
Operator : JC/MD  
Sample : Q2345-07  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BPOW6-5-20250616

Quant Time: Jun 19 01:24:11 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061725W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 18 03:09:16 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.568 min Scan# 71  
Delta R.T. 0.006 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58




Instrument :  
MSVOA\_X  
ClientSampleId :  
BPOW6-5-20250616

Tgt Ion:168 Resp: 138024  
Ion Ratio Lower Upper  
168 100  
99 61.2 48.5 72.7



#33  
1,2-Dichloroethane-d4  
Concen: 48.244 ug/l  
RT: 5.964 min Scan# 800  
Delta R.T. -0.000 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58

Tgt Ion: 65 Resp: 92104  
Ion Ratio Lower Upper  
65 100  
67 51.7 0.0 105.4








#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58




Instrument :  
MSVOA\_X  
ClientSampleId :  
BPOW6-5-20250616


Tgt Ion:114 Resp: 229333  
Ion Ratio Lower Upper  
114 100  
63 21.9 0.0 44.2  
88 17.0 0.0 33.2



#35  
Dibromofluoromethane  
Concen: 48.631 ug/l  
RT: 5.403 min Scan# 708  
Delta R.T. 0.006 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58

Tgt Ion:113 Resp: 73789  
Ion Ratio Lower Upper  
113 100  
111 103.3 82.0 123.0  
192 18.7 15.3 22.9









#50  
Toluene-d8  
Concen: 49.192 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. 0.006 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58




Instrument :  
MSVOA\_X  
ClientSampleId :  
BPOW6-5-20250616


Tgt Ion: 98 Resp: 268168  
Ion Ratio Lower Upper  
98 100  
100 66.9 53.0 79.4



#62  
4-Bromofluorobenzene  
Concen: 49.720 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. -0.000 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58

Tgt Ion: 95 Resp: 101037  
Ion Ratio Lower Upper  
95 100  
174 74.8 0.0 150.4  
176 72.1 0.0 145.0









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58

Instrument :  
MSVOA\_X  
ClientSampleId :  
BPOW6-5-20250616

Tgt Ion:117 Resp: 204785  
Ion Ratio Lower Upper  
117 100  
82 55.7 44.6 66.8  
119 32.0 25.8 38.8



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.018 min Scan# 1793  
Delta R.T. -0.000 min  
Lab File: VX046739.D  
Acq: 18 Jun 2025 13:58

Tgt Ion:152 Resp: 98514  
Ion Ratio Lower Upper  
152 100  
115 61.2 43.2 129.6  
150 157.7 0.0 346.8

