

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345
 Lab File ID: VX046728.D Date Analyzed: 06/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:33
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	153144	5.56	247293	6.76	216990	10.06
UPPER LIMIT	306288	6.056	494586	7.263	433980	10.555
LOWER LIMIT	76572	5.056	123647	6.263	108495	9.555
EPA SAMPLE NO.						
TB03-20250616	131404	5.57	213527	6.77	190496	10.06
BPOW6-1-20250616	208218	5.56	333734	6.77	292626	10.06
BPOW6-2-20250616	133570	5.56	218146	6.77	194756	10.06
DUP05-20250616	201684	5.57	330858	6.77	287321	10.06
BPOW6-3-20250616	203124	5.56	332589	6.77	290067	10.06
BPOW6-5-20250616	138024	5.57	229333	6.77	204785	10.06
BPOW6-5-20250616MS	143057	5.56	240999	6.77	215546	10.06
BPOW6-5-20250616MSD	136146	5.56	229178	6.77	208450	10.06
BPOW6-6-20250616	202843	5.57	334138	6.77	291485	10.06
BPOW6-10-20250616	221164	5.56	359780	6.77	315084	10.06
BPOW6-9-20250616	133892	5.57	222699	6.78	198876	10.06
TT189D2-20250617	217878	5.57	358276	6.77	311829	10.06
TT150S1-20250617	128949	5.57	213744	6.77	192588	10.06
RW8-MW01D1-20250617	226176	5.57	367631	6.77	322256	10.06
TT192D2-20250617	133130	5.57	221280	6.77	195519	10.06
BPOW6-4-20250616	210931	5.56	348901	6.77	308900	10.06
VX0618WBL01	167417	5.56	270618	6.76	242123	10.06
VX0618WBS01	156160	5.56	258286	6.77	234034	10.06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345
 Lab File ID: VX046728.D Date Analyzed: 06/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:33
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	153144	5.56	247293	6.76	216990	10.06
UPPER LIMIT	306288	6.056	494586	7.263	433980	10.555
LOWER LIMIT	76572	5.056	123647	6.263	108495	9.555
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.