

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345
 Lab File ID: VX046728.D Date Analyzed: 06/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:33
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	104970	12.018				
UPPER LIMIT	209940	12.518				
LOWER LIMIT	52485	11.518				
EPA SAMPLE NO.						
TB03-20250616	88415	12.02				
BPOW6-1-20250616	140376	12.02				
BPOW6-2-20250616	93153	12.02				
DUP05-20250616	137822	12.02				
BPOW6-3-20250616	141430	12.02				
BPOW6-5-20250616	98514	12.02				
BPOW6-5-20250616MS	107663	12.02				
BPOW6-5-20250616MSD	102802	12.02				
BPOW6-6-20250616	140253	12.02				
BPOW6-10-20250616	151979	12.02				
BPOW6-9-20250616	94577	12.02				
TT189D2-20250617	149381	12.02				
TT150S1-20250617	93953	12.02				
RW8-MW01D1-20250617	153646	12.02				
TT192D2-20250617	93964	12.02				
BPOW6-4-20250616	147369	12.02				
VX0618WBL01	121135	12.02				
VX0618WBS01	121898	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2345 SAS No.: Q2345 SDG NO.: Q2345
 Lab File ID: VX046728.D Date Analyzed: 06/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:33
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	104970	12.018				
UPPER LIMIT	209940	12.518				
LOWER LIMIT	52485	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.