

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2348
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2348 **SAS No.:** Q2348
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3950	4000	99	95 - 105	P	06/19/2025	14:45	LB136210

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2348
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2348 **SAS No.:** Q2348
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	96.2	100	96	80 - 120	P	06/19/2025	14:57	LB136210

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2348
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2348 **SAS No.:** Q2348
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4980	5000	100	90 - 110	P	06/19/2025	15:33	LB136210
CCV02	Iron	5290	5000	106	90 - 110	P	06/19/2025	16:31	LB136210
CCV03	Iron	5130	5000	103	90 - 110	P	06/19/2025	16:44	LB136210
CCV04	Iron	4890	5000	98	90 - 110	P	06/19/2025	17:32	LB136210
CCV05	Iron	4900	5000	98	90 - 110	P	06/19/2025	18:25	LB136210
CCV06	Iron	4810	5000	96	90 - 110	P	06/19/2025	19:19	LB136210
CCV07	Iron	4890	5000	98	90 - 110	P	06/19/2025	19:53	LB136210
