



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q2348
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2348 **SAS No.:** Q2348
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	106	100	106	65 - 135	P	06/19/2025	15:06	LB136210