

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q2348
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2348 **SAS No.:** Q2348
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	104000	100000	104	85600	116500	06/19/2025	15:10	LB136210
ICSAB01	Iron	103000	99000	104	84400	114500	06/19/2025	15:16	LB136210
ICSA	Iron	104000	100000	104	85600	116500	06/19/2025	15:23	LB136210
ICSAB	Iron	103000	99000	104	84400	114500	06/19/2025	15:28	LB136210